



INTERNATIONAL
ACCREDITATION
SERVICE®

CERTIFICATE OF ACCREDITATION

This is to attest that

INTERTEK TESTING SERVICES NA, LTD.

1500 BRIGANTINE DRIVE
COQUITLAM, BRITISH COLUMBIA, V3K 7C1, CANADA

Calibration Laboratory CL-102

has met the requirements of AC204, *IAS Accreditation Criteria for Calibration Laboratories*, and has demonstrated compliance with ISO/IEC Standard 17025:2017, *General requirements for the competence of testing and calibration laboratories*. This organization is accredited to provide the services specified in the scope of accreditation.

Effective Date October 30, 2023

Expiration Date January 1, 2026



A handwritten signature in black ink that reads "Raj Nathan".

President

Visit www.iasonline.org for current accreditation information.

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

INTERTEK TESTING SERVICES NA, LTD.

www.intertek.com

Contact Name Peter Gildenstern

Contact Phone +1-604-528-8715

Accredited to ISO/IEC 17025:2017

Effective Date October 30, 2023

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)*

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (\pm)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Dimensional			
Length ⁶ Dial Indicators Calipers	0 in to 2 in 2 in to 6 in	0.0011 in 0.0011 in	WI-L-CA-Cali-1393 Gauge Blocks
Mechanical			
Force – Compression ⁶ (Structural Wood Bending & Compression Testers)	1 lbf to 50 lbf 50 lbf to 50,000 lbf 50,000 lbf to 200,000 lbf	0.1 % 0.32 % 0.54 %	ASTM E4, Load Cell WI-L-CA-Cali-1387 WI-L-CA-Cali-1388
Force – Tension ⁶ (Structural Wood Tension Testers)	1 lbf to 100,000 lbf	710 lbf	WI-L-CA-Cali-1389 WI-L-CA-Cali-1390 WI-L-CA-Cali-1391 WI-L-CA-Cali-1393
Electrical – DC/LF			
DC Voltage – Generate ³	0 mV to 329.9999 mV 0 V to 3.299999 V 0 V to 32.99999 V 30 V to 329.9999 V 100 V to 1000.000 V	20 μ V/V + 1 μ V 11 μ V/V + 2 μ V 12 μ V/V + 20 μ V 18 μ V/V + 150 μ V 18 μ V/V + 1500 μ V	WI-L-CA-Cali-1402 Fluke 5522A
DC Current – Generate ³	0 μ A to 329.999 μ A 0 mA to 3.29999 mA 0 mA to 32.9999 mA 0 mA to 329.999 mA 0 A to 1.09999 A 1.1 A to 2.99999 A 0 A to 10.9999 A 10 A to 150 A 150 A to 1000 A	150 μ A/A + 0.02 μ A 100 μ A/A + 0.05 μ A 100 μ A/A + 0.25 μ A 100 μ A/A + 2.5 μ A 200 μ A/A + 40 μ A 380 μ A/A + 40 μ A 500 μ A/A + 500 μ A 0.25 % + 0.02 A 0.25 % + 0.05 A	WI-L-CA-Cali-1402 Fluke 5522A WI-L-CA-Cali-1402 Fluke 5522A w/50 Turn Coil

* If information in this CMC is presented in non-SI units, the conversion factors stated in NIST Special Publication 811 "Guide for the Use of the International System of Units (SI)" apply.

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (\pm)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
AC Voltage – Generate ³	1.0 mV to 32.999 mV (10 Hz to 45 Hz) (45 Hz to 10 kHz) (10 kHz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 100 kHz) (100 kHz to 500 kHz)	800 μ V/V + 6 μ V 150 μ V/V + 6 μ V 200 μ V/V + 6 μ V 1000 μ V/V + 6 μ V 3500 μ V/V + 12 μ V 8000 μ V/V + 50 μ V	WI-L-CA-Cali-1402 Fluke 5522A
	33 mV to 329.999 mV (10 Hz to 45 Hz) (45 Hz to 10 kHz) (10 kHz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 100 kHz) (100 kHz to 500 kHz)	300 μ V/V + 8 μ V 145 μ V/V + 8 μ V 160 μ V/V + 8 μ V 350 μ V/V + 8 μ V 800 μ V/V + 32 μ V 2000 μ V/V + 70 μ V	
	0.33 V to 3.29999 V (10 Hz to 45 Hz) (45 Hz to 10 kHz) (10 kHz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 100 kHz) (100 kHz to 500 kHz)	300 μ V/V + 50 μ V 150 μ V/V + 60 μ V 190 μ V/V + 60 μ V 300 μ V/V + 50 μ V 700 μ V/V + 125 μ V 2400 μ V/V + 600 μ V	
	3.3 V to 32.9999 V (10 Hz to 45 Hz) (45 Hz to 10 kHz) (10 kHz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 100 kHz)	300 μ V/V + 650 μ V 150 μ V/V + 600 μ V 240 μ V/V + 600 μ V 350 μ V/V + 600 μ V 900 μ V/V + 1600 μ V	
	33 V to 329.999 V (10 Hz to 45 Hz) (45 Hz to 10 kHz) (10 kHz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 100 kHz)	190 μ V/V + 2 mV 200 μ V/V + 6 mV 250 μ V/V + 6 mV 300 μ V/V + 6 mV 2000 μ V/V + 50 mV	
	330 V to 1020 V (45 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz)	300 μ V/V + 10 mV 250 μ V/V + 10 mV 300 μ V/V + 10 mV	

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

AC Current – Generate ³	30 µA to 330 µA (10 Hz to 20 Hz) (20 Hz to 45 Hz) (45 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz) (10 kHz to 30 kHz)	0.2 % + 0.1 µA 0.15 % + 0.1 µA 0.125 % + 0.1 µA 0.3 % + 0.15 µA 0.8 % + 0.2 µA 1.6 % + 0.4 µA	WI-L-CA-Cali-1402 Fluke 5522A, LCOMP off
	330 µA to 3.3 mA (10 Hz to 20 Hz) (20 Hz to 45 Hz) (45 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz) (10 kHz to 30 kHz)	0.2 % + 0.15 µA 0.125 % + 0.15 µA 0.1 % + 0.15 µA 0.2 % + 0.2 µA 0.5 % + 0.3 µA 1.0 % + 0.6 µA	
	3.3 mA to 33 mA (10 Hz to 20 Hz) (20 Hz to 45 Hz) (45 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz) (10 kHz to 30 kHz)	0.18 % + 2 µA 0.09 % + 2 µA 0.04 % + 2 µA 0.08 % + 2 µA 0.2 % + 3 µA 0.4 % + 4 µA	
	33 mA to 330 mA (10 Hz to 20 Hz) (20 Hz to 45 Hz) (45 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz) (10 kHz to 30 kHz)	0.18 % + 20 µA 0.09 % + 20 µA 0.04 % + 20 µA 0.10 % + 50 µA 0.2 % + 100 µA 0.4 % + 200 µA	
	330 mA to 1.1 A (10 Hz to 45 Hz) (45 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz)	0.18 % + 0.1 mA 0.05 % + 0.1 mA 0.6 % + 1 mA 2.5 % + 5 mA	
	1.1 A to 3 A (10 Hz to 45 Hz) (45 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz)	0.18 % + 0.1 mA 0.06 % + 0.1 mA 0.6 % + 1 mA 2.5 % + 5 mA	
	3 A to 11 A (45 Hz to 100 Hz) (100 Hz to 1 kHz) (1 kHz to 5 kHz)	0.06 % + 2 mA 0.10 % + 2 mA 3 % + 2 mA	

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (\pm)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
AC Current – Generate ³ continued	11 A to 20.5 A (45 Hz to 100 Hz) (100 Hz to 1 kHz) (1 kHz to 5 kHz)	0.12 % + 5 mA 0.15 % + 5 mA 3 % + 5 mA	WI-L-CA-Cali-1402 Fluke 5522A, LCOMP off
	20 A to 1000 A (45 Hz to 440 Hz)	0.28 %+ 0.11 A	WI-L-CA-Cali-1402 Fluke 5522A w/50 Turn Coil
DC Resistance – Generate ³	0 Ω to 11 Ω	40 μΩ/Ω + 1 mΩ	WI-L-CA-Cali-1402
	11 Ω to 33 Ω	30 μΩ/Ω + 1.5 mΩ	Fluke 5522A
	33 Ω to 110 Ω	28 μΩ/Ω + 1.4 mΩ	
	110 Ω to 330 Ω	28 μΩ/Ω + 2 mΩ	
	330 Ω to 1.1 kΩ	28 μΩ/Ω + 2 mΩ	
	1.1 kΩ to 3.3 kΩ	28 μΩ/Ω + 0.02 Ω	
	3.3 kΩ to 11 kΩ	28 μΩ/Ω + 0.02 Ω	
	11 kΩ to 33 kΩ	28 μΩ/Ω + 0.2 Ω	
	33 kΩ to 110 kΩ	28 μΩ/Ω + 0.2 Ω	
	110 kΩ to 330 kΩ	32 μΩ/Ω + 2 Ω	
	330 kΩ to 1.1 MΩ	32 μΩ/Ω + 2 Ω	
	1.1 MΩ to 3.3 MΩ	60 μΩ/Ω + 30 Ω	
	3.3 MΩ to 11 MΩ	0.013 % + 50 Ω	
	11 MΩ to 33 MΩ	0.025 % + 2.5 kΩ	
	33 MΩ to 110 MΩ	0.05 % + 3 kΩ	
	110 MΩ to 1.1 GΩ	1.5 % + 500 kΩ	
DC Voltage – Measure ⁴	Up to 100 mV	37 μV/V + 35 μV	WI-L-CA-Cali-1402
	100 mV to 1 V	25 μV/V + 7 μV	Fluke 8846A
	1 V to 10 V	24 μV/V + 50 μV	
	10 V to 100 V	38 μV/V + 600 μV	
	100 V to 1000 V	41 μV/V + 10 mV	
	Up to 10 kV	0.03 % + 0.035 V	WI-L-CA-Cali-1402 Vitrek 4700
	10 kV to 100 kV	0.05 % + 0.35 V	WI-L-CA-Cali-1402 Vitrek 4700, HVL-100
DC Current Measure ⁴	up to 100 μA	0.05 % + 0.025 μA	WI-L-CA-Cali-1402
	100 μA to 1 mA	0.05 % + 0.05 μA	Fluke 8846A
	1 mA to 10 mA	0.05 % + 2 μA	
	10 mA to 100 mA	0.05 % + 5 μA	
	100 mA to 400 mA	0.05 % + 20 μA	
	400 mA to 1 A	0.05 % + 200 μA	
	1 A to 3 A	0.1 % + 600 μA	
	3 A to 10 A	0.15 % + 800 μA	
AC Voltage – Measure ⁴	up to 100 mV (3 Hz to 5 Hz)	1.0 % + 0.04 mV	WI-L-CA-Cali-1402
	(5 Hz to 10 Hz)	0.35 % + 0.04 mV	Fluke 8846A
	(10 Hz to 20 kHz)	0.06 % + 0.04 mV	
	(20 kHz to 50 kHz)	0.12 % + 0.05 mV	
	(50 kHz to 100 kHz)	0.6 % + 0.08 mV	
	(100 kHz to 300 kHz)	4.0 % + 0.50 mV	

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (\pm)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
AC Voltage – Measure ⁴ continued	100 mV to 1 V (3 Hz to 5 Hz) (5 Hz to 10 Hz) (10 Hz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 100 kHz) (100 kHz to 300 kHz)	1.0 % + 0.3 mV 0.35 % + 0.3 mV 0.06 % + 0.3 mV 0.12 % + 0.5 mV 0.6 % + 0.8 mV 4.0 % + 5 mV	WI-L-CA-Cali-1402 Fluke 8846A (cont'd.)
	1 V to 10 V (3 Hz to 5 Hz) (5 Hz to 10 Hz) (10 Hz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 100 kHz) (100 kHz to 300 kHz)	1.0 % + 3 mV 0.35 % + 3 mV 0.06 % + 3 mV 0.12 % + 5 mV 0.6 % + 8 mV 4.0 % + 50 mV	
	10 V to 100 V (3 Hz to 5 Hz) (5 Hz to 10 Hz) (10 Hz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 100 kHz) (100 kHz to 300 kHz)	1.0 % + 30 mV 0.35 % + 30 mV 0.06 % + 30 mV 0.12 % + 50 mV 0.6 % + 80 mV 4.0 % + 500 mV	
	100 V to 1000 V (3 Hz to 5 Hz) (5 Hz to 10 Hz) (10 Hz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 100 kHz) (100 kHz to 300 kHz)	1.0 % + 225 mV 0.35 % + 225 mV 0.06 % + 225 mV 0.12 % + 375 mV 0.6 % + 600 mV 4.0 % + 3750 mV	
	Up to 10 kV (30 Hz to 200 Hz) (200 Hz to 450 Hz)	0.12 % + 0.15 V 0.4 % + 0.15 V	WI-L-CA-Cali-1402 Vitrek 4700
	10 kV to 75 kV (30 Hz to 70 Hz) (70 Hz to 200 Hz)	0.12 % + 0.65 V 1 % + 0.65 V	WI-L-CA-Cali-1402 Vitrek 4700, HVL-100
AC Current – Measure ⁴	5 μ A to 100 μ A (3 Hz to 5 Hz) (5 Hz to 10 Hz) (10 Hz to 5 kHz) (5 kHz to 10 kHz)	1.1 % + 0.06 μ A 0.35 % + 0.06 μ A 0.15 % + 0.06 μ A 0.35 % + 0.7 μ A	WI-L-CA-Cali-1402 Fluke 8846A
	100 μ A to 1 mA (3 Hz to 5 Hz) (5 Hz to 10 Hz)	1.0 % + 0.4 μ A 0.3 % + 0.4 μ A	

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (\pm)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
AC Current – Measure ⁴ continued	100 μ A to 1 mA (10 Hz to 5 kHz) (5 kHz to 10 kHz)	0.1 % + 0.4 μ A 0.2 % + 2.5 μ A	WI-L-CA-Cali-1402 Fluke 8846A
	1 mA to 10 mA (3 Hz to 5 Hz) (5 Hz to 10 Hz) (10 Hz to 5 kHz) (5 kHz to 10 kHz)	1.1 % + 6 μ A 0.35 % + 6 μ A 0.15 % + 6 μ A 0.35 % + 70 μ A	
	10 mA to 100 mA (3 Hz to 5 Hz) (5 Hz to 10 Hz) (10 Hz to 5 kHz) (5 kHz to 10 kHz)	1.0 % + 40 μ A 0.3 % + 40 μ A 0.1 % + 40 μ A 0.2 % + 250 μ A	
	100 mA to 400 mA (3 Hz to 5 Hz) (5 Hz to 10 Hz) (10 Hz to 5 kHz) (5 kHz to 10 kHz)	1.0 % + 400 μ A 0.3 % + 400 μ A 0.1 % + 400 μ A 0.2 % + 2.8 mA	
	400 mA to 1A (3 Hz to 5 Hz) (5 Hz to 10 Hz) (10 Hz to 5 kHz) (5 kHz to 10 kHz)	1.0 % + 400 μ A 0.3 % + 400 μ A 0.1 % + 400 μ A 0.35 % + 7 mA	
	1A to 3A (3 Hz to 5 Hz) (5 Hz to 10 Hz) (10 Hz to 5 kHz) (5 kHz to 10 kHz)	1.1 % + 1.8 mA 0.35 % + 1.8 mA 0.15 % + 1.8 mA 0.35 % + 21 mA	
	3 A to 10 A (3 Hz to 5 Hz) (5 Hz to 10 Hz) (10 Hz to 5 kHz) (5 kHz to 10 kHz)	1.1 % + 6 mA 0.35 % + 6 mA 0.15 % + 6 mA 0.35 % + 70 mA	

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (\pm)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Resistance – Measure ⁴	Up to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 kΩ 1 kΩ to 10 kΩ 10 kΩ to 100 kΩ 100 kΩ to 1 MΩ 1 MΩ to 10 MΩ 10 MΩ to 100 MΩ 100 MΩ to 1 GΩ	0.01 % + 3 mΩ 0.01 % + 4 mΩ 0.01 % + 0.01 Ω 0.01 % + 0.1 Ω 0.01 % + 1 Ω 0.01 % + 10 Ω 0.04 % + 100 Ω 0.8 % + 10 kΩ 2.0 % + 100 kΩ	WI-L-CA-Cali-1402 Fluke 8846A
Capacitance – Generate ^{3,5}	(10 Hz to 10 kHz) 0.22 nF to 0.4 nF 0.4 nF to 1.1 nF (10 Hz to 3 kHz) 1.1 nF to 3.3 nF (10 Hz to 1 kHz) 3.3 nF to 11 nF 11 nF to 33 nF 33 nF to 110 nF 110 nF to 330 nF (10 Hz to 600 Hz) 330 nF to 1.1 μF (10 Hz to 300 Hz) 1.1 μF to 3.3 μF (10 Hz to 150 Hz) 3.3 μF to 11 μF (10 Hz to 120 Hz) 11 μF to 33 μF (10 Hz to 80 Hz) 33 μF to 110 μF (0 Hz to 50 Hz) 110 μF to 330 μF (0 Hz to 20 Hz) 330 μF to 1.1 mF (0 Hz to 6 Hz) 1.1 mF to 3.3 mF (0 Hz to 2 Hz) 3.3 mF to 11 mF	0.5 % + 0.01 nF 0.5 % + 0.01 nF 0.5 % + 0.01 nF 0.25 % + 0.01 nF 0.25 % + 0.1 nF 0.25 % + 0.1 nF 0.25 % + 0.3 nF 0.25 % + 1 nF 0.25 % + 3 nF 0.25 % + 10 nF 0.40 % + 30 nF 0.45 % + 100 nF 0.45 % + 300 nF 0.45 % + 1 μF 0.45 % + 3 μF 0.45 % + 10 μF	WI-L-CA-Cali-1402 Fluke 5522A

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Capacitance – Generate ^{3,5} (cont'd.)	(0 Hz to 0.6 Hz) 11 mF to 33 mF (0 Hz to 0.2 Hz) 33 mF to 110 mF	0.75 % + 30 µF 1.1 % + 100 µF	WI-L-CA-Cali-1402 Fluke 5522A (cont'd.)
Temperature Simulation - Thermocouple Indicators & Sources			WI-L-CA-Cali-1402 Fluke 5522A
Type E	-250 °C to -100 °C -100 °C to -25 °C -25 °C to 350 °C 350 °C to 650 °C 650 °C to 1000 °C	0.5 °C 0.16 °C 0.14 °C 0.16 °C 0.21 °C	
Type J	-210 °C to -100 °C -100 °C to -30 °C -30 °C to 150 °C 150 °C to 760 °C 760 °C to 1200 °C	0.27 °C 0.16 °C 0.14 °C 0.17 °C 0.23 °C	
Type K	-200 °C to -100 °C -100 °C to -25 °C -25 °C to 120 °C 120 °C to 1000 °C 1000 °C to 1372 °C	0.33 °C 0.18 °C 0.16 °C 0.26 °C 0.4 °C	
Type N	-200 °C to -100 °C -100 °C to -25 °C -25 °C to 120 °C 120 °C to 410 °C 410 °C to 1300 °C	0.23 °C 0.23 °C 0.20 °C 0.20 °C 0.28 °C	
Type T	-250 °C to -150 °C -150 °C to 0 °C 0 °C to 120 °C 120 °C to 400 °C	0.63 °C 0.24 °C 0.16 °C 0.14 °C	
Temperature Simulation – RTD Indicators & Sources			WI-L-CA-Cali-1402 Fluke 5522A Fluke 8846A
RTD, Pt 385 (100 Ω)	-200 °C to 799 °C	0.076 °C	
RTD, Pt 385 (500 Ω)	-190 °C to 630 °C	0.11 °C	
RTD, Pt 385 (1000 Ω)	-190 °C to 630 °C	0.23 °C	

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

¹The uncertainty covered by the Calibration and Measurement Capability (CMC) is expressed as the expanded uncertainty having a coverage probability of approximately 95 %. It is the smallest measurement uncertainty that a laboratory can achieve within its scope of accreditation when performing calibrations of a best existing device. The measurement uncertainty reported on a calibration certificate may be greater than that provided in the CMC due to the behavior of the calibration item and other factors that may contribute to the uncertainty of a specific calibration.

²When uncertainty is stated in relative terms (such as percent, a multiplier expressed as a decimal fraction or in scientific notation), it is in relation to instrument reading or instrument output, as appropriate, unless otherwise indicated.

³Capability is suitable for the calibration of measuring devices in the stated ranges.

⁴Capability is suitable for the calibration of devices intended to generate the indicated quantity in the stated ranges.

⁵The actual frequency applied by the calibrator cannot be selected and may be dependent on the measurement device under calibration. Approximate frequency ranges for a given capacitance or capacitance range may be found in the Fluke 5522A's published specifications.

⁶ On Site Calibration Only