



INTERNATIONAL
ACCREDITATION
SERVICE®

CERTIFICATE OF ACCREDITATION

This is to attest that

INTERNATIONAL LABORATORIES AND AL HOTY STANGER W.L.L.

BLDG NO. 397, ROAD NO. 1507, P.O. BOX 16464
HIDD - 115 INDUSTRIAL AREA, KINGDOM OF BAHRAIN

Calibration Laboratory CL-136

has met the requirements of AC204, *IAS Accreditation Criteria for Calibration Laboratories*, and has demonstrated compliance with ISO/IEC Standard 17025:2017, *General requirements for the competence of testing and calibration laboratories*. This organization is accredited to provide the services specified in the scope of accreditation.

Effective Date November 22, 2023

Expiration Date March 1, 2026



A handwritten signature in black ink, reading "Raj Nathan".

President

Visit www.iasonline.org for current accreditation information.

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

INTERNATIONAL LABORATORIES AND AL HOTY STANGER W.L.L.

www.interlabbh.com

Contact Name Murali Pokkunuri

Contact Phone +973-17464402

Accredited to ISO/IEC 17025:2017

Effective Date November 22, 2023

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)*

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
<i>Dimensional</i>			
Micrometer-Outside	0 mm to 150 mm 0 mm to 300 mm 300 mm to 600 mm	2 µm 10 µm 20 µm	Gauge Blocks 0 Grade / Optical Flat / Optical Parallel BS EN 3611 2010
Micrometer - Inside	0 mm to 600 mm	10 µm	Internal Microchecker IS 2966
Caliper - Vernier, Dial & Electronic	0 mm to 600 mm	10 µm	Gauge Blocks 0 Grade, Internal Micro-checker IS 3651-2
Height Gauges - Vernier, Dial & Electronic	0 mm to 600 mm	10 µm	Gauge Blocks 0 Grade BS EN ISO 13225
Dial Indicator/ Gauge (Plunger)	0 mm to 25 mm	4 µm	Dial calibration tester IS 2092
Plain Plug Gauge (Go/No Go)	0 mm to 150 mm	3 µm	Digital Micrometer BS 1044-1:2008
Dial Thickness Gauge	0 mm to 50 mm	10 µm	Gauge Blocks 0 Grade ILS/CM/M/008
Bore Gauge (Only for Transmission)	0 mm to 2 mm	4 µm	Dial Calibration Tester ILS/CM/M/018
Feeler Gauge	0 mm to 2 mm	3.5 µm	Digital Micrometer IS 3179
Pin Gauge	0 mm to 25 mm	3 µm	Digital Micrometer IS 11103
Bevel Protector	0° to 360°	4'	Sine Bar / Gauge Blocks IS 2102-1
Combination Set	0° to 180°	45'	Sine Bar / Gauge Blocks IS 2102-1

* If information in this CMC is presented in non-SI units, the conversion factors stated in NIST Special Publication 811 "Guide for the Use of the International System of Units (SI)" apply.

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Dial Calibration Tester	0 mm to 25 mm	1.5 µm	Electronic Comparator with Probe ILS/CM/M/019
Test Sieve	63 µm to 2 mm	0.4 %	Master Sieves, Glass Beads & Weighing Balances BS1377-1:1990
	2 mm to 125 mm	30 µm	Digital Caliper BS ISO 3310-2:2013 BS 410-1:2000
Volumetric Glassware	5 µL to 1000 µL 1 mL to 10 mL 10 mL to 50 mL 50 mL to 100 mL 100 mL to 500 mL 500 mL to 1000 mL 1000 mL to 2000 mL	1.3 µL 3 µL 0.06 mL 0.6 mL 1 mL 3 mL 6 mL	OIML Class E2 Weighing Balances & distilled water by gravimetric method IS 8897
Mass	1 mg to 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg 20 kg	0.17 mg 0.18 mg 0.18 mg 0.18 mg 0.18 mg 0.16 mg 1.7 mg 10 mg 13 mg 18 mg 160 mg 160 mg	Standard weight of E2 Class & Weighing Balances IS 9865
Weighing Scale and Balances ⁵	1 mg to 200 g 200 g to 1 kg 1 kg to 4 kg 4 kg to 50 kg	1.5 mg 1.7 mg 1.0 mg 0.15 g	Standard weight of E2 Class IS 9440
Hydraulic - Pressure Gauges	1 bar to 1000 bar	0.6 %	Reference Test Gauge by comparison method BS EN 837-1:1998
Vacuum Gauge	-1 mmHg to -710 mmHg	0.5 %	Reference Test Gauge by comparison method IS 3624
Compression Testing Machine ⁵	1 kN to 3000 kN	0.25 %	Load Cell BS EN ISO 376:2011 BS EN 12390-4:2000

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Thermal			
RTD / Thermocouples with or without Indicators & Temperature Gauges	-25 °C to 50 °C 50 °C to 600 °C 600 °C to 1100 °C	0.96 °C 2.8 °C 2.3 °C	Temperature Baths, Reference PRT or S/K Type TC Sensor and Universal Calibrator – by Comparison Method IS 7358 / IS 6274
Digital Thermometers	-25 °C to 50 °C 50 °C to 600 °C 600 °C to 1100 °C	0.96 °C 2.8 °C 2.3 °C	
Mechanical (Dial) Thermometers	0 °C to 1200 °C	1.4 °C	
Liquid in Glass Thermometers	-25 °C to 50 °C 50 °C to 250 °C	0.52 °C 0.78 °C	Temperature Bath Reference PRT and Universal Calibrator - Partial immersion only IS 6274
Resistance Thermometers	-30 °C to 50 °C 50 °C to 600 °C 250 °C to 1200 °C	0.52 °C 0.78 °C 1.3 °C	Temperature Bath, Reference PRT and Fluke 8846A Multimeter UNICAL 3100 Manufacturing Manual
Controller/Indicator of Ovens, Incubator, Muffle Furnace, Refrigerator ⁵	-10 °C to 50 °C 50 °C to 600 °C 600 °C to 1200 °C	0.60 °C 0.60 °C 2.3 °C	Reference PRT/ S Type Thermocouple with Universal Calibrator by single point Calibration (at measuring location in DUC) EURAMET cg-13
Electrical – DC/LF			
DC Voltage – Source ³	1 mV to 33 V 33 V to 1000 V	0.013 % 0.0095 %	Fluke 5500E Calibrator by Direct Method EURAMET cg-15 Fluke Manufacturing
DC Voltage -Measure ⁴	10 mV to 100 V 100 V to 1000 V	0.045 % 0.06 %	Fluke 8846A DMM: NICT Journal-Vol 63 No 1-02-03
AC Voltage - Source ³ @ 50 Hz	10 mV to 330 mV 330 mV to 1000 V	0.27 % 0.09 %	Fluke 5500E Calibrator by Direct Method EURAMET cg-15, Fluke Manufacturing
AC Voltage - Measure ⁴ @ 50 Hz	100 mV to 1000 V	0.08 %	Fluke 8846A DMM by Direct Method NICT Journal-Vol 63 No 1-02-03

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
DC Current - Source ³	1 mA to 330 mA 330 mA to 10 A 10 A to 550 A	0.01 % 0.08 % 0.6 %	Fluke 5500E Calibrator & 50 Turn Coil by Direct method EURAMET cg-15, Fluke Manufacturing
DC Current - Measure ⁴	1 µA to 100 µA 0.1 mA to 10 mA 10 mA to 3 A 3 A to 10 A	0.012 % 0.08 % 0.002 % 0.02 %	Fluke 8846A DMM by Direct Method. NICT Journal-Vol 63 No 1-02- 03 Fluke Manufacturing
AC Current - Source ³ @ 50 Hz	30 µA to 330 mA 0.33 A to 10 A 10 A to 550 A	0.3 % 0.06 % 0.05 %	Fluke 5500E Calibrator & 50 Turn Coil by Direct method EURAMET cg-15, Fluke Manufacturing
AC Current - Measure ⁴ @ 50 Hz	30 µA to 100 µA 100 µA to 1 mA 1 mA to 10 A	0.065 % 0.019 % 0.038 %	Fluke 8846A DMM by Direct Method NICT Journal-Vol 63 No 1-02- 03
DC Resistance - Source ³	0.01 Ω to 32.999 Ω 33 Ω to 32.9999 kΩ 33 kΩ to 3.29999 MΩ 3.3 MΩ to 10.9999 MΩ 11 MΩ to 32.9999 MΩ 33 MΩ to 329.999 MΩ	0.012 % 0.009 % 0.015 % 0.06 % 0.12 % 0.5 %	Fluke 5500E Calibrator by Direct method EURAMET cg-15
Resistance Measure ⁴	1 Ω to 10 Ω 10 Ω to 100 Ω 0.1 kΩ to 10 MΩ 10 MΩ to 100 MΩ 100 MΩ to 1 GΩ	0.04 % 0.014 % 0.011 % 0.041 % 0.8 %	Fluke 8846A DMM by Direct Method: NICT Journal-Vol 63 No 1-02-03 Fluke Manufacturing
Capacitance - Source ^{3,6}	0.33 nF to 11 nF 11 nF to 1.1 µF 1.1 µF to 33 µF 33 µF to 110 µF 110 µF to 330 µF 330 µF to 1.1 mF	0.5 % 0.3 % 0.4 % 0.5 % 0.7 % 1 %	Fluke 5500E Calibrator by Direct Method EURAMET cg-15, Fluke Manufacturing
Capacitance – Measure ⁴	0.01 nF to 1 nF 1 nF to 10 mF 10 mF to 100 mF	4.5 % 1.5 % 4.2 %	Fluke 8846A Multimeter by Direct Method: NICT Journal- Vol 63 No 1-02-03

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Temperature Simulation Thermocouples			Fluke 5500E Calibrator by Direct Method EURAMET cg-11, Fluke Manufacturing
Type B	600 °C to 1820 °C	0.44 °C	
Type C	0 °C to 2316 °C	0.84 °C	
Type E	-250 °C to 1000 °C	0.27 °C	
Type J	-210 °C to 1200 °C	0.27 °C	
Type K	-200 °C to 1372 °C	0.40 °C	
Type L	-200 °C to 900 °C	0.37 °C	
Type N	-200 °C to 1300 °C	0.40 °C	
Type R	0 °C to 1767 °C	0.57 °C	
Type S	0 °C to 1767 °C	0.47 °C	
Type T	-250 °C to 400 °C	0.63 °C	
Type U	-200 °C to 600 °C	0.27 °C	
RTDs			
Pt 385, 100 Ω	-200 °C to 800 °C	0.23 °C	
Pt 3926, 100 Ω	-200 °C to 630 °C	0.12 °C	
Pt 3916, 100 Ω	-200 °C to 630 °C	0.25 °C	
Pt 385, 200 Ω	-200 °C to 630 °C	0.16 °C	
Pt 385, 500 Ω	-200 °C to 630 °C	0.11 °C	
Pt 385 1000 Ω	-200 °C to 630 °C	0.23 °C	
Pt 385, 120 Ω	-80 °C to 260 °C	0.14 °C	
U 427, 10 Ω	-100 °C to 260 °C	0.3 °C	
Time and Frequency			
Frequency – Source ³	2 Hz to 1 MHz	2.9 %	Fluke 5500E Calibrator by Direct Method ILS/CM/E/022
Frequency – Measure ⁴	3 Hz to 1 MHz	1.5 %	Fluke 8846A Multimeter by Direct Method ILS/CM/E/022
Stopwatches and Timers	5 s to 1 h 1 h to 24 h	1.0 s 2.1 s	Stopwatch by Comparison Method ILS/CM/E/011
Chemical			
pH meters	4.0 pH 7.0 pH 10.2 pH	0.03 pH 0.03 pH 0.03 pH	Standard Buffer Solutions ILS/CM/E/014
Conductivity Meters	3.30 mS to 100 mS	0.5 %	Using conductivity buffer solutions: Direct method ILS/CM/E/024

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

Notes:

¹The uncertainty covered by the Calibration and Measurement Capability (CMC) is expressed as the expanded uncertainty having a coverage probability of approximately 95 %. It is the smallest measurement uncertainty that a laboratory can achieve within its scope of accreditation when performing calibrations of a best existing device. The measurement uncertainty reported on a calibration certificate may be greater than that provided in the CMC due to the behavior of the calibration item and other factors that may contribute to the uncertainty of a specific calibration.

²When uncertainty is stated in relative terms (such as percent, a multiplier expressed as a decimal fraction or in scientific notation), it is in relation to instrument reading or instrument output, as appropriate, unless otherwise indicated.

³Capability is suitable for the calibration of measuring devices in the stated ranges.

⁴Capability is suitable for the calibration of devices intended to generate the indicated quantity in the stated ranges.

⁵Also available as site calibration. Note that actual measurement uncertainties achievable at a customer's site can normally be expected to be larger than the uncertainties listed on this Scope of Accreditation.

⁶The actual frequency applied by the calibrator cannot be selected and may be dependent on the measurement device under calibration. Approximate frequency ranges for a given capacitance or capacitance range may be found in the calibrator's published specifications.

DUC = device under calibration