



# CERTIFICATE OF ACCREDITATION

*This is to attest that*

**UL LLC**

1285 WALT WHITMAN ROAD  
MELVILLE, NEW YORK 11747, U.S.A.

**Calibration Laboratory CL-166**

has met the requirements of AC204, *IAS Accreditation Criteria for Calibration Laboratories*, and has demonstrated compliance with ISO/IEC Standard 17025:2017, *General requirements for the competence of testing and calibration laboratories*. This organization is accredited to provide the services specified in the scope of accreditation.

Effective Date December 31, 2022

Expiration Date May 1, 2025



A handwritten signature in black ink that reads 'Raj Nathan'.

**President**

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

**UL LLC**

[www.ul.com](http://www.ul.com)

**Contact Name** Scott Cimino

**Contact Phone** +1-631-546-2518

*Accredited to ISO/IEC 17025:2017*

*Effective Date December 31, 2022*

## CALIBRATION AND MEASUREMENT CAPABILITY (CMC)\*

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
<b>Dimensional</b>			
Calipers	0 in to 6 in	290 µin	Grade '0' inch gage blocks CP000010
Micrometers	0 in to 1 in	59 µin	Grade '0' inch gage blocks CP000010
<b>Thermal</b>			
Electrical Simulation of Thermocouple - Generate <sup>3</sup> Type J	-210 °C to -100 °C	0.22 °C	Fluke 5522A VP000044CMC
	-100 °C to -30 °C	0.14 °C	
	-30 °C to 150 °C	0.12 °C	
Type K	150 °C to 760 °C	0.14 °C	
	760 °C to 1200 °C	0.19 °C	
	-200 °C to -100 °C	0.26 °C	
	-100 °C to -25 °C	0.15 °C	
Type T	-25 °C to 120 °C	0.14 °C	
	120 °C to 1000 °C	0.21 °C	
	1000 °C to 1372 °C	0.66 °C	
	-250 °C to -150 °C	0.39 °C	
	-150 °C to 0 °C	0.19 °C	
Temperature Generate <sup>3</sup>	0 °C to 120 °C	0.25 °C	Hart Scientific 9100S Dry Block CP000054
	120 °C to 400 °C	0.25 °C	
	35 °C to 375 °C	1.0 x 10 <sup>-3</sup> + 0.35 °C	
	0 °C to 60 °C	0.27 °C	Rotronic HG2-S CP000129

\* If information in this CMC is presented in non-SI units, the conversion factors stated in NIST Special Publication 811 "Guide for the Use of the International System of Units (SI)" apply.

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Temperature Measure <sup>4</sup>	-50 °C to 0 °C 0 °C to 125 °C 125 °C to 300 °C	1.3 °C 0.66 °C 1.4 °C	Fluke 726 W/Type T probe VP000045CMC
Humidity Generate <sup>3</sup>	5 %RH to 65 %RH at 23 °C  65 %RH to 95 %RH at 23 °C	1.5 %RH  1.8 %RH	Rotronic HG2-S CP000129
Humidity Measure <sup>4</sup>	5 %RH to 95 %RH (10 °C to 30 °C)  5 %RH to 95 %RH (30 °C to 60 °C)	1.5 %RH  1.8 %RH	Rotronic HL-NT w/ Hygroclip 2 Probe CP000015
<b>Electrical – DC/LF</b>			
DC Voltage Generate <sup>3</sup>	0 mV to 329.9999 mV 0 V to 3.299999 V 0 V to 32.99999 V 30 V to 329.9999 V 100 V to 1020.000 V	$1.6 \times 10^{-5} + 0.78 \mu\text{V}$ $8.5 \times 10^{-6} + 1.6 \mu\text{V}$ $9.3 \times 10^{-6} + 16 \mu\text{V}$ $1.4 \times 10^{-5} + 0.12 \text{ mV}$ $1.4 \times 10^{-5} + 1.2 \text{ mV}$	Fluke 5522A VP000040CMC
AC Voltage Generate <sup>3</sup>	1.0 mV to 32.999 mV (0.01 Hz to 9.99 Hz) (10 Hz to 45 Hz) (45 Hz to 10 kHz) (10 kHz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 100 kHz) (100 kHz to 500 kHz)  33 mV to 329.999 mV (0.01 Hz to 9.99 Hz) (10 Hz to 45 Hz) (45 Hz to 10 kHz) (10 kHz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 100 kHz) (100 kHz to 500 kHz)  0.33 V to 3.29999 V (0.01 Hz to 9.99 Hz) (10 Hz to 45 Hz) (45 Hz to 10 kHz) (10 kHz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 100 kHz) (100 kHz to 500 kHz)	$3.9 \times 10^{-2} + 0.13 \text{ mV}$ $6.2 \times 10^{-4} + 4.7 \mu\text{V}$ $1.2 \times 10^{-4} + 4.7 \mu\text{V}$ $1.6 \times 10^{-4} + 4.7 \mu\text{V}$ $7.8 \times 10^{-4} + 4.7 \mu\text{V}$ $2.7 \times 10^{-3} + 9.3 \mu\text{V}$ $6.2 \times 10^{-3} + 39 \mu\text{V}$  $3.9 \times 10^{-2} + 1.3 \text{ mV}$ $2.3 \times 10^{-4} + 6.2 \mu\text{V}$ $1.1 \times 10^{-4} + 6.2 \mu\text{V}$ $1.2 \times 10^{-4} + 6.2 \mu\text{V}$ $2.7 \times 10^{-4} + 6.2 \mu\text{V}$ $6.2 \times 10^{-4} + 25 \mu\text{V}$ $1.6 \times 10^{-3} + 54 \mu\text{V}$  $3.9 \times 10^{-2} + 13 \text{ mV}$ $2.3 \times 10^{-4} + 39 \mu\text{V}$ $1.2 \times 10^{-4} + 47 \mu\text{V}$ $1.5 \times 10^{-4} + 47 \mu\text{V}$ $2.3 \times 10^{-4} + 39 \mu\text{V}$ $5.4 \times 10^{-4} + 97 \mu\text{V}$ $1.9 \times 10^{-3} + 0.47 \text{ mV}$	Fluke 5522A VP000040CMC

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
AC Voltage Generate <sup>3</sup> (continued)	3.3 V to 32.9999 V (0.01 Hz to 9.99 Hz) (10 Hz to 45 Hz) (45 Hz to 10 kHz) (10 kHz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 100 kHz)	3.9 x 10 <sup>-2</sup> + 0.13 V 2.3 x 10 <sup>-4</sup> + 0.5 mV 1.2 x 10 <sup>-4</sup> + 0.47 mV 1.9 x 10 <sup>-4</sup> + 0.47 mV 2.7 x 10 <sup>-4</sup> + 0.47 mV 7.0 x 10 <sup>-4</sup> + 1.2 mV	Fluke 5522A VP000040CMC
	33 V to 329.999 V (45 Hz to 1 kHz) (1 kHz to 10 kHz) (10 kHz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 100 kHz)	1.5 x 10 <sup>-4</sup> + 1.6 mV 1.6 x 10 <sup>-4</sup> + 4.7 mV 1.9 x 10 <sup>-4</sup> + 4.7 mV 2.3 x 10 <sup>-4</sup> + 4.7 mV 1.6 x 10 <sup>-3</sup> + 39 mV	
	330 V to 1020 V (45 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz)	2.3 x 10 <sup>-4</sup> + 7.8 mV 1.9 x 10 <sup>-4</sup> + 7.8 mV 2.3 x 10 <sup>-4</sup> + 7.8 mV	
DC Current Generate <sup>3</sup>	0 µA to 329.999 µA 0 mA to 3.29999 mA 0 mA to 32.9999 mA 0 mA to 329.999 mA 0 A to 1.09999 A 1.1 A to 2.99999 A 0 A to 10.9999 A 11 A to 20.5 A	1.2 x 10 <sup>-4</sup> + 16 nA 7.8 x 10 <sup>-5</sup> + 39 nA 7.8 x 10 <sup>-5</sup> + 0.19 µA 7.8 x 10 <sup>-5</sup> + 1.9 µA 1.6 x 10 <sup>-4</sup> + 31 µA 2.9 x 10 <sup>-4</sup> + 31 µA 3.9 x 10 <sup>-4</sup> + 0.39 mA 7.8 x 10 <sup>-4</sup> + 0.58 mA	Fluke 5522A VP000040CMC
	20 A to 120 A	7.8 x 10 <sup>-5</sup> + 4.7 mA	Fluke 52120A (Sourced by 5522A) CP000068
AC Current Generate <sup>3</sup>	29.00 µA to 329.99 µA (10 Hz to 20 Hz) (20 Hz to 45 Hz) (45 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz) (10 kHz to 30 kHz)	1.6 x 10 <sup>-3</sup> + 78 nA 1.2 x 10 <sup>-3</sup> + 78 nA 9.7 x 10 <sup>-4</sup> + 78 nA 2.3 x 10 <sup>-3</sup> + 0.12 µA 6.2 x 10 <sup>-3</sup> + 0.16 µA 1.2 x 10 <sup>-2</sup> + 0.31 µA	Fluke 5522A VP000040CMC
	0.33 mA to 3.29999 mA (10 Hz to 20 Hz) (20 Hz to 45 Hz) (45 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz) (10 kHz to 30 kHz)	1.6 x 10 <sup>-3</sup> + 0.12 µA 9.7 x 10 <sup>-4</sup> + 0.12 µA 7.8 x 10 <sup>-4</sup> + 0.12 µA 1.6 x 10 <sup>-3</sup> + 0.16 µA 3.9 x 10 <sup>-3</sup> + 0.23 µA 7.8 x 10 <sup>-3</sup> + 0.47 µA	

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)	
AC Current Generate <sup>3</sup> (continued)	3.3 mA to 32.9999 mA (10 Hz to 20 Hz) (20 Hz to 45 Hz) (45 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz) (10 kHz to 30 kHz)	1.4 x 10 <sup>-3</sup> + 1.6 μA 7.0 x 10 <sup>-4</sup> + 1.6 μA 3.1 x 10 <sup>-4</sup> + 1.6 μA 6.2 x 10 <sup>-4</sup> + 1.6 μA 1.6 x 10 <sup>-3</sup> + 2.3 μA 3.1 x 10 <sup>-3</sup> + 3.1 μA	Fluke 5522A VP000040CMC	
	33 mA to 329.999 mA (10 Hz to 20 Hz) (20 Hz to 45 Hz) (45 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz) (10 kHz to 30 kHz)	1.4 x 10 <sup>-3</sup> + 16 μA 7.0 x 10 <sup>-4</sup> + 16 μA 3.1 x 10 <sup>-4</sup> + 16 μA 7.8 x 10 <sup>-4</sup> + 39 μA 1.6 x 10 <sup>-3</sup> + 78 μA 3.1 x 10 <sup>-3</sup> + 0.16 mA		
	0.33 A to 1.09999 A (10 Hz to 45 Hz) (45 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz)	1.4 x 10 <sup>-3</sup> + 78 μA 3.9 x 10 <sup>-4</sup> + 78 μA 4.7 x 10 <sup>-3</sup> + 0.78 mA 1.9 x 10 <sup>-2</sup> + 3.9 mA		
	1.1 A to 2.99999 A (10 Hz to 45 Hz) (45 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz)	1.4 x 10 <sup>-3</sup> + 78 μA 4.7 x 10 <sup>-4</sup> + 78 μA 4.7 x 10 <sup>-3</sup> + 0.78 mA 1.9 x 10 <sup>-2</sup> + 3.9 mA		
	3 A to 10.9999 A (45 Hz to 100 Hz) (100 Hz to 1 kHz) (1 kHz to 5 kHz)	4.7 x 10 <sup>-4</sup> + 1.6 mA 7.8 x 10 <sup>-4</sup> + 1.6 mA 2.3 x 10 <sup>-2</sup> + 1.6 mA		
	11 A to 20.5 A (45 Hz to 100 Hz) (100 Hz to 1 kHz) (1 kHz to 5 kHz)	9.3 x 10 <sup>-4</sup> + 3.9 mA 1.2 x 10 <sup>-3</sup> + 3.9 mA 2.3 x 10 <sup>-2</sup> + 3.9 mA		
	20 A to 120 A (50/60 Hz)	1.2 x 10 <sup>-4</sup> + 0.47 A		
	0 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V	5.0 x 10 <sup>-6</sup> + 0.1 μV 3.5 x 10 <sup>-6</sup> + 0.4 μV 3.5 x 10 <sup>-6</sup> + 4 μV 5.5 x 10 <sup>-6</sup> + 40 μV 5.5 x 10 <sup>-6</sup> + 0.5 mV		Fluke 8508A VP000040CMC



# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
DC Voltage Measure <sup>4</sup> continued	6 kV to 30 kV	0.12 %	Ross Engineering VD60-6.2Y-A-LB-AL CP000678
DC Current Measure <sup>4</sup>	0 µA to 200 µA 200 µA to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 20 A	1.2 x 10 <sup>-5</sup> + 0.4 nV 1.2 x 10 <sup>-5</sup> + 4 nV 1.4 x 10 <sup>-5</sup> + 40 nV 4.8 x 10 <sup>-5</sup> + 0.8 µV 1.9 x 10 <sup>-4</sup> + 16 µV 4.0 x 10 <sup>-4</sup> + 0.4 mV	Fluke 8508A VP000040CMC
AC Voltage Measure <sup>4</sup>	0 mV to 200 mV (1 Hz to 10 Hz) (10 Hz to 40 Hz) (40 Hz to 100 Hz) (100 Hz to 2 kHz) (2 kHz to 10 kHz) (10 kHz to 30 kHz) (30 kHz to 100 kHz)  200 mV to 2 V (1 Hz to 10 Hz) (10 Hz to 40 Hz) (40 Hz to 100 Hz) (100 Hz to 2 kHz) (2 kHz to 10 kHz) (10 kHz to 30 kHz) (30 kHz to 100 kHz) (100 kHz to 300 kHz) (300 kHz to 1 MHz)  2 V to 20 V (1 Hz to 10 Hz) (10 Hz to 40 Hz) (40 Hz to 100 Hz) (100 Hz to 2 kHz) (2 kHz to 10 kHz) (10 kHz to 30 kHz) (30 kHz to 100 kHz) (100 kHz to 300 kHz) (300 kHz to 1 MHz)  20 V to 200 V (1 Hz to 10 Hz) (10 Hz to 40 Hz) (40 Hz to 100 Hz) (100 Hz to 2 kHz) (2 kHz to 10 kHz) (10 kHz to 30 kHz)	1.7 x 10 <sup>-4</sup> + 14 µV 1.4 x 10 <sup>-4</sup> + 4 µV 1.2 x 10 <sup>-4</sup> + 4 µV 1.1 x 10 <sup>-4</sup> + 2 µV 1.4 x 10 <sup>-4</sup> + 4 µV 3.4 x 10 <sup>-4</sup> + 8 µV 7.7 x 10 <sup>-4</sup> + 20 µV  1.5 x 10 <sup>-4</sup> + 0.12 mV 1.2 x 10 <sup>-4</sup> + 20 µV 9.0 x 10 <sup>-5</sup> + 20 µV 7.5 x 10 <sup>-5</sup> + 20 µV 1.1 x 10 <sup>-4</sup> + 20 µV 2.2 x 10 <sup>-4</sup> + 40 µV 5.7 x 10 <sup>-4</sup> + 0.2 mV 3.0 x 10 <sup>-3</sup> + 2 mV 1.0 x 10 <sup>-2</sup> + 20 mV  1.5 x 10 <sup>-4</sup> + 1.2 mV 1.2 x 10 <sup>-4</sup> + 0.2 mV 9.0 x 10 <sup>-5</sup> + 0.2 mV 7.5 x 10 <sup>-5</sup> + 0.2 mV 1.1 x 10 <sup>-4</sup> + 0.2 mV 2.2 x 10 <sup>-4</sup> + 0.4 mV 5.7 x 10 <sup>-4</sup> + 2 mV 3.0 x 10 <sup>-3</sup> + 20 mV 1.0 x 10 <sup>-2</sup> + 0.2 V  1.5 x 10 <sup>-4</sup> + 12 mV 1.2 x 10 <sup>-4</sup> + 2 mV 9.0 x 10 <sup>-5</sup> + 2 mV 7.5 x 10 <sup>-5</sup> + 2 mV 1.1 x 10 <sup>-4</sup> + 2 mV 2.2 x 10 <sup>-4</sup> + 4 mV	Fluke 8508A VP000040CMC

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
AC Voltage Measure <sup>4</sup> continued	20 V to 200 V (30 kHz to 100 kHz) (100 kHz to 300 kHz) (300 kHz to 1 MHz)  200 V to 1000 V (1 Hz to 10 Hz) (10 Hz to 40 Hz) (40 Hz to 10 kHz) (10 kHz to 30 kHz) (30 kHz to 100 kHz)  4 kV to 30 kV	5.7 x 10 <sup>-4</sup> + 20 mV 3.0 x 10 <sup>-3</sup> + 0.2 V 1.0 x 10 <sup>-2</sup> + 2 V  1.5 x 10 <sup>-4</sup> + 70 mV 1.2 x 10 <sup>-4</sup> + 20 mV 1.2 x 10 <sup>-4</sup> + 20 mV 2.3 x 10 <sup>-4</sup> + 40 mV 5.8 x 10 <sup>-4</sup> + 0.2 V  0.59 %	Fluke 8508A VP000040CMC      Ross Engineering VD60-6.2Y-A-LB-AL CP000678
AC Current Measure <sup>4</sup>	0 µA to 200 µA (1 Hz to 10 Hz) (10 Hz to 10 kHz) (10 kHz to 30 kHz) (30 kHz to 100 kHz)  200 µA to 2 mA (1 Hz to 10 Hz) (10 Hz to 10 kHz) (10 kHz to 30 kHz) (30 kHz to 100 kHz)  200 µA to 20 mA (1 Hz to 10 Hz) (10 Hz to 10 kHz) (10 kHz to 30 kHz) (30 kHz to 100 kHz)  20 mA to 200 mA (1 Hz to 10 Hz) (10 Hz to 10 kHz) (10 kHz to 30 kHz)  200 mA to 2 A (10 Hz to 2 kHz) (2 kHz to 10 kHz) (10 kHz to 30 kHz)  2 A to 20 A (10 Hz to 2 kHz) (2 kHz to 10 kHz)	5.0 x 10 <sup>-4</sup> + 20 nA 5.0 x 10 <sup>-4</sup> + 20 nA 7.1 x 10 <sup>-4</sup> + 20 nA 4.0 x 10 <sup>-3</sup> + 20 nA  3.1 x 10 <sup>-4</sup> + 0.2 µA 3.0 x 10 <sup>-4</sup> + 0.2 µA 7.1 x 10 <sup>-4</sup> + 0.2 µA 4.0 x 10 <sup>-3</sup> + 0.2 µA  3.1 x 10 <sup>-4</sup> + 2 µA 3.0 x 10 <sup>-4</sup> + 2 µA 7.1 x 10 <sup>-4</sup> + 2 µA 4.0 x 10 <sup>-3</sup> + 2 µA  3.1 x 10 <sup>-4</sup> + 20 µA 2.9 x 10 <sup>-4</sup> + 20 µA 6.3 x 10 <sup>-4</sup> + 20 µA  6.2 x 10 <sup>-4</sup> + 0.2 mA 7.4 x 10 <sup>-4</sup> + 0.2 mA 3.0 x 10 <sup>-3</sup> + 0.2 mA  8.2 x 10 <sup>-4</sup> + 2 mA 2.5 x 10 <sup>-3</sup> + 2 mA	Fluke 8508A VP000040CMC

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Resistance Generate <sup>3</sup>	0 Ω to 10.9999 Ω 11 Ω to 32.9999 Ω 33 Ω to 109.9999 Ω 110 Ω to 329.9999 Ω 330 Ω to 1.099999 kΩ 1.1 kΩ to 3.299999 kΩ 3.3 kΩ to 10.99999 kΩ 11 kΩ to 32.99999 kΩ 33 kΩ to 109.9999 kΩ 110 kΩ to 329.99999 kΩ 330 kΩ to 1.099999 MΩ 1.1 MΩ to 3.299999 MΩ 3.3 MΩ to 10.99999 MΩ 11 MΩ to 32.99999 MΩ 33 MΩ to 109.9999 MΩ 110 MΩ to 329.9999 MΩ 330 MΩ to 1100 MΩ	$3.1 \times 10^{-5} + 0.78 \text{ m}\Omega$ $2.3 \times 10^{-5} + 1.2 \text{ m}\Omega$ $2.2 \times 10^{-5} + 1.1 \text{ m}\Omega$ $2.2 \times 10^{-5} + 1.6 \text{ m}\Omega$ $2.2 \times 10^{-5} + 1.6 \text{ m}\Omega$ $2.2 \times 10^{-5} + 16 \text{ m}\Omega$ $2.2 \times 10^{-5} + 16 \text{ m}\Omega$ $2.2 \times 10^{-5} + 0.16 \Omega$ $2.2 \times 10^{-5} + 0.16 \Omega$ $2.5 \times 10^{-5} + 1.6 \Omega$ $2.5 \times 10^{-5} + 1.6 \Omega$ $4.7 \times 10^{-5} + 23 \Omega$ $1.0 \times 10^{-4} + 39 \Omega$ $1.9 \times 10^{-4} + 1.9 \text{ k}\Omega$ $3.9 \times 10^{-4} + 2.3 \text{ k}\Omega$ $2.3 \times 10^{-3} + 78 \text{ k}\Omega$ $1.2 \times 10^{-2} + 0.39 \text{ M}\Omega$	Fluke 5522A VP000040CMC
Resistance Measure <sup>4</sup>	0 Ω to 2 Ω 2 Ω to 20 Ω 20 Ω to 200 Ω 200 Ω to 2 kΩ 2 kΩ to 20 kΩ 20 kΩ to 200 kΩ 200 kΩ to 2 MΩ 2 MΩ to 20 MΩ 20 MΩ to 200 MΩ 200 MΩ to 2 GΩ	$1.7 \times 10^{-5} + 4 \mu\Omega$ $9.5 \times 10^{-6} + 14 \mu\Omega$ $8.0 \times 10^{-6} + 50 \mu\Omega$ $8.0 \times 10^{-6} + 0.5 \text{ M}\Omega$ $8.0 \times 10^{-6} + 5 \text{ M}\Omega$ $8.0 \times 10^{-6} + 50 \text{ M}\Omega$ $9.0 \times 10^{-6} + 1 \Omega$ $2.0 \times 10^{-5} + 0.1 \text{ k}\Omega$ $1.2 \times 10^{-4} + 10 \text{ k}\Omega$ $1.5 \times 10^{-3} + 1 \text{ M}\Omega$	Fluke 8508A VP000040CMC
Capacitance Generate <sup>3,5</sup>	220 pF to 399.9 pF 0.4 nF to 1.0999 nF 1.1 nF to 3.2999 nF 3.3 nF to 10.9999 nF 11 nF to 32.9999 nF 33 nF to 109.999 nF 110 nF to 329.999 nF 0.33 μF to 1.09999 μF 1.1 μF to 3.29999 μF 3.3 μF to 10.9999 μF 11 μF to 32.9999 μF 33 μF to 109.999 μF 110 μF to 329.999 μF 0.33 mF to 1.09999 mF 1.1 mF to 3.29999 mF 3.3 mF to 10.9999 mF 11 mF to 32.9999 mF 33 mF to 110 mF	$3.9 \times 10^{-3} + 7.8 \text{ pF}$ $3.9 \times 10^{-3} + 7.8 \text{ pF}$ $3.9 \times 10^{-3} + 7.8 \text{ pF}$ $1.9 \times 10^{-3} + 7.8 \text{ pF}$ $1.9 \times 10^{-3} + 78 \text{ pF}$ $1.9 \times 10^{-3} + 78 \text{ pF}$ $1.9 \times 10^{-3} + 0.23 \text{ nF}$ $1.9 \times 10^{-3} + 0.78 \text{ nF}$ $1.9 \times 10^{-3} + 2.3 \text{ nF}$ $1.9 \times 10^{-3} + 7.8 \text{ nF}$ $3.1 \times 10^{-3} + 23 \text{ nF}$ $3.5 \times 10^{-3} + 78 \text{ nF}$ $3.5 \times 10^{-3} + 0.23 \mu\text{F}$ $3.5 \times 10^{-3} + 0.78 \mu\text{F}$ $3.5 \times 10^{-3} + 2.3 \mu\text{F}$ $3.5 \times 10^{-3} + 7.8 \mu\text{F}$ $5.8 \times 10^{-3} + 23 \mu\text{F}$ $8.5 \times 10^{-3} + 78 \mu\text{F}$	Fluke 5522A VP000043CMC



# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
DC Power Generate <sup>3</sup>	33 mV to 1020 V (0.33 mA to 329.99 mA) (0.33 A to 2.9999 A) (3 A to 20.5 A)	(relative to output in Watts) $1.8 \times 10^{-4}$ $1.7 \times 10^{-4}$ $5.4 \times 10^{-4}$	Fluke 5522A VP000041CMC
AC Power – Generate <sup>3</sup> (45 Hz to 65 Hz)	33 mV to 329.999 mV (3.3 mA to 8.999 mA) (9 mA to 32.999 mA) (33 mA to 89.99 mA) (90 mA to 329.99 mA) (0.33 A to 0.8999 A) (0.9 A to 2.1999 A) (2.2 A to 4.4999 A) (4.5 A to 20.5 A)	(relative to output in Watts) $1.1 \times 10^{-3}$ $7.8 \times 10^{-4}$ $1.1 \times 10^{-3}$ $7.8 \times 10^{-4}$ $1.0 \times 10^{-3}$ $8.5 \times 10^{-4}$ $1.0 \times 10^{-3}$ $8.5 \times 10^{-4}$	Fluke 5522A VP000041CMC
	330 mV to 1020 V (3.3 mA to 8.999 mA) (9 mA to 32.999 mA) (33 mA to 89.99 mA) (90 mA to 329.99 mA) (0.33 A to 0.8999 A) (0.9 A to 2.1999 A) (2.2 A to 4.4999 A) (4.5 A to 20.5 A)	$9.3 \times 10^{-4}$ $6.2 \times 10^{-4}$ $9.3 \times 10^{-4}$ $6.2 \times 10^{-4}$ $8.5 \times 10^{-4}$ $7.0 \times 10^{-4}$ $9.3 \times 10^{-4}$ $7.8 \times 10^{-4}$	
Input Impedance	2 Ω to 100 kΩ (20 Hz to 100 kHz)	$1.2 \times 10^{-3}$	Wayne Kerr 6440B CP000072
	30 Ω to 16 kΩ (100 kHz to 1 MHz)	$2.3 \times 10^{-3}$	
Transfer Impedance	0.345 Ω to 500 Ω (nominal) (20 Hz to 1 MHz)	$2.4 \times 10^{-2}$	Wayne Kerr 6440B Keysight 33220A Tektronix TBS 1202B CP000072
<b>Time and Frequency</b>			
Frequency Generate <sup>3</sup>	0.01 Hz to 119.99 Hz 120.0 Hz to 1199.9 Hz 1.200 kHz to 11.999 kHz 12.00 kHz to 119.99 kHz 120.0 kHz to 1199.9 kHz 1.200 MHz to 2.000 MHz	$7.5 \times 10^{-6} + 3.9 \mu\text{Hz}$ $7.5 \times 10^{-6} + 3.9 \mu\text{Hz}$ $7.5 \times 10^{-6} + 3.9 \mu\text{Hz}$ $7.5 \times 10^{-6} + 3.9 \mu\text{Hz}$ $7.5 \times 10^{-6} + 3.9 \mu\text{Hz}$ $7.5 \times 10^{-6} + 3.9 \mu\text{Hz}$	Fluke 5522A VP000040CMC
Frequency Measure <sup>4</sup>	10 Hz to 19.52 Hz 19.53 Hz to 195.2 Hz 195.3 Hz to 1.952 kHz 1.953 kHz to 19.52 kHz 19.53 kHz to 195.2 kHz 195.3 kHz to 1 MHz	$1.2 \times 10^{-5} + 20 \mu\text{Hz}$ $1.2 \times 10^{-5} + 0.2 \text{ mHz}$ $1.2 \times 10^{-5} + 2.0 \text{ mHz}$ $1.2 \times 10^{-5} + 20 \text{ mHz}$ $1.2 \times 10^{-5} + 0.2 \text{ Hz}$ $1.2 \times 10^{-5} + 2.0 \text{ Hz}$	Fluke 8508A VP000040CMC

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> ( $\pm$ )	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Timer / Stopwatch	Deviation per day	0.035 s/24 h	Helmut Klein Timometer 4500 CP000020

<sup>1</sup>The uncertainty covered by the Calibration and Measurement Capability (CMC) is expressed as the expanded uncertainty having a coverage probability of approximately 95 %. It is the smallest measurement uncertainty that a laboratory can achieve within its scope of accreditation when performing calibrations of a best existing device. The measurement uncertainty reported on a calibration certificate may be greater than that provided in the CMC due to the behavior of the calibration item and other factors that may contribute to the uncertainty of a specific calibration.

<sup>2</sup>When uncertainty is stated in relative terms (such as percent, a multiplier expressed as a decimal fraction or in scientific notation), it is in relation to instrument reading or instrument output, as appropriate, unless otherwise indicated.

<sup>3</sup>Capability is suitable for the calibration of measuring devices in the stated ranges.

<sup>4</sup>Capability is suitable for the calibration of devices intended to generate the indicated quantity in the stated ranges.

<sup>5</sup>The actual frequency applied by the calibrator cannot be selected and may be dependent on the measurement device under calibration. Approximate frequency ranges for a given capacitance or capacitance range may be found in the Fluke 552xA's published specifications.