



INTERNATIONAL
ACCREDITATION
SERVICE®

CERTIFICATE OF ACCREDITATION

This is to attest that

GLOBAL INSIGNIA FOR TECHNICAL SERVICES COMPANY

BANDER ALBADEEN BUSINESS CENTER, BUILDING NO. 7023-SALMAN AL FARISI ST AL KHALIDIYYAH
DAMMAM 32232, KINGDOM OF SAUDI ARABIA

Calibration Laboratory CL-186

has met the requirements of AC204, *IAS Accreditation Criteria for Calibration Laboratories*, and has demonstrated compliance with ISO/IEC Standard 17025:2017, *General requirements for the competence of testing and calibration laboratories*. This organization is accredited to provide the services specified in the scope of accreditation.

Effective Date January 21, 2024

Expiration Date May 1, 2026



A handwritten signature in black ink, reading "Raj Nathan".

President

Visit www.iasonline.org for current accreditation information.

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

GLOBAL INSIGNIA FOR TECHNICAL SERVICES COMPANY

www.globalinsignia.com

Contact Name Renju J. Cherian

Contact Phone +966 558443064

Accredited to ISO/IEC 17025:2017

Effective Date January 21, 2024

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)*

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
<i>Dimensional</i>			
Gauge Blocks (Grade 1 & Grade 2 Only)	1 mm to 25 mm 25 mm to 50 mm 50 mm to 75 mm 75 mm to 100 mm	0.68 µm 0.71 µm 0.80 µm 0.90 µm	LABMASTER 1000M Grade '0' Gauge Block set ASME B89.1.2M-1991 GIFTS Procedure: CALPRGB-01
Caliper (Digital, Dial & Vernier types)	0 mm to 300 mm	7.7 µm	Grade '0', Gauge Block set ISO 13385-1: 2019 GIFTS Procedure: CALPRDVC-01
Dial Indicator	1 mm Up to 12 mm	1 µm 2 µm	LABMASTER 1000M Grade '0' Gauge Block set ASME B89.1.10M-2001 GIFTS Procedure: CALPRDDIG-01
3-Wire & 2-Wire Sets for Thread Measurements	Ø 0.007 mm to Ø 6 mm Ø 6 mm to Ø 10 mm	1.2 µm 2.5 µm	LABMASTER 1000M Grade '0' Gauge Block set ASME B89.1.17-2001 GIFTS Procedure: CALPRDTMW-01
Setting Rods	25 mm to 1000 mm	10 µm	LABMASTER 1000M Grade '0' Gauge Block set IS 7014 -1973(RA 1995) GIFTS Procedure: CALPRDTSR-01
Micrometer	Up to 200 mm	9.4 µm	LABMASTER 1000M Grade '0' Gauge Block set ASME B89.1.13-2013 GIFTS Procedure: CALPRDOM-01

* If information in this CMC is presented in non-SI units, the conversion factors stated in NIST Special Publication 811 "Guide for the Use of the International System of Units (SI)" apply.

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Height Gauge	0 mm to 1000 mm	12 µm	Standard Setting Rod BS: 1643:2008 GIFTS Procedure: CALPRDHG-01
Cylindrical Ring Gages (Inside Diameter) (Type: Plain Cylindrical Ring Gages, Setting Ring Gages, Master Setting ring Gages and GO/NO-GO Ring Gages)	Up to Ø 200 mm	3.2 µm	LABMASTER 1000M Grade '0' Gauge Block set IS 3485-1983 (RA1998) GIFTS Procedure: CALPRDCRG-01
Thread Plug Gauge (Major and Pitch diameter only)	Up to M 60 (also, up to Ø2.5"X 4UNC) Major diameter Pitch diameter	 1.6 µm 1.9 µm	LABMASTER 1000M Grade '0' Gauge Block set ANSI/ASME B1.2-R2017 GIFTS Procedure: CALPRDTT-01
Mechanical			
Pressure Gauges	0 bar to 40 bar 0 bar to 300 bar 0 bar to 1000 bar 0 bar to 4000 bar	0.04 bar 0.14 bar 0.58 bar 5.3 bar	BS EN 837-1, WIKA Digital Pressure Gauge. GIFTS Procedure: CALPRMPG-01
Pressure Transmitter	0 bar to 1000 bar	0.1 bar	WIKA Pressure Indicator Euramet cg-17 GIFTS Procedure: CALPRMTX-01
Pressure Transducer	0 bar to 3000 bar	4.5 bar	WIKA Pressure Indicator Euramet cg-17 GIFTS Procedure: CALPRMT-01
Safety Relief Valve	0 bar to 300 bar	1.7 bar	API 527, API 526 and API 576 Ventil Testing Unit with AE Sensor GIFTS Procedure: TESPRPRV-01
Vacuum Gauge	-0.9 bar to 5 bar	0.02 bar	WIKA Digital Vacuum Gauge ISO 3567: 2011 GIFTS Procedure: CALPRMVG-01
Weight Set (Mass)	2 g to 2 kg	17 mg	Mettler Toledo Analytical Balance. NPL MGP Guide No.:71 GIFTS Procedure: CALPRPWS -01

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Torque Wrench	2.5 N·m to 25 N·m 10 N·m to 100 N·m 50 N·m to 500 N·m 100 N·m to 800 N·m 300 N·m to 3000 N·m	0.11 N·m 0.25 N·m 1.9 N·m 8.8 N·m 12 N·m	ISO 6789, NORBAR Static Torque Transducer GIFTS Procedure: CALPRMTW-01
Hydraulic Torque Wrench	5000 N·m to 20000 N·m	100 N·m	Hydraulic Torque Wrench Tester GIFTS Procedure: CALPRHTW-01
Volumetric Glassware (Glass Syringe, Pipette)	Up to 100 mL	0.42 mL	METTLER-TOLEDO Analytical Balance Euramet cg-19 GIFTS Procedure: CALPRAVG-01
Volumetric Glassware (Measuring Cylinder, Receiving Cylinder, Volumetric Flask, Burette, and Measuring Jar.)	Up to 1000 L	1.5 mL	METTLER-TOLEDO Analytical Balance. Euramet cg-19 GIFTS Procedure: CALPRAVG-01
Thermal			
RTD	-45 °C to 200 °C	0.72 °C	ISOTECH Precision Thermometer ASTM E1137/E1137M-08 GIFTS Procedure: CALPRRTD-01
Thermocouple	100 °C to 1100 °C	1.8 °C	WIKA Digital Thermometer ASTM E0220-19 GIFTS Procedure: CALPRTTTC-01
Liquid in Glass Thermometer	-45 °C to 200 °C	0.56 °C	ISOTECH Precision Thermometer NBS Monograph-90 GIFTS Procedure: CALPRTLGL-01
Temperature Transmitter	-45 °C to 1100 °C	2.8 °C	ISOTECH Precision Thermometer Euramet cg-11 GIFTS Procedure: CALPRTTX-01
Oven & Furnace	up to 300 °C	2.7 °C	WIKA Digital Thermometer ASTM E145-19 GIFTS Procedure: CALPRTOF-01

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Electrical – DC/LF			
DC Current Generate ³	0 µA to 200 µA 200 µA to 2 mA 2 mA to 20 mA 20 mA to 200 mA 0 A to 100 A 100 A to 500 A 500 A to 1000 A	1.2 µA 0.0029 mA 0.0071 mA 0.019 mA 0.6 A 3 A 6 A	EURAMET cg-15, Multifunction Calibrator, Time Electronics LTD GIFTS Procedure: CALPREMM-01
DC Voltage Generate ³	0 mV to 20 mV 0 mV to 200 mV 0 V to 2 V 0 V to 20 V 0 V to 200 V 0 V to 500 V 0 V to 1000 V	0.0013 mV 0.0035 mV 0.055 mV 5.1 mV 9.7 mV 60 mV 83 mV	EURAMET cg-15, Multifunction Calibrator, Time Electronics LTD GIFTS Procedure: CALPREMM-01
AC Current Generate ³ 20 Hz to 1 kHz	0 µA to 200 µA 0 mA to 2 mA 0 mA to 20 mA 0 mA to 200 mA	1.3 µA 0.0012 mA 0.012 mA 0.12 mA	EURAMET cg-15, Multifunction Calibrator, Time Electronics LTD
AC Current Generate ³ 20 Hz to 500 Hz	200 mA to 20A	0.1 A	GIFTS Procedure: CALPREMM-01
AC Current Generate ³ @ 60 Hz	20 A to 500 A 500 A to 1000 A	0.11 A 0.32 A	
AC Voltage Generate ³ 20 Hz to 1 kHz	20 mV to 200 mV 200 mV to 2 V 2 V to 20 V	1.2 mV 0.29 mV 22 mV	EURAMET cg-15, Multifunction Calibrator, Time Electronics LTD
AC Voltage Generate ³ 40 Hz to 1 kHz	20 V to 200 V 200 V to 500 V 500 V to 1000 V	0.26 V 0.53 V 0.99 V	GIFTS Procedure: CALPREMM-01
DC Resistance Generate ³	10 Ω to 19 Ω 20 Ω to 100 Ω 101 Ω to 1k Ω 1 kΩ to 10 kΩ 10 kΩ to 100 kΩ 100 kΩ to 1 MΩ 1 MΩ to 4 MΩ	17 mΩ 26 mΩ 130 mΩ 2.3 Ω 0.023 kΩ 0.28 kΩ 0.0072 MΩ	EURAMET cg-15, Multifunction Calibrator, Time Electronics LTD GIFTS Procedure: CALPREMM-01
Electrical Simulation of Thermocouples (Generate) Type K Type J	-200 °C to 1200 °C -180 °C to 1200 °C	0.15 °C 0.15 °C	EURAMET cg-15, Multifunction Calibrator, Time Electronics LTD

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Electrical Simulation of Thermocouples (Generate) (cont'd.)			GIFTS Procedure: CALPRECT-01
Type T	-200 °C to 395 °C	0.12 °C	
Type R	0 °C to 1750 °C	0.34 °C	
Type S	0 °C to 1750 °C	0.34 °C	
Type E	-50 °C to 995 °C	0.09 °C	
Electrical Simulation of RTD (Generate) Pt-100	-180 °C to 800 °C	0.45 °C	EURAMET cg-15, Multifunction Calibrator, Time Electronics LTD GIFTS Procedure: CALPRECR-01
Capacitance Generate ³ @1 kHz	1 nF to 1000 nF	0.67 nF	EURAMET cg-15, Multifunction Calibrator, Time Electronics LTD GIFTS Procedure: CALPREC-01
Time and Frequency			
Non-Contact Tachometer	6 rpm to 60000 rpm	3 rpm	EURAMET cg-15, Multifunction Calibrator, Time Electronics LTD GIFTS Procedure: CALPRENT-01

¹The uncertainty covered by the Calibration and Measurement Capability (CMC) is expressed as the expanded uncertainty having a coverage probability of approximately 95 %. It is the smallest measurement uncertainty that a laboratory can achieve within its scope of accreditation when performing calibrations of a best existing device. The measurement uncertainty reported on a calibration certificate may be greater than that provided in the CMC due to the behavior of the calibration item and other factors that may contribute to the uncertainty of a specific calibration.

²When uncertainty is stated in relative terms (such as percent, a multiplier expressed as a decimal fraction or in scientific notation), it is in relation to instrument reading or instrument output, as appropriate, unless otherwise indicated.

³Capability is suitable for the calibration of measuring devices in the stated ranges.