



**International Accreditation Service, Inc.**  
3060 Saturn Street, Suite 100  
Brea, CA 92821 USA  
t: 562.364.8201  
t: 866.427.4422  
f: 562.699.8031  
www.iasonline.org

## **ACCREDITATION CRITERIA FOR CALIBRATION LABORATORIES**

### **AC204**

**April 2017  
(Effective June 1, 2017)**

### **PREFACE**

The attached accreditation criteria have been issued to provide all interested parties with guidelines on implementing performance features of the applicable standards referenced herein. The criteria were developed and adopted following public hearings conducted by the International Accreditation Service, Inc. (IAS), Accreditation Committee and are effective on the date shown above. All accreditations issued or reissued on or after the effective date must comply with these criteria. If the criteria are an updated version from a previous edition, solid vertical lines (|) in the outer margin within the criteria indicate a technical change or addition from the previous edition. Deletion indicators (→) are provided in the outer margins where a paragraph or item has been deleted if the deletion resulted from a technical change. These criteria may be further revised as the need dictates.

IAS may consider alternate criteria provided the proponent submits substantiating data demonstrating that the alternate criteria are at least equivalent to the attached criteria and otherwise meet applicable accreditation requirements.

Copyright © 2017

# ACCREDITATION CRITERIA FOR CALIBRATION LABORATORIES

## 1. INTRODUCTION

1.1. **Scope:** These criteria set forth the requirements for obtaining and maintaining International Accreditation Service, Inc. (IAS), Calibration Laboratory accreditation. These criteria supplement the IAS Rules of Procedure for Laboratory Accreditation.

1.2. **Normative and Reference Documents:** Publications listed below refer to current editions (unless otherwise stated).

1.2.1. ISO/IEC Standard 17025, General requirements for the competence of testing and calibration laboratories.

1.2.2. ILAC-P9, ILAC Policy for Participation in Proficiency Testing Activities.

1.2.3. ISO/IEC Standard 17000, Conformity assessment — Vocabulary and general principles.

## 2. DEFINITIONS

Applicable definition of ISO/IEC Standard 17000 series apply.

## 3. ELIGIBILITY

Accreditation services are available to calibration laboratories that:

3.1. Perform calibrations of its own company (internal).

3.2. Perform calibrations of products of external organizations (external).

## 4. REQUIRED BASIC INFORMATION

Calibration laboratories must demonstrate compliance with the following requirements:

4.1. ISO/IEC Standard 17025, General requirements for the competence of testing and calibration laboratories.

4.2. IAS Rules of Procedure for Laboratory Accreditation.

## 5. ADDITIONAL INFORMATION (AS APPLICABLE)

5.1. **Measurement Uncertainty:** Accredited calibration laboratories must report measurement uncertainty in all calibration reports.

5.2. **Proficiency Testing Activity:** Proficiency testing activity shall be completed in accordance with Clause 4.2 of ILAC-P9.

## 6. LINKS TO ADDITIONAL REFERENCES:

- 6.1 Asia Pacific Laboratory Accreditation Cooperation – [www.aplac.org](http://www.aplac.org)
- 6.2 International Laboratory Accreditation Cooperation – [www.ilac.org](http://www.ilac.org)
- 6.3 International Organization for Standardization – [www.iso.org](http://www.iso.org)
- 6.4 International Electrotechnical Commission – [www.iec.ch](http://www.iec.ch)
- 6.5 International Accreditation Service – [www.iasonline.org](http://www.iasonline.org)

*These criteria were previously issued September 2002, May 2004, August 2006, April 2008, May 2010, June 2013 and February 2016*