



CERTIFICATE OF ACCREDITATION

This is to attest that

TRANSCAL

NO. 100, 10TH CROSS, BETWEEN SAMPIGE ROAD AND MARGOSA ROAD, MALLESWARAM
BANGALORE, KARNATAKA 560003, REPUBLIC OF INDIA

Calibration Laboratory CL-233

has met the requirements of AC204, *IAS Accreditation Criteria for Calibration Laboratories*, and has demonstrated compliance with the ISO/IEC Standard 17025:2017, *General requirements for the competence of testing and calibration laboratories*. This organization is accredited to provide the services specified in the scope of accreditation maintained on the following page(s).

This certificate is valid up to September 1, 2021.

(See laboratory's scope of accreditation for fields of calibration and accredited calibration.)



This accreditation certificate supersedes any IAS accreditation bearing an earlier effective date. The certificate becomes invalid upon suspension, cancellation or revocation of accreditation. See www.iasonline.org for current accreditation information, or contact IAS at 562-364-8201.



A handwritten signature in black ink, reading 'Raj Nathan', positioned above a horizontal line.

Raj Nathan
President



SCOPE OF ACCREDITATION

IAS Accreditation Number	CL-233
Accredited Entity	TRANSCAL
Address	No. 100, 10 th cross, Between Sampige Road and Margosa Road, Malleswaram Bangalore, Karnataka 560003, Republic of India
Contact Name	Saraswathi k
Telephone	+91-8095111880
Effective Date of Scope	August 4, 2020
Accreditation Standard	ISO/IEC 17025:2017

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)^{1,2}

CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ³ (±)	TECHNIQUE, REFERENCE STANDARD, EQUIPMENT
<i>Dimensional</i>			
Height Gauge (Vernier/Dial/Digital)	0 mm to 600 mm 0 mm to 1000 mm	2.2 µm 2.4 µm	Using Gauge Blocks/Caliper Checker
V- Block (Parallelism, Symmetry)	300x125x200 mm	4.4 µm	Using Lever Dial Gauge & Mandrel
Caliper (Vernier/Dial/Digital)	0 mm to 600 mm 0 mm to 1000 mm 0 mm to 2000 mm	7.8 µm 10 µm 12 µm	Using Gauge Block Set, Caliper Checker
External Micrometer (Mechanical / Electronic / Digital)	0 mm to 25 mm	0.4 µm	Using Gauge Blocks
	0 mm to 150 mm	0.6 µm	Using Gauge Blocks
	0 mm to 1000 mm 0 mm to 2000 mm	1.7 µm 11 µm	
Electronic Probe	0 to 25mm	0.5 µm	Using Gauge Blocks
	0 to 50 mm	0.9 µm	
Wire Gauge	0.19 mm to 7.62 mm	8 µm	Using Vision System
Hegman Gauge	up to 1 mm	2.3 µm	Using Plunger Dial Gauge
Wet Film Thickness Gauge	0.025 mm to 5 mm	5.9 µm	Vision System
Depth Micrometer	0 mm to 300 mm	6 µm	Using Gauge Block Set
	0 mm to 600 mm	11 µm	
Depth Caliper	0 mm to 300 mm	6.4 µm	Using Gauge Block Set
	0 mm to 600 mm	13 µm	
Micrometer Setting Rod	25 mm to 1000 mm	4.7 µm	Using Gauge Block Set
	1000 mm to 1950 mm	7.9 µm	
Angle Graticule	0 deg to 360°°	1.8'	Using Vision System, Comparison method
Laser Distance Meter	0 mm to 2000 mm	350 µm	Using Slip Gauge, Comparison Method
Inclinometer	0° to 90°	1.7'	Using Angle Gauge Blocks, Comparison Method



SCOPE OF ACCREDITATION

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)^{1,2}

CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ³ (±)	TECHNIQUE, REFERENCE STANDARD, EQUIPMENT
Bench Center (Co-axiality)	0 mm to 500 mm	8.9 µm	Using Straight & Taper Mandrels, Dial Gauge. IS 5980
Angle Plate	450 mm x 300 mm x 200 mm	11 µm	Using Liver Dial Gauge, IS 2554, IS 6973
Flakiness Gauge	0 mm to 100 mm	5.9 µm	Using Vision System & 2D Height Gauge, Comparison Method
Elongation Gauge	0 mm to 100 mm	5.9 µm	Using 2D Height Gauge, Comparison Method
Dial Gauge (Plunger /Digital/ Dial Thickness Gauge)	0 mm to 100 mm	0.7 µm	Using Length Measuring Machine
Dial Gauge (Lever Type)	0 mm to 2 mm	0.5 µm	Using Length Measuring Machine
Bore Dial Gauge for Transmission Accuracy check	0 mm to 2 mm	1.8 µm	Using Length Measuring Machine
Surface Plate	5 m X 3 m	$1.3 \sqrt{((L+W)/100)}$ L = Length in mm W = Width in mm	Using Electronic Level
Plain Plug Gauge	0 mm to 100 mm 100 mm to 400 mm	1 µm 1.6 µm	Using Length Measuring Machine, Master Disc, FCDM
Feeler Gauge	Up to 1 mm	1.4 µm	Using Digital Micrometer
Cylindrical Measuring Pin	0.1 mm to 26 mm	0.9 µm	Using Length Measuring Machine, Master Disc, FCDM
Thread Plug Gauge (Major Diameter & Effective Diameter)	3 mm to 100 mm 100 mm to 400 mm	1.2 µm 1.7 µm	Using Length Measuring Machine, Master Disc, FCDM
Snap Gauge	3 mm to 500 mm	2.7 µm	Using Gauge Block Set
Bevel Protractor / Combination Set	0° to 360°	2.9'	Using Angle Block Set
Thread Ring Gauge (For Effective Diameter & Minor Diameter only)	3 mm to 325 mm	1.9 µm	Using Length Measuring Machine, Master Ring
Spirit Level (Type 1, 2 & 3)	Up to 4 mm/m	7.1 µm	Using Electronic Level
Plain Ring Gauge	3 mm to 325 mm	1.9 µm	Using Length Measuring Machine, Master Ring
Measuring Scales	Up to 3000 mm	$114 \sqrt{(L)} \mu\text{m}$ where L is length in m	Using Length Measuring Machine
Measuring Tape/ Pi Tape	Up to 50 m	$114 \sqrt{(L)} \mu\text{m}$ where L is length in m	Using Length Measuring Machine



SCOPE OF ACCREDITATION

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)^{1,2}

CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ³ (±)	TECHNIQUE, REFERENCE STANDARD, EQUIPMENT
Thread Pitch Gauge Pitch	0.25 mm to 6.35 mm	4.8 µm	Using Vision System
Thread Pitch Gauge Angle	55° 60°	1.2' 1.2'	Using Vision System
Radius Gauge	0.4 mm to 50 mm	5.6 µm	Using Vision System
Internal / Stick Micrometer (2 Point)	50 mm to 5000 mm	(1.25 + 3.27L) µm where L is length in m	Using Gauge Block Set
Engineering Square (Squareness)	Up to 400 mm	6.2 µm	Using Granite Square & Slip Gauge
Test Sieves	0.03 mm to 125 mm	4.8 µm	Using Vision System
Comparator Stand (Flatness of worktable)	200 mm x 200 mm	2.4 µm	Using Lever Dial Gauge
Straight Edge/Parallels	Up to 2000 mm	2.8 µm	Using Electronic Level
Inside/ Outside Dial Caliper	0 mm to 150 mm	0.7 µm	Using Gauge Block Set
Pistol Caliper	0 mm to 100 mm	60 µm	Using Gauge Block Set
Thickness Plate/Foils	Up to 2.5 mm	1.6 µm	Using Digital Micrometer
Limit Gauge /Test probes (Length, Radius / Diameter, Angle)	0 mm to 400 mm 360°	4.6 µm 2.4'	Using Vision System
Coating Thickness Gauge with Foils	0 mm to 2000 µm	1.8 µm	Using Standard Thickness Foils
Taper Thread Plug Gauge (Up to 150 mm diameter)	Taper angle and Flank angle Diameter at ends	3.2" 1.4 µm	Using Length Measuring Machine, Master Disc, FCDM; Vision system
3-Point Micrometer	2.5 mm to 100 mm	2.0 µm	Using set of Setting Ring Gauges
Taper Scale	1 mm to 15 mm	4.8 µm	Using Vision System
Ultrasonic Thickness Gauge	0 mm to 300 mm	38 µm	Using Gauge Block Set
Length Bars	50 mm to 500 mm	2.9 µm	Using Length Measuring Machine
Micrometer Head	0 mm to 50 mm	0.8 µm	Using Length Measuring Machine
Fillet Gauge/ Form Gauge	0 mm to 150 mm 0° to 90°	5.8 µm 5.4'	Using Vision System
Taper Plain Plug Gauge	Taper Half Angle Up to 100 mm	0.6" 0.9 µm	Using Length Measuring Machine
Taper Thread Ring Gauge	Up to 100 mm	0.9 µm	Using Length Measuring Machine



SCOPE OF ACCREDITATION

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)^{1,2}

CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ³ (±)	TECHNIQUE, REFERENCE STANDARD, EQUIPMENT
Taper Plain Ring Gauge	Taper Half Angle Up to 100 mm	5.5" 0.9 µm	Using Length Measuring machine
Sine Bar/ Sine Centre / Sine Table	0° to 45°	2.8"	Using Gauge Blocks, Angle Blocks, Lever Dial Gauge
Profile Projector / Video Measuring machine /Microscope	10 X to 50 X	0.06 %	Using Gauge Blocks/Glass Scale
	0° to 360°	2.8"	Using Angle Gauge Blocks/Glass Scale
	400 mm	2.9 µm	Using Gauge Blocks/Glass Scale
Floating Carriage Micrometer	0 mm to 100 mm	1.5 µm	Using Mandrels & Master Cylinders, MOY/SCMI/9
Thread Measuring Wire	0.170 mm to 6.350 mm	0.5 µm	Using Electronic Probe with DRO, Comparison Method
Cylindrical Master (Diameter & Concentricity)	3 mm to 100 mm	1.1 µm (diameter) 1.3 µm (concentricity)	Electronic Probe with DRO, IS 6311
Slip Gauges (Gauge Blocks)	0.5 mm to 25 mm	0.07 µm	Slip Gauge Calibrator & k Grade Slip Gauges, IS 2984, ISO 3650
	25 mm to 50 mm	0.10 µm	
	50 mm to 75 mm	0.12 µm	
	75 mm to 100 mm	0.15 µm	
Dial Calibration Tester	0 mm to 25 mm	0.7 µm	Using Electronic Probe with DRO, by Comparison Method
Gauge Block Comparators	0 mm to 100 mm	0.03 µm	Using K grade Gauge Blocks by Direct Method
Surface Roughness Master specimen	Ra up to 10 µm Rz up to 25 µm	8 % 8 %	Using Surface Roughness Tester
Surface Roughness Tester	1 µm to 800 µm	8.1 %	Using Depth Master and roughness master
Caliper Checker	20 mm to 600 mm	2.1 µm	Electronic Comparator and gauge block (0 grade)
Length Measuring Machine	0 mm to 100 mm	0.3 µm	Using Gauge Block Set (0 Grade)
	0 mm to 1000 mm (Tape & Scale Calibration)	3.7 µm	Using Gauge Block Set and Long Gauge Block (0 Grade)
Long Gauge Blocks	100 mm to 200 mm	0.7 µm	Electronic Comparator and gauge blocks (0 Grade)
	200 mm to 300 mm	0.9 µm	
	300 mm to 500 mm	1.5 µm	
Glass Scale	0 mm to 300 mm	6.6 µm	Using Vision Measuring System
Mechanical			
Sound Level Meter (at 1 kHz)	94 dB and 114 dB	0.2 dB	Using Sound Level Calibrator



SCOPE OF ACCREDITATION

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)^{1,2}

CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ³ (±)	TECHNIQUE, REFERENCE STANDARD, EQUIPMENT
Vibration Machine/Vibration Measurement (10 Hz to 6.8 kHz)	1 g to 40 g (where g is acceleration due to gravity)	3.2 %	Using Accelerometer
One Mark Pipette, Graduated Pipette, Graduated Burette, Volumetric Flask, Measuring Jar (Volumetric Glassware)	1 mL to 10 mL 10 mL to 100 mL 100 mL to 1000 mL 1000 mL to 5000 mL	0.00018 mL 0.0004 mL 0.005 mL 0.017 mL	Calibration based on Gravimetric method as per ISO 4787 using weighing balance of d = 0.1 mg, 1 mg, 10 mg and Distilled Water
Micro Pipette	1 µL to 10 µL 10 µL to 100 µL 100 µL to 1000 µL 1000 µL to 10000 µL	0.02 µL 0.04 µL 0.11 µL 1.9 µL	Calibration based on Gravimetric method as per ISO 8655 using weighing balance of d = 0.001 mg, 0.01 mg and Distilled Water
Digital / Analogue Pressure Gauges, Differential Pressure Gauge, Transducers, Transmitters, Switches	0.1 mbar to 10 mbar 10 mbar to 100 mbar 100 mbar to 2 bar 2 bar to 20 bar 20 bar to 40 bar 40 bar to 700 bar 700 bar to 1000 bar	0.93 % 0.06 % 0.045 % 0.043 % 0.016 % 0.023 % 0.02 %	Digital Pressure Gauge using Hydraulic Comparator Pump based on DKD – R 6 - 1
	1 bar to 35 bar 35 bar to 1200 bar	0.026 % 0.02 %	Using Hydraulic (Oil operated) Dead Weight Tester based on DKD – R6 - 1
Pirani gauges	0.006 mbar to 1013 mbar	1.7 %	Using Standard Pirani Gauge
Digital/Analogue Vacuum Gauges, Transducers/ Transmitters	-0.01 bar to -0.9 bar	0.22 %	Digital Pressure Gauge using Pneumatic Vacuum Comparator Pump based on DKD – R6 - 1
Absolute Pressure Gauges	200 mbar to 915 mbar (abs)	0.26 %	Standard absolute Gauge using desiccator and vacuum pump
Altimeter Chamber	30 mbar to 915 mbar	3 mbar	Using Barometer



SCOPE OF ACCREDITATION

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)^{1,2}

CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ³ (±)	TECHNIQUE, REFERENCE STANDARD, EQUIPMENT	
Calibration of E1 Class Weights and coarser	1 mg	0.0010 mg	Using E1 class Standard weights and Mass comparator of $d = 0.1 \mu\text{g}$, per OIML R – 111 by subdivision through ABBA cycles	
	2 mg	0.0010 mg		
	5 mg	0.0010 mg		
	10 mg	0.0010 mg		
	20 mg	0.0010 mg		
	50 mg	0.0010 mg		
	100 mg	0.0013 mg		
	200 mg	0.0014 mg		
	500 mg	0.0016 mg		
	1 g	0.003 mg		Using Mass Comparator of $d = 0.001 \text{ mg}$, per OIML R 111 by Substitution Method
	2 g	0.004 mg		
	5 g	0.005 mg		
10 g	0.006 mg			
20 g	0.007 mg			
	50 g	0.01 mg	Using Mass Comparator of $d = 0.01 \text{ mg}$, per OIML R 111 by Substitution Method	
	100 g	0.02 mg		
	200 g	0.03 mg		
	500 g	0.10 mg		
	1 kg	0.20 mg		
Calibration of E2 Class Weights and coarser	2 kg	1 mg	Using Mass Comparator of $d = 1 \text{ mg}$, by Substitution Methods through ABBA Cycles as per OIML R - 111	
	5 kg	2 mg		
	10 kg	3 mg		
	20 kg	7 mg		
	50 kg	100 mg		Using Mass Comparator of $d = 100 \text{ mg}$, by Substitution Methods through ABBA Cycles as per OIML R - 111
Calibration of Class 1 Weighing Balances and coarser	1mg to 2 g	0.0035 mg	E1 Class Standard Weights 1 mg to 20 kg, as per OIML R - 76	
	1 mg to 5 g	0.005 mg		
	1 mg to 20 g	0.007 mg		
	1mg to 50 g	0.02 mg		
	1mg to 200 g	0.03 mg		
	1mg to 500 g	0.10 mg		
	1mg to 1 kg	0.20 mg		
	500 mg to 5 kg	1 mg		
	500 mg to 20 kg	7 mg		
	Calibration of Class 2 Weighing Balances and coarser	500 mg to 50 kg		100 mg
500 mg to 150 kg		1 g		
1 kg to 300 kg		20 g		
1 kg to 1000 kg		100 g		
2 kg to 3000 kg		500 g		
Spring Balance	10 g to 1500 g	0.28 %	Using F1 class weights	
	1500 g to 100 kg	0.1 %		



SCOPE OF ACCREDITATION

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)^{1,2}

CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ³ (±)	TECHNIQUE, REFERENCE STANDARD, EQUIPMENT
Torque Wrench, Torque Driver, Type I - Class B, C, D, E Type II - Class A, B, D, E	0.1 N·m to 10 N·m 10 N·m to 20 N·m 20 N·m to 200 N·m 200 N·m to 1000 N·m	0.89 % 0.38 % 0.43 % 0.62 %	Torque sensors of various capacities using Torque Calibration Rig Based on ISO 6789
Torque Measuring Devices	0.1 N·m to 1 N·m 1 N·m to 10 N·m 10 N·m to 500 N·m	0.15 % 0.3 % 0.3 %	Using Dead Weight Torque Calibration System as per BS 7882
Load Cells / Force Proving Instruments	0.1 N to 100 N 100 N to 2000 N	0.07 % 0.15 %	Using Newton weights as per ISO 373
Universal Testing Machine (UTM)	Compression: 20 N to 10 kN 20 kN to 1000 kN	0.52 % 0.52 %	Using Force Proving Instruments as per IS 1828
	Tension: 20 N to 10 kN 20 kN to 100 kN	0.29 % 0.29 %	
Displacement Measuring System and Devices used in Material Testing Machine	0 mm to 600 mm	0.27 mm	Using Gauge Blocks as per ASTM E2309
Extensometer used in Material Testing Machine - Gauge Length up to 200 mm	0 mm to 12.7 mm	0.16 mm	Using Digital Micrometer as per ISO 9513 and ASTM E83
Speed of Material Testing Machine	0 mm/min to 500 mm/min	0.6 mm/min	As per ASTM E2658
Push Pull Gauge	1 N to 2000 N	0.21 %	Newtonian Weights And Frame Fixture VDI/VDE 2624 – Part 2.1
Hydrometer	SG =0.6 to 1.0	0.00014	Calibration of Hydrometers by Cuckows method
	SG =1.0 to 2.0	0.00014	
Viscosity Cups, Zahn Cups	30 cSt to 240 cSt	0.33 %	Calibration as per IS 3944, ASTM D4212 using liquids of known kinematic viscosity
Liquid Flow Meter	1 m ³ /h to 350 m ³ /h	0.62 %	Using Clamp-On Ultra Sonic Flow Meter by Comparison Method
Flow Meters (Air)	0.1 L/min to 5 L/min	0.77 %	Using Mass Flow Meter by Comparison Method
	5 L/min to 500 L/min	1.0 %	
Velocity (Medium Air)	0.3 m/s to 30 m/s	1.3 %	Using Digital Anemometer
Liquid Flow Measuring Devices	0.001 t/h to 0.2 t/h	0.03 %	Using Weighing System by Gravimetric Method
	0.2 t/h to 4.5 t/h	0.16 %	



SCOPE OF ACCREDITATION

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)^{1,2}

CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ³ (±)	TECHNIQUE, REFERENCE STANDARD, EQUIPMENT
Rubber Hardness Tester	Shore A, B, E, O Shore C, D, DO	0.1 % 0.1 %	Using Load Cell with Indicator
Rockwell Hardness Testing Machines	HRA HRBW HRC	0.62 HRA 0.65 HRBW 0.63 HRC	Using Standard Hardness Blocks as per IS 1586 (Part II) / ISO 6508 - 2 by Indirect Method
Brinell Hardness Testing Machines	HBW 2.5 / 187.5 HBW 5 / 750 HBW 10 / 3000	1.5 % 1.5 % 1.5 %	Using Standard Hardness Blocks as per IS 1500 (Part II) / ISO 6506 - 2 by Indirect Method
Vickers Hardness Testing Machines	HV 10 HV 20 HV 30 HV 50	1.1 % 1.1 % 1.1 % 1.1 %	Using Standard Hardness Blocks as per IS 1501 (Part II) / ISO 6507 - 2 by Indirect Method
Thermal			
Humidity Meters (Dial / Digital)	0.5 %RH 5 %RH 95 %RH	0.33 %RH 0.33 %RH 0.85 %RH	Using Humidity standard solution by direct method
Dial /Digital Humidity Meters	10 %RH to 95 %RH (at 10 °C to 60 °C)	0.80 %RH	Using Temperature & Humidity Meter with Humidity Chamber by Comparison Method
Humidity Transmitters (for temperature scale)	0 °C to 60 °C	0.18 °C	Using Class 'A' RTD Sensor /PRT sensor with Digital Indicator by comparison
RTD, Thermocouples, Indicator with sensor	-196 °C	0.08 °C	Using LN ₂ and cryo bath by comparison method
	-100 °C to -45 °C -45 °C to 140 °C 140 °C to 650 °C	0.09 °C 0.02 °C 0.08 °C	Using Dry Temperature Bath, SPRT with Digital Indicator by comparison Method
	650 °C to 1000 °C 1000 °C to 1200 °C	0.43 °C 1.5 °C	Using Dry Temperature Bath, S- Type Thermo couple with Digital Indicator by Comparison Method
Glass Thermometer	-40 °C to 200 °C	0.18 °C	Using Liquid Bath, SPRT with Digital Indicator by Comparison Method



SCOPE OF ACCREDITATION

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)^{1,2}

CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ³ (±)	TECHNIQUE, REFERENCE STANDARD, EQUIPMENT
Oil Bath, Low & High Temperature Bath, Dry Bath, Incubators & Autoclave, Thermal Chambers / Ovens, Water Bath, Furnace	-100 °C to 140 °C	0.072 °C	Using SPRT with Digital Indicator by Direct Method
	140 °C to 650 °C	0.08 °C	
	650 °C to 1200 °C	1.5 °C	Using S- Type Thermo couple with Digital Indicator by Direct Method
Humidity Chambers, Climatic Chambers, Dry cabinet, De-humidifier	10 %RH to 95 %RH (at 10 °C to 60 °C)	0.98 %RH	Using Temperature & Humidity Meter by Direct Method
	5 %RH to 10 %RH @ ambient temperature	1.3 %RH	
IR Thermometer	0 °C	1.0 °C	By Comparison Method
	50 °C to 650 °C	1.0 °C	
	650 °C to 1200 °C	1.7 °C	
Electrical/DC/Low Frequency			
AC Voltage Generate ⁴	50 µV to 2 mV (50 Hz to 1 kHz)	1 %	Using Calibrator Fluke 5700A/5522A by Direct Method
	2 mV to 20 mV (10 Hz to 40 Hz)	700 µV/V + 5 µV	
	(40 Hz to 20 kHz)	190 µV/V + 5 µV	
	(20 kHz to 50 kHz)	460 µV/V + 5.6 µV	
	(50 kHz to 100 kHz)	850 µV/V + 7 µV	
	(100 kHz to 300 kHz)	0.12 % + 16 µV	
	(300 kHz to 500 kHz)	0.17 % + 25 µV	
	(500 kHz to 1 MHz)	0.45 % + 5 µV	
	20 mV to 200 mV (10 Hz to 40 Hz)	320 µV/V + 8 µV	
	(40 Hz to 20 kHz)	140 µV/V + 9 µV	
	(20 kHz to 50 kHz)	380 µV/V + 9 µV	
	(50 kHz to 300 kHz)	0.13 % + 30 µV	
	(300 kHz to 1 MHz)	0.41 % + 100 µV	
	200 mV to 2 V (10 Hz to 40 Hz)	175 µV/V + 35 µV	
	(40 Hz to 20 kHz)	95 µV/V + 10 µV	
	(20 kHz to 50 kHz)	145 µV/V + 20 µV	
(50 kHz to 300 kHz)	0.04 % + 70 µV		
(300 kHz to 1 MHz)	0.3 % + 1 mV		
2 V to 20 V (10 Hz to 40 Hz)	175 µV/V + 350 µV		
(40 Hz to 20 kHz)	95 µV/V + 100 µV		
(20 kHz to 50 kHz)	145 µV/V + 200 µV		
(50 kHz to 300 kHz)	0.04 % + 700 µV		



SCOPE OF ACCREDITATION

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)^{1,2}

CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ³ (±)	TECHNIQUE, REFERENCE STANDARD, EQUIPMENT
AC Voltage Generate ⁴ continued	2 V to 20 V (300 kHz to 1 MHz)	0.3 % + 12 mV	Using Calibrator Fluke 5700A/5522A by Direct Method
	20 V to 30 V (300 kHz to 500 kHz)	0.9 % + 15 mV	
	20 V to 220 V (10 Hz to 40 Hz) (40 Hz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 100 kHz)	175 µV/V + 3.5 mV 100 µV/V + 1.1 mV 240 µV/V + 3.5 mV 0.06 % + 10 mV	
	200 V to 1000 V (50 Hz to 1 kHz)	150 µV/V + 4 mV	
AC Current Generate ⁴	10 µA to 200 µA (40 Hz to 1 kHz)	190 µA/A + 16 nA	Using Calibrator Fluke 5700A/5522A by Direct Method
	200 µA to 2 mA (10 Hz to 40 Hz) (40 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz)	650 µA/A + 35 nA 190 µA/A + 35 nA 700 µA/A + 0.4 µA 0.16 % + 0.8 µA	
	2 mA to 20 mA (10 Hz to 40 Hz) (40 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz)	650 µA/A + 350 nA 190 µA/A + 350 nA 700 µA/A + 4 µA 0.16 % + 8 µA	
	20 mA to 200 mA (10 Hz to 40 Hz) (40 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz)	650 µA/A + 3.5 µA 190 µA/A + 3.5 µA 700 µA/A + 40 µA 0.16 % + 80 µA	
	200 mA to 2 A (40 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz)	650 µA/A + 35 µA 900 µA/A + 80 µA 0.85 % + 160 µA	
	2 A to 10 A (45 Hz to 100 Hz) (100 Hz to 1 kHz) (1 kHz to 5 kHz)	0.08 % + 2 mA 0.08 % + 2 mA 3.4 % + 2 mA	
	10 A to 20 A (45 Hz to 100 Hz) (100 Hz to 1 kHz)	0.15 % + 5 mA 0.18 % + 5 mA	



SCOPE OF ACCREDITATION

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)^{1,2}

CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ³ (±)	TECHNIQUE, REFERENCE STANDARD, EQUIPMENT	
AC Current Generate ⁴ continued	10 A to 20 A (1 kHz to 5 kHz)	3.4 % +5 mA	Using Calibrator Fluke 5700A/5522A by Direct Method	
	20 A to 120 A (50 Hz to 1 kHz)	0.1 %	Using current source Omicron, Direct Method	
	120 A to 1000 A (50 Hz to 400 Hz)	0.4 %	Using current source current coil	
	1000 A to 3000 A (50 Hz)	1.2%		
AC Power Generate ⁴ Single Phase, 50 Hz, UPF 120 V to 1000 V 0.01 A to 20 A	0.01 W to 4.8 kW 4.8 kW to 1 MW	0.12 % 0.8 %	Using Calibrator Fluke 5522A Direct Method	
	AC Power Generate ⁴ Single Phase, 50 Hz, 0.2 PF lead/lag, 120 V to 1000 V 0.1 A to 20 A	2.4 W to 200 kW		1 %
	AC Power Generate ⁴ Single Phase, 50Hz, 0.5 PF lead/lag, 120 V to 1000 V 0.1 A to 20 A	6 W to 500 kW		0.50 %
	AC Power Generate ⁴ Single Phase, 50Hz, 0.8 PF lead/lag, 120V to 1000 V, 0.1 A to 20 A	9.6 W to 800 kW		0.23 %
	AC Energy Generate ⁴ Active /Reactive Single & Three Phase, 40 V to 300 V, 0.05 A to 20 A, 40 Hz to 70 Hz	0.5 Wh to 6 kWh (0.25 pF to 1 pF)		0.25 %
Power Factor, Three phase	0.25 PF Lag to Unity 0.25 PF Lead to Unity	0.006 PF	Using Edutech Energy source by direct method	



SCOPE OF ACCREDITATION

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)^{1,2}

CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ³ (±)	TECHNIQUE, REFERENCE STANDARD, EQUIPMENT
Power Factor, Single phase	0.2 PF Lead to unity 0.2 PF Lag to Unity	0.002 PF	Using Calibrator Fluke 5522A by Direct Method
DC Voltage Generate ⁴	1.018 V, 10 V	3 µV/V	Using 732B Reference Standard by Direct Method
	50 µV to 0.5 mV 0.5 mV to 220 mV 220 mV to 2.2 V 2.2 V to 11 V 11 V to 22 V 22V to 220 V 220 V to 1000 V	0.2 % 7 µV/V + 1 µV 8 µV/V + 1 µV 8 µV/V + 3.5 µV 8 µV/V + 6.5 µV 9 µV/V + 80 µV 11 µV/V + 500 µV	Using Calibrator Fluke 5700A by Direct Method
DC Current Generate ⁴ / Measure ⁵	10 nA to 10 µA	0.15 %	Using Reference 732B, Decade Megohm box, DMM by VI method
DC Current Generate ⁴	10 µA to 200 µA 200 µA to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 10 A 10 A to 20 A	70 µA/A + 8 nA 62 µA/A + 8 nA 72 µA/A + 80 nA 82 µA/A + 0.8 µA 100 µA/A + 25 µA 0.06 % + 500 µA 0.12 % + 750 µA	Using Fluke 5700A/5522A
	20 A to 1000 A 1000 A to 3000 A	0.4 % 1.2 %	Using Current Source with current coil
DC Power Generate ⁴ (1V to 1000 V, 1mA to 1000 A)	1 mW to 1 kW 1 kW to 1 MW	0.08 % 0.65 %	Using Calibrator Fluke 5522A with 50 turns current coil, Direct Method
DC Resistance Generate ⁴	75 µΩ 0.001 Ω to 0.1 Ω 0.1 Ω to 1 Ω	1 % 0.6 % 0.1 %	Using Standard Resistors & Shunts by V-I Method
	1 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 kΩ 1 kΩ to 10 kΩ 10 kΩ to 100 kΩ 100 kΩ to 1 MΩ 1 MΩ to 10 MΩ 10 MΩ to 100 MΩ 100 MΩ to 300 MΩ 300 MΩ to 1 GΩ	35 uΩ/Ω + 1 mΩ 35 uΩ/Ω + 1.4 mΩ 35 uΩ/Ω + 2 mΩ 35 uΩ/Ω + 0.02 Ω 35 uΩ/Ω + 0.2 Ω 40 uΩ/Ω + 2 Ω 150 uΩ/Ω + 50 Ω 600 uΩ/Ω + 3 kΩ 0.36 % + 100 kΩ 1.7 % + 500 kΩ	Using Calibrator Fluke 5522 A by Direct Method
	1 GΩ to 1 TΩ	1 %	Using Decade MegOhm Box
Discrete DC Resistance Generate ⁴	1 Ohm 1.9 Ohm 10 Ohm 19 Ohm 100 Ohm 190 Ohm	110 µΩ/Ω 59 µΩ/Ω 28 µΩ/Ω 32 µΩ/Ω 20 µΩ/Ω 29 µΩ/Ω	Using Calibrator Fluke 5700 with DMM 3458A by Direct Method



SCOPE OF ACCREDITATION

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)^{1,2}

CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ³ (±)	TECHNIQUE, REFERENCE STANDARD, EQUIPMENT
Discrete DC Resistance Generate ⁴ continued	1 kOhm 1.9 kOhm 10 kOhm 19 kOhm 100 kOhm 190 kOhm 1 MOhm 1.9 MOhm 10 MOhm 19 Mohm 100 MOhm	15 μΩ/Ω 16 μΩ/Ω 14 μΩ/Ω 14 μΩ/Ω 16 μΩ/Ω 16 μΩ/Ω 23 μΩ/Ω 24 μΩ/Ω 47 μΩ/Ω 55 μΩ/Ω 140 μΩ/Ω	Using Calibrator Fluke 5700 with DMM 3458A by Direct Method
Temperature Simulation – Generate ⁴ (temperature indicator / controller / recorder)	B Type: 600 °C to 1800 °C K Type: -200 °C to 1372 °C E Type: -200 °C to 1000 °C U Type: -200 °C to 400 °C L Type: -200 °C to 900 °C J Type: -200 °C to 1200 °C T Type: -200 °C to 400 °C N Type: -200 °C to 1300 °C R Type: 0 °C to 1750 °C S Type: 0 °C to 1750 °C RTD: -200 °C to 800 °C	0.5 °C 0.06 °C 0.08 °C 0.07 °C 0.08 °C 0.06 °C 0.1 °C 0.07 °C 0.07 °C 0.07 °C 0.07 °C	Using 5700A Calibrator DC mV measurement method
Capacitance Generate ⁴	(1 kHz) 220 pF to 1 μF 1 μF to 10 μF (100 Hz) 10 μF to 100 μF 100 μF to 1 mF 1 mF to 10 mF 10 mF to 110 mF	0.05 % + 10 pF 0.25 % + 15 nF 0.25 % + 150 nF 0.45 % + 1 μF 0.45 % + 10 μF 1.3 %	Using Resistance Method Using Calibrator Fluke 5520A, Decade Capacitance Box by Direct Method
Harmonics Generate ⁴ (45 Hz to 5 kHz)	2nd order to 39th order (33 mV to 1000 V and 3.3 mA to 20 A)	0.5 %	FLUKE 5522A Source, Direct Method
AC Voltage Measure ⁵	1 mV to 200 mV (50 Hz to 2 kHz) 100 mV to 2 V (20 Hz to 50 Hz) (2 kHz to 100 kHz) 200 mV to 2 V (50 Hz to 2 kHz) 1 V to 20 V (100 kHz to 1 MHz) 2 V to 20 V (20 Hz to 50 Hz)	130 uV/V + 4 uV 130 uV/V + 20 uV 600 uV/V + 0.2 mV 95 uV/V + 20 uV 1 % + 0.2 V 130 uV/V + 0.2 mV	Using 8½ DMM 8508A/HP 3458 A, by Direct Method



SCOPE OF ACCREDITATION

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)^{1,2}

CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ³ (±)	TECHNIQUE, REFERENCE STANDARD, EQUIPMENT
AC Voltage Measure ⁵ continued	2 V to 20 V (50 Hz to 2 kHz) (2 kHz to 100 kHz)	95 μ V/V + 0.2 mV 600 μ V/V + 2 mV	Using 8½ DMM 8508A/HP 3458 A, by Direct Method
	20 V to 200 V (20 Hz to 50 Hz) (50 Hz to 2 kHz) (2 kHz to 100 kHz)	130 μ V/V + 2 mV 95 μ V/V + 2 mV 600 μ V/V + 20 mV	
	100 V to 1000 V (50 Hz to 10 kHz) (10 kHz to 20 kHz)	130 μ V/V + 20 mV 270 μ V/V + 40 mV	
	1 kV to 5 kV (50 Hz)	0.23 %	
	5 kV to 28 kV (50 Hz)	2 %	Using High Voltage Divider with DMM's, Sources and HV Probe with DMM by Direct Method/Comparison Method
	28 kV to 100 kV (50 Hz)	2.3 %	
AC Current Measure ⁵	20 μ A to 200 μ A (50 Hz to 1 kHz)	250 μ A/A + 0.02 μ A	
	200 μ A to 20 mA (50 Hz to 1 kHz)	280 μ A/A + 2 μ A	
	20 mA to 200 mA (50 Hz to 10 kHz)	250 μ A/A + 0.02 mA	
	200 mA to 2 A (50 Hz to 10 kHz)	600 μ A/A + 0.2 mA	
	2 A to 20 A (50 Hz to 2 kHz) (2 kHz to 10 kHz)	800 μ A/A + 2 mA 0.25 % + 2 mA	
	1 A to 30 A (50 Hz to 5 kHz)	0.5 %	
	30 A to 1000 A (50 Hz)	0.6 %	
1000 A to 3000 A (50 Hz)	2.5 %	Using Shunt with DMM by V-I Method	
			Using Shunt with DMM by direct method Current coil & Clamp Meter
AC Power Measure ⁵ 1 Phase, 50 Hz @ UPF 120V to 240 V, 0.01 A to 20 A	1.2 W to 4.8 kW	0.25 %	Using Digital Power Meter WT 210 by Direct Method



SCOPE OF ACCREDITATION

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)^{1,2}

CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ³ (±)	TECHNIQUE, REFERENCE STANDARD, EQUIPMENT
AC Power Measure ⁵ 1 Phase, 50 Hz, 0.8 PF Lead/Lag 120V to 240 V, 0.1 A to 20 A	9.6 W to 3.8 kW	0.23 %	Using Digital Power Meter WT 210 by Direct Method
AC Power Measure ⁵ 1 Phase, 50 Hz, 0.5 PF Lead/Lag, 120 to 240V, 0.1 A to 20 A	6 W to 2.4 kW	0.5 %	Using Digital Power Meter WT 210 by Direct Method
AC Power Measure ⁵ 1 Phase, 50 Hz, 0.2 PF Lead/Lag, 120V to 240 V, 0.1A to 20 A	2.4 W to 960 W	0.5 %	Using Digital Power Meter WT 210 by Direct Method
High Frequency Power Measure ⁵ (400 kHz)	100 mW to 400 W	4 %	Using Differential probe in combination with Oscilloscope by comparison method
DC Voltage Measure ⁵	50 µV to 0.5 mV	0.8 %	Direct / Comparison Method using 8½ DMM HP 3458A/Fluke 8508A
	0.5 mV to 200 mV	7.2 µV/V + 0.1 µV	
	200 mV to 2 V	4.6 µV/V + 0.4 µV	
	2 V to 20 V	4.6 µV/V + 4 µV	
	20 V to 200 V	6.7 µV/V + 0.04 mV	
	200 V to 1000 V	6.8 µV/V + 0.5 mV	
	1 kV to 5 kV	0.20 %	Using HV Divider with DMM by Direct Comparison Method
	5 kV to 40 kV	2 %	Using Source & HV Probe with DMM by Direct Comparison Method
	40 kV to 100 kV	1.7 %	
DC Current Measure ⁵	10 nA to 100 nA	0.05 %	Using Shunt with DMM V-I / Comparison Method
	100 nA to 10 µA	0.08 %	Using 8½ DMM HP 3458A FLUKE8508A, by Direct /Comparison Method
	10 µA to 200 µA	13 µA/A + 0.4 µA	
	200 µA to 20 mA	14 µA/A + 0.4 µA	
	20 mA to 200 mA	55 µA/A + 0.8 µA	
	200 mA to 2 A	0.02 % + 16 µA	
	2 A to 20 A	0.04 % + 0.4 mA	
	20 A to 75 A	0.08 %	
	75 A to 100 A	0.60 %	
	100 A to 1000 A	1.0 %	



SCOPE OF ACCREDITATION

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)^{1,2}

CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ³ (±)	TECHNIQUE, REFERENCE STANDARD, EQUIPMENT
DC Power Measure ⁵ (1V to 600V, 1 mA to 20 A)	1 mW to 10 W	0.09 %	Using Digital Power Meter WT 210 by Direct Method
	10 W to 1 kW	0.50 %	
	1 kW to 12 kW	0.65 %	
Pulse Energy Measure ⁵ (1 ms to 10 ms pulse length)	1 J to 360 J	2.8 %	Using oscilloscope DPO7254, Differential Probe, DP-100 and P 6015 for measurements. Defibrillator used as source
Capacitance Measure ⁵	(1 kHz)		Using LCR Meter by Direct / Comparison Method
	1 pF to 1 µF	0.06 %	
	1 µF to 100 µF	0.13 %	
Inductance Measure ⁵	(1 kHz)		Using LCR Meter by Direct / Comparison Method
	100 µH to 1 mH	0.09 %	
	1 mH to 10 H	0.07 %	
AC Resistance Measure ⁵	(1 kHz)		Using LCR Meter
	1 Ω to 10 kΩ	0.06 %	
	(1 kHz to 100 kHz)		Using LCR Meter & calibrator
	100 Ω to 1 kΩ	0.16 %	
DC Resistance Measure ⁵	75 µΩ to 1 mΩ	0.60 %	Using micro-ohmmeter and Fluke 8508A DMM, standard resistors and shunts by VI method
	1 mΩ to 10 mΩ	0.60 %	
	10 mΩ to 1 Ω	0.035 % + 4 µΩ	Using DMM 81/2 8508A by Direct Comparison method
	0.1 Ω to 1 Ω	0.035 % + 40 µΩ	
	1 Ω to 2 Ω	12 µΩ/Ω + 4 µΩ	
	2 Ω to 20 Ω	7 µΩ/Ω + 30 µΩ	
	20 Ω to 200 Ω	9 µΩ/Ω + 50 µΩ	
	200 Ω to 2 kΩ	9 µΩ/Ω + 0.5 mΩ	
	2 kΩ to 20 kΩ	7.5 µΩ/Ω + 22 mΩ	
	20 kΩ to 200 kΩ	9 µΩ/Ω + 100 mΩ	
200 kΩ to 2 MΩ	6.5 µΩ/Ω + 7 Ω		
2 MΩ to 20 MΩ	7 µΩ/Ω + 160 Ω		
20 MΩ to 200 MΩ	4 µΩ/Ω + 7 kΩ		
200 MΩ to 2 GΩ	30 µΩ/Ω + 0.28 MΩ		
2 GΩ to 20 GΩ	0.05 % + 20 MΩ		
	20 GΩ to 1 TΩ	1 %	Using Fluke Calibrator & 8½ DMM by VI Method/Comparison
Temperature Simulation - Measure ⁵	B Type: 600 °C to 1800 °C	0.5 °C	Using 8½-digit DMM 3458A/8508A DC mV measurement method
	K Type: -200 °C to 1372 °C	0.05 °C	
	E Type: -200 °C to 1000 °C	0.08 °C	
	U Type: -200 °C to 400 °C	0.07 °C	
	L Type: -200 °C to 900 °C	0.08 °C	
	J Type: -200 °C to 1200 °C	0.06 °C	
	T Type: -200 °C to 400 °C	0.1 °C	
	N Type: -200 °C to 1300 °C	0.07 °C	
	R Type: 0 °C to 1750 °C	0.07 °C	
	S Type: 0 °C to 1750 °C	0.07 °C	



SCOPE OF ACCREDITATION

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)^{1,2}

CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ³ (±)	TECHNIQUE, REFERENCE STANDARD, EQUIPMENT
Temperature Simulation - Measure ⁵ continued	RTD: -200 °C to 800 °C	0.02 °C	Using 8½-digit DMM 3458A/8508A DC mV measurement method
Oscilloscope – DC Signal	1 mV to 130 V	0.08 %	Using Fluke Calibrator 5520A/5522A with 1.1 GHz option by Direct Method
Oscilloscope - Band Width @ 50 kHz Reference	50 kHz to 1 GHz	0.07 %	
Oscilloscope - Scope Amplitude, Square Wave Signal (10 Hz to 10 kHz)	1 mV to 55 Vp-p	0.12 %	
Oscilloscope, Time Marker	1 ns to 1000 s	0.00005 %	Using Rubidium standard, Fluke Calibrator 5520A/5522A with 1.1 GHz option by Direct Method
RF/Microwave			
3 dB Bandwidth Measure ⁵ (Filter, Power Meter, Power Sensor)	Up to 40 GHz	1 %	Using RF Reference Source - 9640ALPNX, Signal Generator and Power Meter by Direct Method
RF Power Measure ⁵ / Generate ⁴	(1 kHz to 18 GHz) -60 dBm to -100 dBm -60 dBm to 15 dBm	0.65 dB 0.24 dB	Using RF Reference source 9640A LPNX, Signal Generator, Attenuator, Multimeter, Power Meter, Spectrum Analyzer Method
	(18 GHz to 40 GHz) 15 dBm to -60 dBm	11 %	Using Signal Generator SMB 100A & USB Power Sensor U2054XA
	(18 GHz to 29.99 GHz) 15 dBm to -100 dBm	12 %	Using Signal Generator SMB 100A & Spectrum analyzer FSV30
RF Attenuation Measure ⁵ / Generate ⁴	(1 kHz) 1 dB to 60 dB	0.09 dB	Using RF Reference Source 9640A, Signal Generator- Multimeter & Power Meter Method
	(1 kHz to 18 GHz) 1 dB to 60 dB	0.14 dB	
	(18 GHz to 40 GHz) 1 dB to 60 dB	7.4 %	
	(10 MHz to 18 GHz) 60 dB to 110 dB	0.50 dB	Using Signal Generator, attenuator & Spectrum Analyzer Method
	(18 GHz to 29.99 GHz) 1 dB to 110 dB	7.4 %	



SCOPE OF ACCREDITATION

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)^{1,2}

CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ³ (±)	TECHNIQUE, REFERENCE STANDARD, EQUIPMENT
Spectral Purity (THD) Measure ⁵	0.17 % to 3.19 % (No. of harmonics n=2 to n=10, 10 Hz to 2.9 GHz)	0.2 %	Using Spectrum Analyzer FSV-30 Up to 30 GHz by Direct Method
Reflection Coefficient Measure ⁵ / Generate ⁴ - 10 MHz to 18 GHz (Maury Microwave Mismatch Test Set/Network Analyzer)	rho = 0.024 to 0.33	0.032	Using Network Analyzer (R&S ZVB20) by direct method, with Cal kit Z270. Maury Microwave Mismatch Test Set used for verification of Network Analyzer performance as check/transfer standard.
Amplitude Modulation (AM) Measure ⁵ / Generate ⁴ CW: 100 kHz to 3.9 GHz Modulation Rate: 50 Hz to 10 kHz AM Depth	1 % to 98 %	0.25 %	Using Rhode & Schwarz Signal generator, reference source, Spectrum Analyzer FSV 30, Modulation analyzer HP 8901B as transfer by Relative Sideband Amplitude Method
Frequency Modulation (FM) Measure ⁵ / Generate ⁴ CW: 100 kHz to 25 GHz Modulation Rate: 50 Hz to 267 kHz FM Deviation	50 Hz to 4 MHz	0.15 %	Using Rhode & Schwarz Signal generator, reference source, Spectrum Analyzer FSV 30, by Bessel Function Method
Time/Frequency			
Frequency Generate ⁴	1 mHz to 40 GHz	0.2 nHz/Hz + 10 μHz	Using Rubidium Frequency Standard, Signal Generator, RF Reference Source by Direct Method
Frequency Measure ⁵	1 mHz to 29.999 GHz	0.2 nHz/Hz + 10 μHz	Using Rubidium Frequency Standard locked to other Equipment, Frequency Counter 5350B, Spectrum Analyzer -- by Direct Method
Time Interval	100 ms to 86400 s	12 μs/s + 10 μs	Using Timer/counter and oscilloscope by Direct method



SCOPE OF ACCREDITATION

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)^{1,2}

CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ³ (±)	TECHNIQUE, REFERENCE STANDARD, EQUIPMENT
Rotational Speed - Non-Contact	6 rpm	2.2 %	Using Digital Tachometer
	7 rpm to 1000 rpm	0.33 %	
	1001 rpm to 5000 rpm	0.09 %	
	5001 rpm to 20000 rpm	0.07 %	
	20001 rpm to 90000 rpm	0.06 %	
Rotational Speed - Contact	6 rpm	2.2 %	Using Digital Tachometer
	7 rpm to 100 rpm	0.33 %	
	101 rpm to 3000 rpm	0.09 %	
Rotational Speed - Centrifuge	6 rpm	2.2 %	Using Digital Tachometer
	7 rpm to 1000 rpm	0.33 %	
	1001 rpm to 5000 rpm	0.09 %	
	5001 rpm to 20000 rpm	0.07 %	
Optical Radiation			
Illuminance meter	1 lux to 20,000 lux	3.4 %	Using Standard Illuminance Meter
OT Lights (Light Intensity Measurement)	1 lux to 200,000 lux	3.3 %	Using Illuminance Meter

¹The uncertainty covered by the Calibration and Measurement Capability (CMC) is expressed as the expanded uncertainty having a specific coverage probability of approximately 95 %. It is the smallest measurement uncertainty that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than that provided in the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

²If information in this CMC is presented in non-SI units, the conversion factors stated in NIST Special Publication 811 "Guide for the Use of the International System of Units (SI)" apply.

³When uncertainty is stated in relative terms (such as percent, a multiplier expressed as a decimal fraction or in scientific notation), it is in relation to instrument reading or instrument output, as appropriate, unless otherwise indicated.

⁴Capability is suitable for the calibration of measuring devices in the stated ranges.

⁵Capability is suitable for the calibration of devices intended to generate the measurand in the stated ranges.

SG = specific gravity
 PF = power factor
 THD = total harmonic distortion
 CW = continuous wave