



INTERNATIONAL
ACCREDITATION
SERVICE®

CERTIFICATE OF ACCREDITATION

This is to attest

MERIT CALIBRATION

7923 WARNER AVENUE, SUITE K
HUNTINGTON BEACH, CALIFORNIA 92647, U.S.A.

Calibration Laboratory CL-234

has met the requirements of AC204, *IAS Accreditation Criteria for Calibration Laboratories*, and has demonstrated compliance with ISO/IEC Standard 17025:2017, *General requirements for the competence of testing and calibration laboratories*. This organization is accredited to provide the services specified in the scope of accreditation.

Expiration Date December 1, 2026

Effective Date June 3, 2025



International Accreditation Service
Issued under the authority of IAS management

Visit www.iasonline.org for current accreditation information.

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MERIT CALIBRATION

www.meritcalibration.com

Contact Name Brandon Howard

Contact Phone +1 657 259 0597

Accredited to ISO/IEC 17025:2017

Effective Date June 3, 2025

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)*

| MEASURED QUANTITY or DEVICE TYPE CALIBRATED | RANGE | UNCERTAINTY ^{1,2} (±) | CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL) |
|---|------------------|-----------------------------------|--|
| <i>Thermal</i> | | | |
| Temperature Measuring Equipment – | 32 °F to 212 °F | 0.22 °F | Comparison Method MCP-1 RTD, Water Bath, Ice Point |
| Data Logger with External Probe & Sensor | -13 °F to 270 °F | 0.25 °F | Comparison Method MCP-1 RTD, Drywell |

¹The uncertainty covered by the Calibration and Measurement Capability (CMC) is expressed as the expanded uncertainty having a coverage probability of approximately 95 %. It is the smallest measurement uncertainty that a laboratory can achieve within its scope of accreditation when performing calibrations of a best existing device. The measurement uncertainty reported on a calibration certificate may be greater than that provided in the CMC due to the behavior of the calibration item and other factors that may contribute to the uncertainty of a specific calibration.

²When uncertainty is stated in relative terms (such as percent, a multiplier expressed as a decimal fraction or in scientific notation), it is in relation to instrument reading or instrument output, as appropriate, unless otherwise indicated.

** If information in this CMC is presented in non-SI units, the conversion factors stated in NIST Special Publication 811 "Guide for the Use of the International System of Units (SI)" apply.*

CL-234

MERIT CALIBRATION

Effective Date June 3, 2025

Page 2 of 2

IAS/CL/100-3

