



INTERNATIONAL
ACCREDITATION
SERVICE®

CERTIFICATE OF ACCREDITATION

This is to attest

LONG JING TECHNOLOGY LTD.

UNIT 21, 10/F, YALE INDUSTRIAL CENTER, 61-63 AU PUI WAN STREET
FO TAN, HONG KONG 852, SAR

Testing Laboratory TL-1008

has met the requirements of AC89, *IAS Accreditation Criteria for Testing Laboratories*, and has demonstrated compliance with ISO/IEC Standard 17025:2017, *General requirements for the competence of testing and calibration laboratories*. This organization is accredited to provide the services specified in the scope of accreditation.

Effective Date September 29, 2025



International Accreditation Service
Issued under the authority of IAS management

Visit www.iasonline.org for current accreditation information.

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

LONG JING TECHNOLOGY LTD.

www.longjing.com.hk

Contact Name Alfred Au

Contact Phone +852-92749899

Accredited to ISO/IEC 17025:2017

Effective Date September 29, 2025

Face Mask Testing	
ASTM F1862	Standard Test Method for Resistance of Medical Face Masks to Penetration by Synthetic Blood (Horizontal Projection of Fixed Volume at a known Velocity)
ASTM F2100	Standard Specification for Performance of Materials Used in Medical Face Masks
ASTM F2299/F2299M	Standard Test Method for Determining the Initial Efficiency of Materials Used in Medical Face Mask to Penetration by Particulates Using Latex Spheres
EN 14683:2019+AC:2019	Medical face masks- Requirements and test methods
EN 14683:2019+AC:2019 Annex C	Method for Determination of Breathability (Differential Pressure)

TL-1008

LONG JING TECHNOLOGY LTD.

Effective Date September 29, 2025

Page 2 of 2

IAS/TL/100-1

