



# CERTIFICATE OF ACCREDITATION

*This is to attest that*

**UNISERV GULF  
UNITED SERV GULF CO. FOR SERVICES**

AL GHADEER STREET, RIKAZ INDUSTRIAL IBN E SINA AREA, AL SHUKRI COMPLEX  
AL KHOBAR 31952, KINGDOM OF SAUDI ARABIA

**Calibration Laboratory CL-245**

has met the requirements of AC204, *IAS Accreditation Criteria for Calibration Laboratories*, and has demonstrated compliance with ISO/IEC Standard 17025:2017, *General requirements for the competence of testing and calibration laboratories*. This organization is accredited to provide the services specified in the scope of accreditation.

Effective Date January 13, 2022

Expiration Date February 1, 2023



A handwritten signature in black ink, reading 'Raj Nathan'.

**President**

Visit [www.iasonline.org](http://www.iasonline.org) for current accreditation information.

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

## UNISERV GULF UNITED SERV GULF CO. FOR SERVICES

[www.uniservgulf.com](http://www.uniservgulf.com)

**Contact Name** Usman Mirza

**Contact Phone** +966 532073214

*Accredited to ISO/IEC 17025:2017*

*Effective Date January 13, 2022*

### CALIBRATION AND MEASUREMENT CAPABILITY (CMC)\*

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
<i>Mechanical</i>			
Pressure (Pressure gauges, Pressure Switches, Pressure Transmitters, Pressure Transducer, Pressure Recorder, Pressure Relief Valve)	100 psi to 3000 psi 3000 psi to 30000 psi	1.2 % FS 0.6 % FS	Digital pressure gauge & Pressure Pump by comparison method Cal procedure - Doc.Uni008-21/ DKD-R 6-1

<sup>1</sup>The uncertainty covered by the Calibration and Measurement Capability (CMC) is expressed as the expanded uncertainty having a coverage probability of approximately 95 %. It is the smallest measurement uncertainty that a laboratory can achieve within its scope of accreditation when performing calibrations of a best existing device. The measurement uncertainty reported on a calibration certificate may be greater than that provided in the CMC due to the behavior of the calibration item and other factors that may contribute to the uncertainty of a specific calibration.

<sup>2</sup>When uncertainty is stated in relative terms (such as percent, a multiplier expressed as a decimal fraction or in scientific notation), it is in relation to instrument reading or instrument output, as appropriate, unless otherwise indicated.

FS = full scale

*\* If information in this CMC is presented in non-SI units, the conversion factors stated in NIST Special Publication 811 "Guide for the Use of the International System of Units (SI)" apply.*