



INTERNATIONAL
ACCREDITATION
SERVICE®

CERTIFICATE OF ACCREDITATION

This is to attest that

DURAT AL JUBAIL CO.LTD

ALRASHID INDUSTRIAL AREA, P.O. BOX 1514
AL JUBAIL, 31951, SAUDI ARABIA

Calibration Laboratory CL-290

has met the requirements of AC204, *IAS Accreditation Criteria for Calibration Laboratories*, and has demonstrated compliance with ISO/IEC Standard 17025:2017, *General requirements for the competence of testing and calibration laboratories*. This organization is accredited to provide the services specified in the scope of accreditation.

Effective Date September 18, 2024

Expiration Date October 1, 2026



A handwritten signature in black ink, reading 'Raj Nathan'.

President

Visit www.iasonline.org for current accreditation information.

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

DURAT AL JUBAIL CO.LTD

www.djcinspection.com

Contact Name Syed Tasneem Shah

Contact Phone 966-501920842

Accredited to ISO/IEC 17025:2017

Effective Date September 18, 2024

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)*

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Mechanical			
Hydraulic Pressure ⁵ - Pressure Gauges/ Pressure Transmitter/ Pressure Records	10 bar to 650 bar	0.8 bar	Comparison method by using Pressure Module & Pressure Hydraulic Pump
Vacuum Gages	0 inHg to 27 inHg	0.06 inHg	Comparison method by using Pressure Module & Pressure Pneumatic Pump
Pneumatic Pressure ⁵ - Pressure Gauges/ Pressure Transmitter/ Pressure Records	1 bar to 20 bar 20 bar to 40 bar	0.06 bar 0.13 bar	Comparison method by using Pressure Module & Pressure Pneumatic Pump
Torque Wrench	50 N·m to 600 N·m	0.65 N·m	Torque Wrench Calibrator CM1000 by Direct Method
Thermal			
Infrared Thermometers ⁵	-15 °C to 120 °C	0.2 °C	Direct method by using Precision Infrared Calibrator
Temperature Sensors with or Without Indicator ⁵ (RTD/ TC) / Temp Transmitter ⁵ / Temp recorders ⁵	-20 °C to 100 °C 100 °C to 700 °C	1.7 °C 1.1 °C	Comparison method by using RTD/ TC and Field Metrology Well
Oven/ Water Bath/ Dry well ⁵ (Single sensor method)	-20 °C to 100 °C 100 °C to 700 °C	1.7 °C 1.1 °C	Direct method by using RTD/ TC (single sensor)
Electrical – DC/LF			
DC Voltage Generate ³	200 mV to 500 mV 500 mV to 2 V 2 V to 1000 V	0.02 mV 0.02 V 0.12 V	Direct Method by using Clamp Type Multimeter Calibrator
DC Voltage Measure ^{4,5}	1 mV to 100 mV 100 mV to 10 V 10 V to 100 V	5 µV 1 mV 5 mV	Direct Method by using 6.5 Digit Precision Multimeter

* If information in this CMC is presented in non-SI units, the conversion factors stated in NIST Special Publication 811 "Guide for the Use of the International System of Units (SI)" apply.

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
DC Voltage Measure ^{4,5} (continued)	100 V to 1000 V	0.05 V	Direct Method by using 6.5 Digit Precision Multimeter
DC Current Generate ³	10 µA to 400 µA 4 mA to 20 mA 20 mA to 40 mA 40 mA to 2 A 2 A to 10 A 10 A to 100 A	0.05 µA 1 µA 3 µA 1 mA 2 mA 0.02 A	Direct Method by using Clamp Type Multimeter Calibrator
DC Current Measure ^{4,5}	10 µA to 100 µA 100 µA to 10 mA 10 mA to 400 mA	0.006 µA 0.001 mA 0.009 mA	Direct Method by using 6.5 Digit Precision Multimeter
AC Voltage Measure ^{4,5} @ 50 Hz	10 mV to 100 mV 100 mV to 1 V 1 V to 100 V 100 V to 1000 V	0.2 µV 0.12 mV 4.2 mV 0.04 V	Direct Method by using 6.5 Digit Precision Multimeter
AC Voltage Generate ³ @ 50 Hz	200 mV to 2 V 2 V to 10 V 10 V to 1000 V	0.7 mV 1.2 mV 0.04 V	Direct Method by using Clamp Type Multimeter Calibrator
AC Current Generate ³ @ 50 Hz	2 mA to 40 mA 40 mA to 1A 1 A to 10 A 10 A to 1000 A	3 µA 0.4 mA 4.6 mA 0.09 A	Direct Method by using Clamp Type Multimeter Calibrator
AC Current Measure ^{4,5} @ 50 Hz	50 µA to 1 mA 1 mA to 100 mA 100 mA to 400 mA 400 mA to 1 A 1 A to 10 A	0.05 µA 0.05 mA 0.4 mA 1 mA 2 mA	Direct Method by using 6.5 Digit Precision Multimeter
DC Resistance Generate ³	2 Ω to 1 kΩ 1 kΩ to 500 kΩ 500 KΩ to 2 MΩ	0.02 Ω 0.02 kΩ 0.03 MΩ	Direct Method by using Clamp Type Multimeter Calibrator
DC Resistance Measure ⁴	1 Ω to 100 Ω 0.1 kΩ to 10 kΩ 10 kΩ to 100 kΩ 0.1 MΩ to 2 MΩ	0.5 mΩ 0.001 kΩ 0.002 kΩ 0.02 MΩ	Direct Method by using 6.5 Digit Precision Multimeter
Electrical Simulation of Thermocouples – Generate ³ N-type J-Type K-type S-Type	-200 °C to 1300 °C -200 °C to 1200 °C -200 °C to 1372 °C -20 °C to 1767 °C	0.28 °C 0.48 °C 0.17 °C 0.17 °C	Direct method by using Document Process Calibrator
Electrical Simulation of Thermocouples – Measure ⁴ N-type J-Type K-type S-Type	-200 °C to 1300 °C -200 °C to 1200 °C -200 °C to 1372 °C -20 °C to 1767 °C	0.19 °C 0.13 °C 0.13 °C 0.11 °C	Direct method by using Document Process Calibrator

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Electrical Simulation of RTD – Generate ³ Pt 100 (3926) Pt 100 (385) Pt 200 (385) Pt 1000 (385)	-200 °C to 630 °C -200 °C to 800 °C -200 °C to 630 °C -200 °C to 630 °C	0007 °C 0.007 °C 0.006 °C 0.007 °C	Direct method by using Document Process Calibrator
Electrical Simulation of RTD – Measure ⁴ Pt 100 (3926) Pt 100 (385) Pt 200 (385) Pt 1000 (385)	-200 °C to 630 °C -200 °C to 800 °C -200 °C to 630 °C -200 °C to 630 °C	0.009 °C 0.011 °C 0.01 °C 0.009 °C	Direct method by using Document Process Calibrator
Time and Frequency			
Stopwatch	60 s to 3600 s	7.3 s	Comparison method by using Reference Stopwatch
Non-Contact Tachometer	60 rpm to 99999 rpm	1 %	Direct method by using Electrical Process Calibrator
Chemical/Gas			
Multi Gas Detector	O ₂ : 18 % CH ₄ : 2.5 % (50 % LEL) H ₂ S: 25 ppm CO: 100 ppm	1.3 % 1.1 % 0.37 part per 10 ⁶ 5 part per 10 ⁶	Direct method by Using Standard Gas
pH Meters (Discrete Values)	4 pH 7 pH 10 pH	0.12 pH 0.12 pH 0.12 pH	Using Certified Reference Buffer Solutions by Direct Method
Conductivity Meter (Discrete Values)	10 µS/cm 447 µS/cm 1413 µS/cm 2500 µS/cm 2764 µS/cm 15000 µS/cm	0.16 µS/cm 0.03 µS/cm 9.3 µS/cm 15 µS/cm 12 µS/cm 90 µS/cm	Using Certified Reference Buffer Solutions by Direct Method

¹The uncertainty covered by the Calibration and Measurement Capability (CMC) is expressed as the expanded uncertainty having a coverage probability of approximately 95 %. It is the smallest measurement uncertainty that a laboratory can achieve within its scope of accreditation when performing calibrations of a best existing device. The measurement uncertainty reported on a calibration certificate may be greater than that provided in the CMC due to the behavior of the calibration item and other factors that may contribute to the uncertainty of a specific calibration.

²When uncertainty is stated in relative terms (such as percent, a multiplier expressed as a decimal fraction or in scientific notation), it is in relation to instrument reading or instrument output, as appropriate, unless otherwise indicated.

³Capability is suitable for the calibration of measuring devices in the stated ranges.

⁴Capability is suitable for the calibration of devices intended to generate the indicated quantity in the stated ranges.

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

⁵Also available as site calibration. Note that actual measurement uncertainties achievable at a customer's site can normally be expected to be larger than the uncertainties listed on this Scope of Accreditation.

Notes=

ppm= parts per million

LEL = Lower Explosive Limit