



INTERNATIONAL
ACCREDITATION
SERVICE®

CERTIFICATE OF ACCREDITATION

This is to attest

UL LLC

333 PFINGSTEN RD
NORTHBROOK, ILLINOIS, 60062, U.S.A.

Calibration Laboratory CL-310

has met the requirements of AC204, *IAS Accreditation Criteria for Calibration Laboratories*, and has demonstrated compliance with ISO/IEC Standard 17025:2017, *General requirements for the competence of testing and calibration laboratories*. This organization is accredited to provide the services specified in the scope of accreditation.

Expiration Date January 1, 2030

Effective Date March 27, 2026



International Accreditation Service
Issued under the authority of IAS management

Visit www.iasonline.org for current accreditation information.

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

UL LLC

www.website.com

Location	Address	Contact Name	Contact Phone	Scope pages
CL-310 (Main location)	333 Pfingsten Rd. Northbrook, IL 60062, USA	Rick A. Titus Dario Rivoltella	+1-847-664-3281 +39-0292526427	Management System only
CL-163	Via delle Industrie, 6 20061 Carugate (MI), Italy	Giancarlo Alzate	+39-0292526435	2-69
CL-171	Admiral - Rosendahl - Strasse 23 Neu-Isenburg, 63263, Germany	Tobias Berg	+49- 69489810231	70-74

Accredited to ISO/IEC 17025:2017

Effective Date March 27, 2026

CL-163

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)*

METERS ⁵				
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
		U1	U2	
DC Voltage – Generate ³	1 mV to 330 mV 330 mV to 3.3 V 3.3 V to 33 V 33 V to 330 V 330 V to 1000 V	2.0E-05 rel 1.1E-05 rel 1.2E-05 rel 1.8E-05 rel 1.8E-05 rel	1.0 µV/U 2.6 µV/U 26 µV/U 0.19 mV/U 1.5 mV/U	PT-TAR-M-VDC1 Direct Method (Note 1s, 2s)
DC Current - Generate ³	29 µA to 330 µA 330 µA to 3.3 mA 3.3 mA to 33 mA 33 mA to 330 mA 330 mA to 1.1 A 1.1 A to 3 A 3 A to 11 A 11 A to 20 A 20 A to 33 A 33 A to 100 A	1.5E-04 rel 1.0E-04 rel 1.0E-04 rel 1.0E-04 rel 2.0E-04 rel 3.8E-04 rel 5.0E-04 rel 1.0E-03 rel 1.6E-04 rel 1.6E-04 rel	20 nA/I 51 nA/I 0.26 µA/I 2.6 µA/I 40 µA/I 41 µA/I 0.50 mA/I 0.75 mA/I 9.6 mA/I 9.6 mA/I	PT-TAR-M-IDC1 Direct Method (Note 1s, 3s)

* If information in this CMC is presented in non-SI units, the conversion factors stated in NIST Special Publication 811 "Guide for the Use of the International System of Units (SI)" apply.



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

METERS ⁵				
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
		U1	U2	
DC Resistance – Generate ³ DC Resistance – Generate ³ continued	10 mΩ	2.0E-03 rel	0.0 mΩ/R	PT-TAR-M-RDC2W1 PT-TAR-M-RDC4W1 Direct Method (Note 1s, 4s)
	20 mΩ	2.0E-03 rel	0.0 mΩ/R	
	50 mΩ	2.0E-03 rel	0.0 mΩ/R	
	100 mΩ	3.0E-04 rel	0.0 mΩ/R	
	200 mΩ	3.0E-04 rel	0.0 mΩ/R	
	500 mΩ	3.0E-04 rel	0.0 mΩ/R	
	1 Ω to 11 Ω	4.0E-05 rel	1.0 mΩ/R	
	11 Ω to 33 Ω	3.0E-05 rel	1.5 mΩ/R	
	33 Ω to 110 Ω	2.8E-05 rel	1.4 mΩ/R	
	110 Ω to 330 Ω	2.8E-05 rel	2 mΩ/R	
	330 Ω to 1.1 kΩ	2.8E-05 rel	2 mΩ/R	PT-TAR-M-RDC2W1 PT-TAR-M-RDC4W1 Direct Method (Note 1s, 4s)
	1.1 kΩ to 3.3 kΩ	2.8E-05 rel	20 mΩ/R	
	3.3 kΩ to 11 kΩ	2.8E-05 rel	20 mΩ/R	
	11 kΩ to 33 kΩ	2.8E-05 rel	0.2 Ω/R	
	33 kΩ to 110 kΩ	2.8E-05 rel	0.2 Ω/R	
	110 kΩ to 330 kΩ	3.2E-05 rel	2 Ω/R	
	330 kΩ to 1.1 MΩ	3.2E-05 rel	2 Ω/R	
	1.1 MΩ to 3.3 MΩ	6.0E-05 rel	30 Ω/R	
	3.3 MΩ to 11 MΩ	1.3E-04 rel	50 Ω/R	
	11 MΩ to 33 MΩ	2.5E-04 rel	2.5 kΩ/R	
33 MΩ to 110 MΩ	5.0E-04 rel	3.0 kΩ/R		
110 MΩ to 330 MΩ	3.0E-03 rel	0.1 MΩ/R		

POWER METERS ⁵					
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	U1	U2	
DC Power – Generate ³	3.3 V to 33 V	29 μA to 330 μA	1.5E-04 rel	5.8 μW/P	PT-TAR-M-PDC2 Direct Method (Note 1s, 8s)
		330 μA to 3.3 mA	1.0E-04 rel	6.0 μW/P	
3.3 mA to 33 mA		1.0E-04 rel	9.9 μW/P		
33 mA to 330 mA		1.0E-04 rel	99 μW/P		
330 mA to 1.1 A		2.0E-04 rel	1.3 mW/P		
1.1 A to 3 A		3.8E-04 rel	1.3 mW/P		
3 A to 11 A		5.0E-04 rel	17 mW/P		
11 A to 20 A		1.0E-03 rel	25 mW/P		
20 A to 33 A		4.1E-04 rel	1.2 W/P		
33 A to 100 A		4.1E-04 rel	1.2 W/P		
	33 V to 330 V	29 μA to 330 μA	1.5E-04 rel	8.8 μW/P	
		330 μA to 3.3 mA	1.0E-04 rel	18 μW/P	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

POWER METERS ⁵						
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)	
	Voltage	Current	U1	U2		
DC Power – Generate ³ continued		3.3 mA to 33 mA	1.0E-04 rel	0.10 mW/P		
		33 mA to 330 mA	1.0E-04 rel	1.0 mW/P		
		330 mA to 1.1 A	2.0E-04 rel	13 mW/P		
		1.1 A to 3 A	3.8E-04 rel	13 mW/P		
		3 A to 11 A	5.0E-04 rel	0.17 W/P		
		11 A to 20 A	1.0E-03 rel	0.25 W/P		
		20 A to 33 A	4.1E-04 rel	12 W/P		
	33 A to 100 A	4.1E-04 rel	12 W/P			
	330 V to 1000 V	29 µA to 330 µA	1.5E-04 rel	13 µW/P		
		330 µA to 3.3 mA	1.0E-04 rel	31 µW/P		
		3.3 mA to 33 mA	1.0E-04 rel	0.17 mW/P		
		33 mA to 330 mA	1.0E-04 rel	1.7 mW/P		
		330 mA to 1.1 A	2.0E-04 rel	24 mW/P		
		1.1 A to 3 A	3.8E-04 rel	25 mW/P		
3 A to 11 A		5.0E-04 rel	0.30 W/P			
330 V to 1000 V	11 A to 20 A	1.0E-03 rel	0.46 W/P			
	20 A to 33 A	4.1E-04 rel	22 W/P			
	33 A to 100 A	4.1E-04 rel	22 W/P			

Thermocouple Meters /Temperature DataLogger ⁵				
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
		U		
Meters - Type B – Generate ³	600 °C to 800 °C	0.35 °C		PT-TAR-M-TEMP1 Direct Method (Note 17s)
	800 °C to 1550 °C	0.28 °C		
	1550 °C to 1820 °C	0.22 °C		
Meters - Type C - Generate ³	0 °C to 1000 °C	0.16 °C		
	1000 °C to 1800 °C	0.23 °C		
	1800 °C to 2000 °C	0.26 °C		
	2000 °C to 2316 °C	0.35 °C		
Meters - Type E - Generate ³	-250 °C to -200 °C	0.25 °C		
	-200 °C to -100 °C	0.12 °C		
	-100 °C to 0 °C	91 m°C		
	0 °C to 600 °C	81 m°C		
	600 °C to 1000 °C	0.10 °C		
Meters - Type J - Generate ³	-210 °C to -100 °C	0.14 °C		
	-100 °C to 800 °C	91 m°C		
	800 °C to 1200 °C	0.10 °C		



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

POWER METERS ⁵				
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	U1	
Meters - Type K - Generate ³	-250 °C to -200 °C		0.46 °C	
	-200 °C to -100 °C		0.16 °C	
	-100 °C to 500 °C		0.10 °C	
	500 °C to 800 °C		0.10 °C	
	800 °C to 1372 °C		0.13 °C	
Meters - Type L - Generate ³	-200 °C to -100 °C		0.10 °C	
	-100 °C to 900 °C		91 m°C	
Thermocouple Meters /Temperature DataLogger ⁵				
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
			U	
Meters - Type N - Generate ³	-250 °C to -200 °C		0.73 °C	PT-TAR-M-TEMP1 Direct Method (Note 17s)
	-200 °C to -100 °C		0.23 °C	
	-100 °C to 0 °C		0.12 °C	
	0 °C to 100 °C		0.11 °C	
	100 °C to 800 °C		0.10 °C	
	800 °C to 1300 °C		0.12 °C	
Meters - Type R - Generate ³	-50 °C to -25 °C		0.55 °C	
	-25 °C to 0 °C		0.45 °C	
	0 °C to 100 °C		0.39 °C	
	100 °C to 400 °C		0.28 °C	
	400 °C to 600 °C		0.22 °C	
	600 °C to 1000 °C		0.21 °C	
	1000 °C to 1600 °C		0.19 °C	
Meters - Type S - Generate ³	-50 °C to -25 °C		0.51 °C	
	-25 °C to 0 °C		0.43 °C	
	0 °C to 100 °C		0.38 °C	
	100 °C to 400 °C		0.29 °C	
	400 °C to 600 °C		0.23 °C	
	600 °C to 1000 °C		0.22 °C	
	1000 °C to 1600 °C		0.22 °C	
Meters - Type T - Generate ³	-250 °C to -200 °C		0.35 °C	
	-200 °C to -100 °C		0.16 °C	
	-100 °C to 0 °C		0.11 °C	
	0 °C to 200 °C		91 m°C	
	200 °C to 400 °C		91 m°C	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Thermocouple Meters /Temperature DataLogger ⁵				
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
		U		
Meters - Type U - Generate ³	-200 °C to 0 °C 0 °C to 200 °C 200 °C to 600 °C	0.16 °C 0.10 °C 0.10 °C		

Meters ⁵					
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	CAPACITANCE	FREQUENCY	U1	U2	
Capacitance – Generate ³	220 pF to 399.9 pF	10 Hz to 10 kHz	5.0E-03 rel	0.01 nF	PT-TAR-M-C1 Direct Method (Note 1s, 25s, 27s)
	0.4 nF to 1.0999 nF	10 Hz to 10 kHz	5.0E-03 rel	0.01 nF	
	1.1 nF to 3.2999 nF	10 Hz to 3 kHz	5.0E-03 rel	0.01 nF	
	3.3 nF to 10.9999 nF	10 Hz to 1 kHz	2.5E-03 rel	0.01 nF	
	11 nF to 32.9999 nF	10 Hz to 1 kHz	2.5E-03 rel	0.12 nF	
	33 nF to 109.999 nF	10 Hz to 1 kHz	2.5E-03 rel	0.12 nF	
	110 nF to 329.999 nF	10 Hz to 1 kHz	2.5E-03 rel	0.65 nF	
	0.33 µF to 1.09999 µF	10 Hz to 600 Hz	2.5E-03 rel	1.2 nF	
	1.1 µF to 3.29999 µF	10 Hz to 300 Hz	2.5E-03 rel	6.5 nF	
	3.3 µF to 10.9999 µF	10 Hz to 150 Hz	2.5E-03 rel	12 nF	
	11 µF to 32.9999 µF	10 Hz to 120 Hz	4.0E-03 rel	65 nF	
	33 µF to 109.999 µF	10 Hz to 80 Hz	4.5E-03 rel	0.12 µF	
	110 µF to 329.999 µF	0 Hz to 50 Hz	4.5E-03 rel	0.65 µF	

Meters ⁵					
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Frequency	U1	U2	
AC Voltage – Generate ³	10 mV to 33 mV	45 Hz to 10 kHz	1.5E-04 rel	6.0 µV/U	PT-TAR-M- VAC1 Direct Method
		10 kHz to 20 kHz	2.0E-04 rel	6.0 µV/U	
		20 kHz to 50 kHz	1.0E-03 rel	6.0 µV/U	
		50 kHz to 100 kHz	3.5E-03 rel	12 µV/U	
		100 kHz to 4 MHz	2.0E-02 rel	5.8 µV/U	
		4 MHz to 20 MHz	2.0E-02 rel	5.8 µV/U	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Meters ⁵					
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Frequency	U1	U2	
AC Voltage – Generate ³ continued	33 mV to 330 mV	45 Hz to 10 kHz	3.0E-04 rel	8.1 µV/U	(Note 1s, 2s)
		10 kHz to 20 kHz	3.6E-04 rel	8.1 µV/U	
		20 kHz to 50 kHz	3.5E-04 rel	8.1 µV/U	
		50 kHz to 100 kHz	8.0E-04 rel	32 µV/U	
		100 kHz to 4 MHz	2.0E-02 rel	58 µV/U	
330 mV to 3.3 V	45 Hz to 10 kHz	45 Hz to 10 kHz	8.9E-01 rel	61 µV/U	PT-TAR-M-VAC1 Direct Method
		10 kHz to 20 kHz	1.9E-04 rel	61 µV/U	
		20 kHz to 50 kHz	3.0E-04 rel	51 µV/U	
		50 kHz to 100 kHz	7.0E-04 rel	0.13 mV/U	
		100 kHz to 500 kHz	2.4E-03 rel	0.60 mV/U	
3.3 V to 33 V	45 Hz to 10 kHz	45 Hz to 10 kHz	1.5E-04 rel	0.61 mV/U	(Note 1s, 2s)
		10 kHz to 20 kHz	2.4E-04 rel	0.61 mV/U	
		20 kHz to 50 kHz	3.5E-04 rel	0.61 mV/U	
		50 kHz to 100 kHz	9.0E-04 rel	1.6 mV/U	
33 V to 330 V	45 Hz to 1 kHz	45 Hz to 1 kHz	1.9E-04 rel	2.1 mV/U	(Note 1s, 2s)
		1 kHz to 10 kHz	2.0E-04 rel	6.1 mV/U	
		10 kHz to 20 kHz	2.5E-04 rel	6.1 mV/U	
		20 kHz to 50 kHz	3.0E-04 rel	6.1 mV/U	
330 V to 1000 V	45 Hz to 1 kHz	45 Hz to 1 kHz	3.0E-04 rel	10 mV/U	(Note 1s, 2s)
		1 kHz to 5 kHz	2.5E-04 rel	10 mV/U	
		5 kHz to 10 kHz	3.0E-04 rel	10 mV/U	

Meters ⁵					
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Current	Frequency	U1	U2	
AC Current – Generate ³	29 µA to 330 µA	45 Hz to 1 kHz	1.3E-03 rel	0.10 µA/I	PT-TAR-M-IAC1 Direct Method
	330 µA to 3.3 mA	45 Hz to 1 kHz	1.0E-03 rel	0.15 µA/I	
		1 kHz to 5 kHz	2.0E-03 rel	0.20 µA/I	
3.3 mA to 33 mA	45 Hz to 1 kHz	45 Hz to 1 kHz	4.0E-04 rel	2.0 µA/I	(Note 1s, 3s)
		1 kHz to 5 kHz	8.0E-04 rel	2.0 µA/I	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Meters ⁵					
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Current	Frequency	U1	U2	
AC Current – Generate ³ (continued)	33 mA to 330 mA	45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 6 kHz 6 kHz to 10 kHz	4.0E-04 rel 1.0E-03 rel 7.8E-03 rel 1.5E-02 rel	20 µA/I 50 µA/I 25 mA/I 62 mA/I	PT-TAR-M-IAC1 Direct Method (Note 1s, 3s)
	330 mA to 1.1 A	45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 6 kHz 6 kHz to 10 kHz	5.0E-04 rel 6.0E-03 rel 7.8E-03 rel 1.5E-02 rel	0.10 mA/I 1.0 mA/I 25 mA/I 62 mA/I	
	1.1 A to 3 A	45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 6 kHz 6 kHz to 10 kHz	6.0E-04 rel 6.0E-03 rel 7.8E-03 rel 1.5E-02 rel	0.10 mA/I 1.0 mA/I 25 mA/I 62 mA/I	
	3 A to 11 A	45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 6 kHz 6 kHz to 10 kHz	6.0E-04 rel 1.0E-03 rel 7.8E-03 rel 2.3E-02 rel	2.0 mA/I 2.0 mA/I 62 mA/I 94 mA/I	
	11 A to 20 A	45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 6 kHz 6 kHz to 10 kHz	1.2E-03 rel 1.5E-03 rel 7.8E-03 rel 2.3E-02 rel	5.0 mA/I 5.0 mA/I 62 mA/I 94 mA/I	
	20 A to 33 A	45 Hz to 65 Hz 65 Hz to 300 Hz 300 Hz to 1 kHz 1 kHz to 3 kHz 3 kHz to 6 kHz 6 kHz to 10 kHz	1.9E-04 rel 2.8E-04 rel 8.0E-04 rel 2.3E-03 rel 7.8E-03 rel 3.1E-02 rel	19 mA/I 28 mA/I 94 mA/I 0.23 A/I 0.42 A/I 0.70 A/I	
	33 A to 120 A	45 Hz to 65 Hz 65 Hz to 300 Hz 300 Hz to 1 kHz 1 kHz to 3 kHz 3 kHz to 6 kHz 6 kHz to 10 kHz	1.9E-04 rel 2.8E-04 rel 8.0E-04 rel 2.3E-03 rel 7.8E-03 rel 3.1E-02 rel	19 mA/I 28 mA/I 94 mA/I 0.23 A/I 0.42 A/I 0.70 A/I	

Meters ⁵					
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Frequency	Voltage	U1	U2	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Frequency (Voltage) – Generate ³	0.01 Hz to 10 Hz	1 mV to 3.3 V	2.6E-06 rel	7.7 µHz/f	PT-TAR-M-F1 Direct Method (Note 1s, 5s)
	10 Hz to 45 Hz	1 mV to 33 V	2.5E-06 rel	7.7 µHz/f	
	45 Hz to 100 Hz	1 mV to 1000 V	2.5E-06 rel	7.7 µHz/f	
	100 Hz to 1199.9 Hz		2.5E-06 rel	7.7 µHz/f	
	1.2 kHz to 10 kHz		2.6E-06 rel	5.8 mHz/f	
	10 kHz to 100 kHz	1 mV to 329.99 V	2.5E-06 rel	5.8 mHz/f	
	100 kHz to 1199.9 kHz	1 mV to 3.3 V	2.5E-06 rel	5.8 mHz/f	
	1.2 MHz to 2 MHz		5.1E-06 rel	5.8 Hz/f	
2 MHz to 20 MHz	1.5E-03 rel		5.8 kHz/f		

Meters ⁵						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Frequency	Crest Factor	U1	U2	
AC Voltage Crest Factor – Measure ⁴	6 V to 60 V	50 Hz to 60 Hz	1.0 to 2.0	1.7 E-02 rel	5.8E-07 Unit	PT-TAR-M-CF1 Direct Method (Note 1s)
	60 V to 270 V	50 Hz to 60 Hz	1.0 to 1.6	1.9 E-02 rel	5.8E-07 Unit	

Power meters ⁵						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	PF	U1	U2	
Power Factor @ 50/60 Hz – Generate ³	3.3 V to 1000 V	29 µA to 20 A	0.10	0.0E+00 rel	2.6E-03 unit/PF	PT-TAR-M-PF2 Direct Method (Note 1s, 7s, 20s, 24s)
			0.35	0.0E+00 rel	2.5E-03 unit/PF	
			0.50	0.0E+00 rel	2.3E-03 unit/PF	
			0.70	0.0E+00 rel	1.9E-03 unit/PF	
			0.90	0.0E+00 rel	1.1E-03 unit/PF	
			1.00	0.0E+00 rel	5.8E-06 unit/PF	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Power meters ⁵						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	cos(φ)	U1	U2	
AC Active Power @ 50/60 Hz – Generate ³	3.3 V to 33 V	29 µA to 330 µA	0.1	1.7E-02 rel	0.32 µW/P	PT-TAR-M-PAC2 Direct Method (Note 1s, 8s)
			0.5	3.3E-03 rel	1.6 µW/P	
			1.0	1.3E-03 rel	3.2 µW/P	
		330 µA to 3.3 mA	0.1	1.7E-02 rel	0.52 µW/P	
			0.5	3.2E-03 rel	2.6 µW/P	
			1.0	1.0E-03 rel	5.3 µW/P	
		3.3 mA to 33 mA	0.1	1.7E-02 rel	8.9 µW/P	
			0.5	3.1E-03 rel	34 µW/P	
			1.0	4.3E-04 rel	68 µW/P	
		33 mA to 330 mA	0.1	1.7E-02 rel	68 µW/P	
0.5	3.1E-03 rel		0.34 mW/P			
1.0	4.3E-04 rel		0.68 mW/P			
330 mA to 1.1 A	0.1	1.7E-02 rel	0.33 mW/P			
	0.5	3.1E-03 rel	1.6 mW/P			
	1.0	5.2E-04 rel	3.3 mW/P			
1.1 A to 3 A	0.1	1.7E-02 rel	0.37 mW/P			
	0.5	3.1E-03 rel	1.8 mW/P			
	1.0	6.2E-04 rel	3.7 mW/P			
3 A to 11 A	0.1	1.7E-02 rel	8.7 mW/P			
	0.5	3.1E-03 rel	33 mW/P			
	1.0	6.2E-04 rel	65 mW/P			
11 A to 20 A	0.1	1.7E-02 rel	17 mW/P			
	0.5	3.3E-03 rel	81 mW/P			
	1.0	1.2E-03 rel	0.16 W/P			
20 A to 33 A	0.1	3.5E-02 rel	0.15 W/P			
	0.5	6.2E-03 rel	0.75 W/P			
	1.0	1.0E-03 rel	1.5 W/P			
33 A to 50 A	0.1	3.5E-02 rel	0.17 W/P			
	0.5	6.2E-03 rel	0.86 W/P			
	1.0	9.3E-04 rel	1.7 W/P			



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Power meters ⁵						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	cos(φ)	U1	U2	
AC Active Power @ 50/60 Hz – Generate ³	33 V to 330 V	29 μA to 330 μA	0.1	1.8E-02 rel	6.7 μW/P	PT-TAR-M-PAC2 Direct Method (Note 1s, 8s)
			0.5	3.3E-03 rel	18 μW/P	
			1.0	1.3E-03 rel	34 μW/P	
		330 μA to 3.3 mA	0.1	1.7E-02 rel	7.6 μW/P	
			0.5	3.2E-03 rel	26 μW/P	
			1.0	1.0E-03 rel	50 μW/P	
		3.3 mA to 33 mA	0.1	1.7E-02 rel	67 μW/P	
			0.5	3.1E-03 rel	0.33 mW/P	
			1.0	4.5E-04 rel	0.67 mW/P	
		33 mA to 330 mA	0.1	1.7E-02 rel	0.67 mW/P	
0.5	3.1E-03 rel		3.3 mW/P			
1.0	4.5E-04 rel		6.7 mW/P			
330 mA to 1.1 A	0.1	1.7E-02 rel	3.3 mW/P			
	0.5	3.1E-03 rel	17 mW/P			
	1.0	5.4E-04 rel	33 mW/P			
1.1 A to 3 A	0.1	1.7E-02 rel	3.4 mW/P			
	0.5	3.1E-03 rel	17 mW/P			
	1.0	6.3E-04 rel	34 mW/P			
3 A to 11 A	0.1	1.7E-02 rel	66 mW/P			
	0.5	3.1E-03 rel	0.33 W/P			
	1.0	6.3E-04 rel	0.66 W/P			
11 A to 20 A	0.1	1.7E-02 rel	0.17 W/P			
	0.5	3.3E-03 rel	0.83 W/P			
	1.0	1.2E-03 rel	1.7 W/P			
20 A to 33 A	0.1	3.5E-02 rel	1.6 W/P			
	0.5	6.2E-03 rel	7.8 W/P			
	1.0	1.0E-03 rel	16 W/P			
33 A to 50 A	0.1	3.5E-02 rel	1.8 W/P			
	0.5	6.2E-03 rel	8.8 W/P			
	1.0	9.4E-04 rel	18 W/P			



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Power meters ⁵						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	cos(φ)	U1	U2	
AC Active Power @ 50/60 Hz – Generate ³	330 V to 1000 V	29 µA to 330 µA	0.1	1.8E-02 rel	8.4 µW/P	PT-TAR-M-PAC2 Direct Method (Note 1s, 8s)
			0.5	3.3E-03 rel	31 µW/P	
			1.0	1.3E-03 rel	61 µW/P	
		330 µA to 3.3 mA	0.1	1.7E-02 rel	11 µW/P	
			0.5	3.2E-03 rel	49 µW/P	
			1.0	1.0E-03 rel	96 µW/P	
		3.3 mA to 33 mA	0.1	1.7E-02 rel	0.13 mW/P	
			0.5	3.1E-03 rel	0.63 mW/P	
			1.0	5.0E-04 rel	1.3 mW/P	
		33 mA to 330 mA	0.1	1.7E-02 rel	1.3 mW/P	
0.5	3.1E-03 rel		6.3 mW/P			
1.0	5.0E-04 rel		13 mW/P			
330 mA to 1.1 A	0.1	1.7E-02 rel	6.1 mW/P			
	0.5	3.1E-03 rel	31 mW/P			
	1.0	5.9E-04 rel	61 mW/P			
1.1 A to 3 A	0.1	1.7E-02 rel	8.8 mW/P			
	0.5	3.1E-03 rel	34 mW/P			
	1.0	6.7E-04 rel	67 mW/P			
3 A to 11 A	0.1	1.7E-02 rel	0.12 W/P			
	0.5	3.1E-03 rel	0.61 W/P			
	1.0	6.7E-04 rel	1.2 W/P			
11 A to 20 A	0.1	1.7E-02 rel	0.30 W/P			
	0.5	3.3E-03 rel	1.5 W/P			
	1.0	1.2E-03 rel	3.0 W/P			
20 A to 33 A	0.1	3.5E-02 rel	2.8 W/P			
	0.5	6.2E-03 rel	14 W/P			
	1.0	1.1E-03 rel	28 W/P			
33 A to 50 A	0.1	3.5E-02 rel	3.2 W/P			
	0.5	6.2E-03 rel	16 W/P			
	1.0	9.7E-04 rel	32 W/P			



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Power Meters ⁵						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	cos(φ)	U1	U2	
AC Active Energy @ 50/60Hz – PF @ 1.0 or 0.5 – 10min Integration - Generate ³	3.3 V to 33 V	29 μA to 330 μA	1.0 0.5	1.3E-03 rel 3.3E-03 rel	0.54 μWh/E 0.27 μWh/E	PT-TAR-M-E2 Direct Method (Note 1s, 11s)
		330 μA to 3.3 mA	1.0 0.5	1.0E-03 rel 3.2E-03 rel	0.87 μWh/E 0.43 μWh/E	
		3.3 mA to 33 mA	1.0 0.5	4.3E-04 rel 3.1E-03 rel	13 μWh/E 8.1 μWh/E	
		33 mA to 330 mA	1.0 0.5	4.3E-04 rel 3.1E-03 rel	0.11 mWh/E 56 μWh/E	
		330 mA to 1.1 A	1.0 0.5	5.2E-04 rel 3.1E-03 rel	0.55 mWh/E 0.27 mWh/E	
		1.1 A to 3 A	1.0 0.5	6.2E-04 rel 3.1E-03 rel	0.61 mWh/E 0.31 mWh/E	
		3 A to 11 A	1.0 0.5	6.2E-04 rel 3.1E-03 rel	12 mWh/E 7.9 mWh/E	
		11 A to 20 A	1.0 0.5	1.2E-03 rel 3.3E-03 rel	27 mWh/E 15 mWh/E	
		20 A to 33 A	1.0 0.5	1.0E-03 rel 6.2E-03 rel	0.25 Wh/E 0.13 Wh/E	
		33 A to 50 A	1.0 0.5	9.3E-04 rel 6.2E-03 rel	0.29 Wh/E 0.14 Wh/E	
	33 V to 330 V	29 μA to 330 μA	1.0 0.5	1.3E-03 rel 3.3E-03 rel	8.0 μWh/E 6.4 μWh/E	
		330 μA to 3.3 mA	1.0 0.5	1.0E-03 rel 3.2E-03 rel	10 μWh/E 7.1 μWh/E	
		3.3 mA to 33 mA	1.0 0.5	4.5E-04 rel 3.1E-03 rel	0.11 mWh/E 56 μWh/E	
		33 mA to 330 mA	1.0 0.5	4.5E-04 rel 3.1E-03 rel	1.1 mWh/E 0.55 mWh/E	
		330 mA to 1.1 A	1.0 0.5	5.4E-04 rel 3.1E-03 rel	5.5 mWh/E 2.8 mWh/E	
		1.1 A to 3 A	1.0 0.5	6.3E-04 rel 3.1E-03 rel	5.6 mWh/E 2.8 mWh/E	
		3 A to 11 A	1.0 0.5	6.3E-04 rel 3.1E-03 rel	0.11 Wh/E 55 mWh/E	
		11 A to 20 A	1.0 0.5	1.2E-03 rel 3.3E-03 rel	0.28 Wh/E 0.14 Wh/E	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Power Meters ⁵						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	cos(φ)	U1	U2	
AC Active Energy @ 50/60Hz – PF @ 1.0 or 0.5 – 10min Integration - Generate ³ (continued)	33 V to 330 V	20 A to 33 A	1.0	1.0E-03 rel	2.6 Wh/E	PT-TAR-M-E2 Direct Method (Note 1s, 11s)
			0.5	6.2E-03 rel	1.3 Wh/E	
		33 A to 50 A	1.0	9.4E-04 rel	2.9 Wh/E	
			0.5	6.2E-03 rel	1.5 Wh/E	
	330 V to 1000V	29 μA to 330 μA	1.0	1.3E-03 rel	12 μWh/E	
			0.5	3.3E-03 rel	7.7 μWh/E	
		330 μA to 3.3 mA	1.0	1.0E-03 rel	17 μWh/E	
			0.5	3.2E-03 rel	9.9 μWh/E	
		3.3 mA to 33 mA	1.0	5.0E-04 rel	0.21 mWh/E	
			0.5	3.1E-03 rel	0.10 mWh/E	
		33 mA to 330 mA	1.0	5.0E-04 rel	2.1 mWh/E	
			0.5	3.1E-03 rel	1.0 mWh/E	
		330 mA to 1.1 A	1.0	5.9E-04 rel	10 mWh/E	
			0.5	3.1E-03 rel	5.1 mWh/E	
1.1 A to 3 A	1.0	6.7E-04 rel	13 mWh/E			
	0.5	3.1E-03 rel	8.0 mWh/E			
3 A to 11 A	1.0	6.7E-04 rel	0.20 Wh/E			
	0.5	3.1E-03 rel	0.10 Wh/E			
11 A to 20 A	1.0	1.2E-03 rel	0.50 Wh/E			
	0.5	3.3E-03 rel	0.25 Wh/E			
20 A to 33 A	1.0	1.1E-03 rel	4.7 Wh/E			
	0.5	6.2E-03 rel	2.4 Wh/E			
33 A to 50 A	1.0	9.7E-04 rel	5.4 Wh/E			
	0.5	6.2E-03 rel	2.7 Wh/E			



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Power meters ⁵						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	sin(φ)	U1	U2	
AC Reactive Power @ 50/60 Hz – Generate ³	3.3 V to 33 V	33 mA to 330 mA	0.1	4.6E-04 rel	68 μvar/Q	PT-TAR-M-Q2 Direct Method (Note 1s, 9s)
			0.5	1.1E-03 rel	0.34 mvar/Q	
			0.9	3.6E-03 rel	0.61 mvar/Q	
		330 mA to 1.1 A	0.1	5.5E-04 rel	0.33 mvar/Q	
			0.5	1.1E-03 rel	1.6 mvar/Q	
		1.1 A to 3 A	0.1	6.5E-04 rel	0.37 mvar/Q	
			0.5	1.2E-03 rel	1.8 mvar/Q	
	3 A to 11 A	0.1	6.5E-04 rel	8.7 mvar/Q		
		0.5	1.2E-03 rel	33 mvar/Q		
	11 A to 20 A	0.1	1.2E-03 rel	17 mvar/Q		
		0.5	1.6E-03 rel	81 mvar/Q		
	20 A to 33 A	0.1	1.1E-03 rel	0.15 var/Q		
		0.5	2.3E-03 rel	0.75 var/Q		
	33 A to 50 A	0.1	1.0E-03 rel	0.17 var/Q		
0.5		2.2E-03 rel	0.86 var/Q			
33 V to 330 V	29 μA to 330 μA	0.1	1.3E-03 rel	6.7 μvar/Q	PT-TAR-M-Q2 Direct Method (Note 1s, 9s)	
		0.5	1.6E-03 rel	18 μvar/Q		
		0.9	3.8E-03 rel	30 μvar/Q		
		330 μA to 3.3 mA	0.1	1.0E-03 rel		7.6 μvar/Q
			0.5	1.4E-03 rel		26 μvar/Q
		3.3 mA to 33 mA	0.1	4.8E-04 rel		67 μvar/Q
	0.5		1.1E-03 rel	0.33 mvar/Q		
	33 mA to 330 mA	0.1	4.8E-04 rel	0.67 mvar/Q		
		0.5	1.1E-03 rel	3.3 mvar/Q		
	330 mA to 1.1 A	0.1	5.7E-04 rel	3.3 mvar/Q		
		0.5	1.1E-03 rel	17 mvar/Q		
	1.1 A to 3 A	0.1	6.6E-04 rel	3.4 mvar/Q		
		0.5	1.2E-03 rel	17 mvar/Q		
			0.9	3.7E-03 rel		30 mvar/Q



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Power meters ⁵						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	sin(φ)	U1	U2	
AC Reactive Power @ 50/60 Hz – Generate ³ (continued)	33 V to 330 V	3 A to 11 A	0.1	6.6E-04 rel	66 mvar/Q	PT-TAR-M-Q2 Direct Method (Note 1s, 9s)
			0.5	1.2E-03 rel	0.33 var/Q	
			0.9	3.7E-03 rel	0.60 var/Q	
		11 A to 20 A	0.1	1.2E-03 rel	0.17 var/Q	
	0.5		1.6E-03 rel	0.83 var/Q		
	0.9		3.8E-03 rel	1.5 var/Q		
	20 A to 33 A	0.1	1.1E-03 rel	1.6 var/Q		
		0.5	2.3E-03 rel	7.8 var/Q		
		0.9	7.3E-03 rel	14 var/Q		
	33 A to 50 A	0.1	1.0E-03 rel	1.8 var/Q		
		0.5	2.2E-03 rel	8.8 var/Q		
		0.9	7.3E-03 rel	16 var/Q		
	330 V to 1000 V	29 µA to 330 µA	0.1	1.3E-03 rel	8.4 µvar/Q	
			0.5	1.6E-03 rel	31 µvar/Q	
			0.9	3.8E-03 rel	55 µvar/Q	
		330 µA to 3.3 mA	0.1	1.1E-03 rel	11 µvar/Q	
			0.5	1.5E-03 rel	49 µvar/Q	
			0.9	3.8E-03 rel	87 µvar/Q	
3.3 mA to 33 mA		0.1	5.3E-04 rel	0.13 mvar/Q		
		0.5	1.1E-03 rel	0.63 mvar/Q		
		0.9	3.7E-03 rel	1.1 mvar/Q		
33 mA to 330 mA		0.1	5.3E-04 rel	1.3 mvar/Q		
		0.5	1.1E-03 rel	6.3 mvar/Q		
		0.9	3.7E-03 rel	11 mvar/Q		
330 mA to 1.1 A		0.1	6.1E-04 rel	6.1 mvar/Q		
		0.5	1.2E-03 rel	31 mvar/Q		
	0.9	3.7E-03 rel	55 mvar/Q			
1.1 A to 3 A	0.1	7.0E-04 rel	8.8 mvar/Q			
	0.5	1.2E-03 rel	34 mvar/Q			
	0.9	3.7E-03 rel	61 mvar/Q			
3 A to 11 A	0.1	7.0E-04 rel	0.12 var/Q			
	0.5	1.2E-03 rel	0.61 var/Q			
	0.9	3.7E-03 rel	1.1 var/Q			
11 A to 20 A	0.1	1.3E-03 rel	0.30 var/Q			
	0.5	1.6E-03 rel	1.5 var/Q			
	0.9	3.8E-03 rel	2.7 var/Q			
20 A to 33 A	0.1	1.1E-03 rel	2.8 var/Q			
	0.5	2.3E-03 rel	14 var/Q			
	0.9	7.3E-03 rel	25 var/Q			



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Power meters ⁵						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	sin(φ)	U1	U2	
AC Reactive Power @ 50/60 Hz – Generate ³ (continued)	330 V to 1000 V	33 A to 50 A	0.1 0.5 0.9	1.0E-03 rel 2.2E-03 rel 7.3E-03 rel	3.2 var/Q 16 var/Q 29 var/Q	PT-TAR-M-Q2 Direct Method Fluke 5522A, Fluke 52120A, Clarke Hess 6000A Clarke Hess 650 Clarke Hess 610 (Note 1s, 9s)

Power Meters ⁵						
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)	
	Voltage	Current	U1	U2		
AC Apparent Power @ 50/60 Hz – PF @ 0.1 to 1.0 Generate ³	3.3 V to 33 V	29 µA to 330 µA	1.3E-03 rel	3.2 µVA/S	PT-TAR-M-S2 Direct Method (Note 1s, 10s)	
		330 µA to 3.3 mA	1.0E-03 rel	5.3 µVA/S		
		3.3 mA to 33 mA	4.3E-04 rel	68 µVA/S		
		33 mA to 330 mA	4.3E-04 rel	0.68 mVA/S		
		330 mA to 1.1 A	5.2E-04 rel	3.3 mVA/S		
		1.1 A to 3 A	6.2E-04 rel	3.7 mVA/S		
		3 A to 11 A	6.2E-04 rel	65 mVA/S		
		11 A to 20 A	1.2E-03 rel	0.16 VA/S		
		20 A to 33 A	1.0E-03 rel	1.3 VA/S		
33 V to 330 V	29 µA to 330 µA	1.3E-03 rel	34 µVA/S			
	330 µA to 3.3 mA	1.0E-03 rel	50 µVA/S			
	3.3 mA to 33 mA	4.5E-04 rel	0.67 mVA/S			
	33 mA to 330 mA	4.5E-04 rel	6.7 mVA/S			
	330 mA to 1.1 A	5.4E-04 rel	33 mVA/S			
	1.1 A to 3 A	6.3E-04 rel	34 mVA/S			
	3 A to 11 A	6.3E-04 rel	0.66 VA/S			
	11 A to 20 A	1.2E-03 rel	1.7 VA/S			
	20 A to 33 A	1.0E-03 rel	14 VA/S			
330 V to 1000 V	29 µA to 330 µA	1.3E-03 rel	61 µVA/S			
	330 µA to 3.3 mA	1.0E-03 rel	96 µVA/S			
	3.3 mA to 33 mA	5.0E-04 rel	1.3 mVA/S			
	33 mA to 330 mA	5.0E-04 rel	13 mVA/S			
	330 mA to 1.1 A	5.9E-04 rel	61 mVA/S			



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Power Meters ⁵						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current		U1	U2	
AC Apparent Power @ 50/60 Hz – PF @ 0.1 to 1.0 Generate ³	330 V to 1000 V	1.1 A to 3 A 3 A to 11 A 11 A to 20 A 20 A to 33 A 33 A to 120 A		6.7E-04 rel 6.7E-04 rel 1.2E-03 rel 1.1E-03 rel 9.6E-04 rel	67 mVA/S 1.2 VA/S 3.0 VA/S 25 VA/S 30 VA/S	PT-TAR-M-S2 Direct Method (Note 1s, 10s)

Voltage Harmonics, Interharmonics and THD meters ⁵						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Par	IHRM Amplitude	IHRM Frequency	U1	U2	
AC Voltage Harmonics and Interharmonics – Measure ⁴	Fund	30 V to 1000 V	50 Hz/60 Hz	5.0E-03 rel	58 µV/U	PT-TAR-M-HRMV1 PT-TAR-M-HRMV2
	IHRM11 to IHRM2000	0.3 V to 60 V	55 Hz to 12 kHz	1.0E-02 rel	58 µV/U	
	THD _v	Fund: 30 V to 1000 V	55 Hz to 12 kHz	7.5E-03 rel	58 µ% _{THD} /THD _v	IEC 61000-4-7 (Note 1s, 2s, 12s)

Current Harmonics, Interharmonics and THD meters ⁵						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Par	IHRM Amplitude	IHRM Frequency	U1	U2	
AC Current Harmonics and Interharmonics – Measure ⁴	Fund	50 mA to 100 A	50 Hz/60 Hz	2.0E-02 rel	0.58 µA/I	PT-TAR-M-HRMI1 PT-TAR-M-HRMI2
	IHRM11 to IHRM2000	4.5 mA to 135 mA (TotRMSmax: 500 mA)	55 Hz to 3 kHz	2.5E-02 rel	0.58 µA/I	
			3 kHz to 12 kHz	6.0E-02 rel	0.58 µA/I	
		50 mA to 1.35 A (TotRMSmax: 5 A)	55 Hz to 3 kHz	2.5E-02 rel	0.58 µA/I	IEC 61000-4-7 (Note 1s, 3s, 12s)
			3 kHz to 12 kHz	6.0E-02 rel	0.58 µA/I	
300 mA to 9 A (TotRMSmax: 32 A)	55 Hz to 3 kHz	2.5E-02 rel	0.58 µA/I			



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

			3 kHz to 12 kHz	6.0E-02 rel	0.58 µA/I
		1 A to 30 A (TotRMSmax: 100 A)	55 Hz to 3 kHz	2.5E-02 rel	0.58 µA/I
			3 kHz to 12 kHz	9.0E-02 rel	0.58 µA/I
	THD _i	Fund: 50 mA to 100 A	55 Hz to 12 kHz	2.5E-02 rel	58 µ% _{THD} /THD _i

Current Clamps ⁵					
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)	
		U1	U2		
DC Current - Generate ³	330 mA to 1.1 A	8.1E-03 rel	16 mA/I	PT-TAR-M-PA-IDC1 PT-TAR-M-PA-IDC2 Direct Method (Note 1s, 3s)	
	1.1 A to 3 A	3.5E-03 rel	17 mA/I		
	3 A to 11 A	1.4E-03 rel	17 mA/I		
	11 A to 20 A	1.3E-03 rel	17 mA/I		
	20 A to 33 A	8.6E-04 rel	40 mA/I		
	33 A to 100 A	8.6E-04 rel	40 mA/I		
	100 A to 150 A	9.3E-04 rel	0.67 A/I		
	150 A to 550 A	5.7E-04 rel	0.68 A/I		
	550 A to 1000 A	1.0E-03 rel	0.67 A/I		
	1000 A to 2500 A	8.6E-04 rel	1.2 A/I		

Clamp-on power meters - Power meters chained with current clamp ⁵					
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	U1	U2	
DC Power – Generate ³	3.3 V to 33 V	330 mA to 1.1 A	2.0E-04 rel	1.3 mW/P	PT-TAR-M-PA-PDC1 Direct Method (Note 1s, 8s)
		1.1 A to 3 A	3.8E-04 rel	1.3 mW/P	
		3 A to 11 A	5.0E-04 rel	18 mW/P	
		11 A to 20 A	1.0E-03 rel	25 mW/P	
		20 A to 33 A	4.1E-04 rel	1.2 W/P	
		33 A to 100 A	4.1E-04 rel	1.2 W/P	
		100 A to 150 A	4.6E-03 rel	66 mW/P	
		150 A to 550 A	1.3E-03 rel	0.83 W/P	
		550 A to 1000 A	1.2E-03 rel	1.2 W/P	
		1000 A to 2500 A	4.7E-04 rel	30 W/P	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

	33 V to 330 V	330 mA to 1.1 A 1.1 A to 3 A 3 A to 11 A 11 A to 20 A 20 A to 33 A 33 A to 100 A 100 A to 150 A 150 A to 550 A 550 A to 1000 A 1000 A to 2500 A	2.0E-04 rel 3.8E-04 rel 5.0E-04 rel 1.0E-03 rel 4.3E-04 rel 4.3E-04 rel 4.6E-03 rel 1.3E-03 rel 1.2E-03 rel 4.9E-04 rel	13 mW/P 13 mW/P 0.17 W/P 0.25 W/P 12 W/P 12 W/P 0.66 W/P 8.3 W/P 12 W/P 0.31 kW/P	
	330 V to 1000 V	330 mA to 1.1 A 1.1 A to 3 A 3 A to 11 A 11 A to 20 A 20 A to 33 A 33 A to 100 A 100 A to 150 A 150 A to 550 A 550 A to 1000 A 1000 A to 2500 A	2.0E-04 rel 3.8E-04 rel 5.0E-04 rel 1.0E-03 rel 4.3E-04 rel 4.3E-04 rel 4.6E-03 rel 1.3E-03 rel 1.2E-03 rel 4.9E-04 rel	24 mW/P 25 mW/P 0.30 W/P 0.46 W/P 23 W/P 23 W/P 1.2 W/P 15 W/P 23 W/P 0.6 kW/P	

Current Clamps⁵ –

CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Current	Frequency	U1	U2	
AC Current - Generate ³	330 mA to 1.1 A	50 Hz to 60 Hz	8.1E-03 rel	5.8 mA/I	PT-TAR-M-PA-IAC1 PT-TAR-M-PA-IAC2 Direct Method (Note 1s,3s)
	1.1 A to 3 A	50 Hz to 60 Hz	2.9E-03 rel	5.8 mA/I	
	3 A to 11 A	50 Hz to 60 Hz	1.0E-03 rel	6.1 mA/I	
	11 A to 20 A	50 Hz to 60 Hz	1.3E-03 rel	7.1 mA/I	
	20 A to 33 A	50 Hz to 60 Hz	1.1E-03 rel	53 mA/I	
	33 A to 100 A	50 Hz to 60 Hz	1.1E-03 rel	53 mA/I	
	100 A to 150 A	50 Hz to 60 Hz	1.3E-03 rel	59 mA/I	
	150 A to 550 A	50 Hz to 60 Hz	1.2E-03 rel	0.16 A/I	
	550 A to 1000 A	50 Hz to 60 Hz	1.2E-03 rel	0.12 A/I	
	1000 A to 3000 A	50 Hz to 60 Hz	1.1E-03 rel	1.3 A/I	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Clamp-on power meters - Power meters chained with current clamp ⁵						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	PF	U1	U2	
Power Factor @ 50/60 Hz – Generate ³	33 V to 330 V	1 A to 1250 A	0.10 ind	4.0E-03 rel	3.5E-03 unit/PF	PT-TAR-M-PA-PF1 Direct Method (Note 1s, 7s, 20s)
			0.35 ind	3.4E-03 rel	3.3E-03 unit/PF	
			0.50 ind	8.0E-04 rel	3.0E-03 unit/PF	
			0.70 ind	1.1E-03 rel	2.5E-03 unit/PF	
			0.90 ind	4.4E-04 rel	1.5E-03 unit/PF	
			1.00	4.0E-04 rel	5.8E-05 unit/PF	
			0.90 cap	4.4E-06 rel	1.5E-03 unit/PF	
			0.70 cap	5.7E-06 rel	2.5E-03 unit/PF	
			0.50 cap	8.0E-06 rel	3.0E-03 unit/PF	
			0.35 cap	2.3E-05 rel	3.3E-03 unit/PF	
0.10 cap	4.0E-05 rel	3.5E-03 unit/PF				

Clamp-on power meters - Power meters chained with current clamp ⁵							
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)	
	Voltage	Current	cos(φ)	U1	U2		
AC Active Power @ 50/60 Hz – Generate ³	33 V to 330 V	330 mA to 1.1 A	0.1	1.7E-02 rel	3.3 mW/P	PT-TAR-M-PA-PAC1 Direct Method (Note 1s, 8s)	
			0.5	3.1E-03 rel	17 mW/P		
			1.0	5.4E-04 rel	33 mW/P		
			1.1 A to 3 A	0.1	1.7E-02 rel		3.4 mW/P
				0.5	3.1E-03 rel		17 mW/P
				1.0	6.3E-04 rel		34 mW/P
		3 A to 11 A	0.1	1.7E-02 rel	66 mW/P		
			0.5	3.1E-03 rel	0.33 W/P		
			1.0	6.3E-04 rel	0.66 W/P		
		11 A to 20 A	0.1	1.7E-02 rel	0.17 W/P		
			0.5	3.3E-03 rel	0.83 W/P		
			1.0	1.2E-03 rel	1.7 W/P		
		20 A to 33 A	0.1	3.5E-02 rel	1.6 W/P		
			0.5	6.2E-03 rel	7.8 W/P		
			1.0	1.0E-03 rel	16 W/P		
		33 A to 50 A	0.1	3.5E-02 rel	1.8 W/P		
			0.5	6.2E-03 rel	8.8 W/P		
			1.0	9.4E-04 rel	18 W/P		



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

	50 A to 150 A	0.1 0.5 1.0	3.5E-02 rel 6.2E-03 rel 1.3E-03 rel	0.33 W/P 1.7 W/P 3.3 W/P
	150 A to 550 A	0.1 0.5 1.0	3.5E-02 rel 6.2E-03 rel 1.3E-03 rel	3.3 W/P 17 W/P 33 W/P
	550 A to 1000 A	0.1 0.5 1.0	3.5E-02 rel 6.2E-03 rel 1.3E-03 rel	3.3 W/P 17 W/P 33 W/P
	1000 A to 1250 A	0.1 0.5 1.0	3.5E-02 rel 6.2E-03 rel 9.4E-04 rel	44 kW/P 0.22 kW/P 0.44 kW/P

Clamp-on power meters - Power meters chained with current clamp ⁵						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	cos(φ)	U1	U2	
AC Active Power @ 50/60 Hz – Generate ³	330 V to 1000 V	330 mA to 1.1 A	0.1	1.7E-02 rel	6.1 mW/P	PT-TAR-M-PAC1 Direct Method (Note 1s, 8s)
			0.5	3.1E-03 rel	31 mW/P	
			1.0	5.9E-04 rel	61 mW/P	
		1.1 A to 3 A	0.1	1.7E-02 rel	8.8 mW/P	
			0.5	3.1E-03 rel	34 mW/P	
			1.0	6.7E-04 rel	67 mW/P	
		3 A to 11 A	0.1	1.7E-02 rel	0.12 W/P	
			0.5	3.1E-03 rel	0.61 W/P	
			1.0	6.7E-04 rel	1.2 W/P	
11 A to 20 A	0.1	1.7E-02 rel	0.30 W/P			
	0.5	3.3E-03 rel	1.5 W/P			
	1.0	1.2E-03 rel	3.0 W/P			
20 A to 33 A	0.1	3.5E-02 rel	2.8 W/P			
	0.5	6.2E-03 rel	14 W/P			
	1.0	1.1E-03 rel	28 W/P			
33 A to 50 A	0.1	3.5E-02 rel	3.2 W/P			
	0.5	6.2E-03 rel	16 W/P			
	1.0	9.7E-04 rel	32 W/P			
50 A to 150 A	0.1	3.5E-02 rel	0.62 W/P			
	0.5	6.2E-03 rel	3.1 W/P			
	1.0	1.3E-03 rel	6.2 W/P			
150 A to 550 A	0.1	3.5E-02 rel	6.1 W/P			
	0.5	6.2E-03 rel	30 W/P			
	1.0	1.3E-03 rel	61 W/P			
550 A to 1000 A	0.1	3.5E-02 rel	6.1 W/P			
	0.5	6.2E-03 rel	31 W/P			
	1.0	1.3E-03 rel	61 W/P			



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

		1000 A to 1250 A	0.1 0.5 1.0	3.5E-02 rel 6.2E-03 rel 9.7E-04 rel	81 kW/P 0.40 kW/P 0.81 kW/P	
--	--	------------------	-------------------	---	-----------------------------------	--

Clamp-on power meters - Power meters chained with current clamp ⁵						
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)			CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	cos(φ)	U1	U2	
AC Active Energy @ 50/60 Hz – PF @ 1.0 or 0.5– 10min Integration - Generate ³	33 V to 330 V	330 mA to 1.1 A	1.0 0.5	5.4E-04 rel 3.1E-03 rel	5.5 mWh/E 2.8 mWh/E	PT-TAR-M-PA-E1 Direct Method (Note 1s, 11s)
		1.1 A to 3 A	1.0 0.5	6.3E-04 rel 3.1E-03 rel	5.6 mWh/E 2.8 mWh/E	
		3 A to 11 A	1.0 0.5	6.3E-04 rel 3.1E-03 rel	0.11 Wh/E 55 mWh/E	
		11 A to 20 A	1.0 0.5	1.2E-03 rel 3.3E-03 rel	0.28 Wh/E 0.14 Wh/E	
		20 A to 33 A	1.0 0.5	1.0E-03 rel 6.2E-03 rel	2.6 Wh/E 1.3 Wh/E	
		33 A to 100 A	1.0 0.5	9.4E-04 rel 6.2E-03 rel	2.9 Wh/E 1.5 Wh/E	
		100 A to 150 A	1.0 0.5	1.3E-03 rel 6.2E-03 rel	0.55 Wh/E 0.28 Wh/E	
		150 A to 550 A	1.0 0.5	1.3E-03 rel 6.2E-03 rel	5.5 Wh/E 2.8 Wh/E	
		550 A to 1000 A	1.0 0.5	1.3E-03 rel 6.2E-03 rel	5.5 Wh/E 2.8 Wh/E	
	1000 A to 1250 A	1.0 0.5	9.4E-04 rel 6.2E-03 rel	74 Wh/E 37 Wh/E		
	330 V to 1000 V	330 mA to 1.1 A	1.0 0.5	5.9E-04 rel 3.1E-03 rel	10 mWh/E 5.1 mWh/E	
		1.1 A to 3 A	1.0 0.5	6.7E-04 rel 3.1E-03 rel	13 mWh/E 8.0 mWh/E	
		3 A to 11 A	1.0 0.5	6.7E-04 rel 3.1E-03 rel	0.20 Wh/E 0.10 Wh/E	
		11 A to 20 A	1.0 0.5	1.2E-03 rel 3.3E-03 rel	0.50 Wh/E 0.25 Wh/E	
		20 A to 33 A	1.0 0.5	1.1E-03 rel 6.2E-03 rel	4.7 Wh/E 2.4 Wh/E	
		33 A to 100 A	1.0 0.5	9.7E-04 rel 6.2E-03 rel	5.4 Wh/E 2.7 Wh/E	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

		100 A to 150 A	1.0 0.5	1.3E-03 rel 6.2E-03 rel	1.0 Wh/E 0.52 Wh/E	
		150 A to 550 A	1.0 0.5	1.3E-03 rel 6.2E-03 rel	10 Wh/E 5.1 Wh/E	
		550 A to 1000 A	1.0 0.5	1.3E-03 rel 6.2E-03 rel	10 Wh/E 5.1 Wh/E	
Clamp-on power meters - Power meters chained with current clamp⁵						
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)			CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	cos(φ)	U1	U2	
AC Active Energy @ 50/60 Hz – PF @ 1.0 or 0.5– 10min Integration - Generate ³ continued	330 V to 1000 V	1000 A to 1250 A	1.0 0.5	9.7E-04 rel 6.2E-03 rel	0.13 kWh/E 67 Wh/E	PT-TAR-M-PA-E1 Direct Method (Note 1s, 11s)

Clamp-on power meters - Power meters chained with current clamp⁵						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	sin(φ)	U1	U2	
AC Reactive Power @ 50/60 Hz – Generate ³	33 V to 330 V	330 mA to 1.1 A	0.1	5.7E-04 rel	3.4 mvar/Q	PT-TAR-M-PA-Q1 Direct Method (Note 1s, 9s)
			0.5	1.1E-03 rel	17 mvar/Q	
			0.9	3.7E-03 rel	30 mvar/Q	
		1.1 A to 3 A	0.1	6.6E-04 rel	3.4 mvar/Q	
			0.5	1.2E-03 rel	17 mvar/Q	
			0.9	3.7E-03 rel	30 mvar/Q	
		3 A to 11 A	0.1	6.6E-04 rel	66 mvar/Q	
0.5	1.2E-03 rel		0.33 var/Q			
0.9	3.7E-03 rel		0.60 var/Q			
11 A to 20 A	0.1	1.2E-03 rel	0.17 var/Q			
	0.5	1.6E-03 rel	0.83 var/Q			
	0.9	3.8E-03 rel	1.5 var/Q			
20 A to 33 A	0.1	1.1E-03 rel	1.6 var/Q			
	0.5	2.3E-03 rel	7.8 var/Q			
	0.9	7.3E-03 rel	14 var/Q			
33 A to 50 A	0.1	1.0E-03 rel	1.8 var/Q			
	0.5	2.2E-03 rel	8.8 var/Q			
	0.9	7.3E-03 rel	16 var/Q			



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

		50 A to 150 A	0.1 0.5 0.9	1.3E-03 rel 2.4E-03 rel 7.3E-03 rel	0.34 var/Q 1.7 var/Q 3.0 var/Q	
		150 A to 550 A	0.1 0.5 0.9	1.3E-03 rel 2.4E-03 rel 7.3E-03 rel	3.4 var/Q 17 var/Q 30 var/Q	

Clamp-on power meters - Power meters chained with current clamp ⁵						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	sin(φ)	U1	U2	
AC Reactive Power @ 50/60 Hz – Generate ³	33 V to 330 V	550 A to 1000 A	0.1 0.5 0.9	1.3E-03 rel 2.4E-03 rel 7.3E-03 rel	3.4 var/Q 17 var/Q 30 var/Q	PT-TAR-M-PA-Q1 Direct Method (Note 1s, 9s)
		1000 A to 1250 A	0.1 0.5 0.9	1.0E-03 rel 2.2E-03 rel 7.3E-03 rel	44 var/Q 0.22 kvar/Q 0.40 kvar/Q	
	330 V to 1000 V	330 mA to 1.1 A	0.1 0.5 0.9	6.1E-04 rel 1.2E-03 rel 3.7E-03 rel	6.1 mvar/Q 31 mvar/Q 55 mvar/Q	
		1.1 A to 3 A	0.1 0.5 0.9	7.0E-04 rel 1.2E-03 rel 3.7E-03 rel	8.8 mvar/Q 34 mvar/Q 61 mvar/Q	
		3 A to 11 A	0.1 0.5 0.9	7.0E-04 rel 1.2E-03 rel 3.7E-03 rel	0.12 var/Q 0.61 var/Q 1.1 var/Q	
		11 A to 20 A	0.1 0.5 0.9	1.3E-03 rel 1.6E-03 rel 3.8E-03 rel	0.31 var/Q 1.5 var/Q 2.7 var/Q	
		20 A to 33 A	0.1 0.5 0.9	1.1E-03 rel 2.3E-03 rel 7.3E-03 rel	2.8 var/Q 14 var/Q 25 var/Q	
		33 A to 50 A	0.1 0.5 0.9	1.0E-03 rel 2.2E-03 rel 7.3E-03 rel	3.2 var/Q 16 var/Q 29 var/Q	
		50 A to 150 A	0.1 0.5 0.9	1.3E-03 rel 2.4E-03 rel 7.4E-03 rel	0.62 var/Q 3.1 var/Q 5.6 var/Q	
		150 A to 550 A	0.1 0.5 0.9	1.3E-03 rel 2.4E-03 rel 7.3E-03 rel	6.1 var/Q 30 var/Q 55 var/Q	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

		550 A to 1000 A	0.1 0.5 0.9	1.3E-03 rel 2.4E-03 rel 7.3E-03 rel	6.1 var/Q 31 var/Q 55 var/Q	
		1000 A to 1250 A	0.1 0.5 0.9	1.0E-03 rel 2.2E-03 rel 7.3E-03 rel	81 var/Q 0.40 kvar/Q 0.73 kvar/Q	

Clamp-on power meters - Power meters chained with current clamp ⁵					
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	U1	U2	
AC Apparent Power @ 50/60Hz – PF @ 0.1 to 1.0 Generate ³	33 V to 330 V	330 mA to 1.1 A 1.1 A to 3 A 3 A to 11 A 11 A to 20 A 20 A to 33 A 33 A to 100 A 100 A to 150 A 150 A to 550 A 550 A to 1000 A 1000 A to 3000 A	5.4E-04 rel 6.3E-04 rel 6.3E-04 rel 1.2E-03 rel 1.0E-03 rel 9.3E-04 rel 1.3E-03 rel 1.3E-03 rel 1.3E-03 rel 9.3E-04 rel	33 mVA/S 34 mVA/S 0.66 VA/S 1.7 VA/S 14 VA/S 16 VA/S 3.3 VA/S 33 VA/S 33 VA/S 3.9 kVA/S	PT-TAR-M-PA-S1 Direct Method (Note 1s, 10s)
	330 V to 1000 V	330 mA to 1.1 A 1.1 A to 3 A 3 A to 11 A 11 A to 20 A 20 A to 33 A 33 A to 100 A 100 A to 150 A 150 A to 550 A 550 A to 1000 A 1000 A to 3000 A	5.9E-04 rel 6.7E-04 rel 6.7E-04 rel 1.2E-03 rel 1.1E-03 rel 9.6E-04 rel 1.3E-03 rel 1.3E-03 rel 1.3E-03 rel 9.6E-04 rel	61 mVA/S 67 mVA/S 1.2 VA/S 3.0 VA/S 25 VA/S 30 VA/S 6.2 VA/S 61 VA/S 61 VA/S 7.1 kVA/S	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Built-in Meters ⁵				
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
		U1	U2	
DC Voltage - Measure ⁴	0.3 V to 3 V	8.0E-04 rel	10 mV/U	PT-TAR-M-VDC2 Comparison Method (Note 1s, 2s)
	3 V to 6 V	8.0E-04 rel	14 mV/U	
	6 V to 12.5 V	8.1E-04 rel	30 mV/U	
	12.5 V to 25 V	8.1E-04 rel	60 V/U	
	25 V to 60 V	8.0E-04 rel	0.14 V/U	
	60 V to 130 V	8.0E-04 rel	0.31 V/U	
	130 V to 250 V	8.0E-04 rel	0.60 V/U	
	250 V to 400 V	8.0E-04 rel	0.96 V/U	
	400 V to 600 V	8.0E-04 rel	1.5 V/U	
	600 V to 1000 V	8.0E-04 rel	3.5 V/U	

Built-in Meters ⁵				
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
		U1	U2	
DC Current - Measure ⁴	0.15 mA to 1.5 mA	6.1E-03 rel	3.6 µA/I	PT-TAR-M-IDC2 Comparison Method ZES Zimmer LMG95 ZES Zimmer LMG-SH020 ZES Zimmer LMG640/LMG611 (Note 1s, 3s, 16s)
	1.5 mA to 3.0 mA	6.1E-03 rel	7.2 µA/I	
	3 mA to 5 mA	8.0E-04 rel	18 µA/I	
	5 mA to 10 mA	8.0E-04 rel	36 µA/I	
	10 mA to 20 mA	8.0E-04 rel	72 µA/I	
	20 mA to 40 mA	8.0E-04 rel	150 µA/I	
	40 mA to 80 mA	8.0E-04 rel	290 µA/I	
	80 mA to 150 mA	8.0E-04 rel	540 µA/I	
	0.15 A to 0.3 A	8.0E-04 rel	1.1 mA/I	
	0.3 A to 0.6 A	8.0E-04 rel	2.2 mA/I	
	0.6 A to 1.2 A	8.0E-04 rel	4.3 mA/I	
	1.2 A to 2.5 A	8.0E-04 rel	9.0 mA/I	
	2.5 A to 5 A	8.0E-04 rel	18 mA/I	
	5 A to 10 A	8.0E-04 rel	36 mA/I	
	10 A to 20 A	8.0E-04 rel	72 mA/I	
	20 A to 32 A	8.0E-04 rel	120 mA/I	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Built-in Meters ⁵					
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Current	Frequency	U1	U2	
AC Current – Measure ⁴	0.5 mA to 5 mA	50 Hz to 60 Hz	4.0E-04 rel	4.0 µA/I	PT-TAR-M-IAC2 Comparison Method (Note 1s, 3s, 16s)
	5 mA to 10 mA	50 Hz to 60 Hz	4.0E-04 rel	8.0 µA/I	
	10 mA to 20 mA	50 Hz to 60 Hz	4.0E-04 rel	16 µA/I	
	20 mA to 40 mA	50 Hz to 60 Hz	4.0E-04 rel	32 µA/I	
	40 mA to 80 mA	50 Hz to 60 Hz	4.0E-04 rel	64 µA/I	
	80 mA to 150 mA	50 Hz to 60 Hz	4.0E-04 rel	120 µA/I	
	0.15 A to 0.3 A	50 Hz to 60 Hz	4.0E-04 rel	0.24 mA/I	
	0.3 A to 0.6 A	50 Hz to 60 Hz	4.0E-04 rel	0.48 mA/I	
	0.6 A to 1.2 A	50 Hz to 60 Hz	4.0E-04 rel	0.96 mA/I	
	1.2 A to 2.5 A	50 Hz to 60 Hz	4.0E-04 rel	2.0 mA/I	
	2.5 A to 5 A	50 Hz to 60 Hz	4.0E-04 rel	4.0 mA/I	
	5 A to 10 A	50 Hz to 60 Hz	4.0E-04 rel	8.0 mA/I	
	10 A to 20 A	50 Hz to 60 Hz	4.0E-04 rel	16 mA/I	
	20 A to 32 A	50 Hz to 60 Hz	4.0E-04 rel	26 mA/I	
32 A to 60 A	50 Hz to 60 Hz	1.6E-03 rel			
60 A to 100 A	50 Hz to 60 Hz	1.6E-03 rel			

Built-in Meters ⁵					
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Frequency	Voltage	U1	U2	
Frequency (Voltage) – Measure ⁴	1 Hz to 10 Hz	1 V to 10 V	4.0E-04 rel	5.8 µHz/f	PT-TAR-M-F2 Comparison Method (Note 1s, 5s)
	10 Hz to 100 Hz	1 V to 600 V	4.0E-04 rel	5.8 µHz/f	
	100 Hz to 1000 Hz	1 V to 600 V	4.0E-04 rel	5.8 µHz/f	
	1 kHz to 10 kHz	1 V to 600 V	4.0E-04 rel	5.8 mHz/f	
	10 kHz to 100 kHz	1 V to 300 V	4.0E-04 rel	5.8 mHz/f	
	100 kHz to 499.9 kHz	1 V to 3 V	4.0E-04 rel	5.8 mHz/f	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Built-in Meters ⁵						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	PF	U1	U2	
Power Factor @ 50/60 Hz – Measure ⁴	60 V to 1000 V	0.6 A to 50 A	0.10 ind	4.0E-05 rel	2.7E-03 unit/PF	PT-TAR-M-PF1 Comparison Method (Note 1s, 7s, 20s)
			0.35 ind	1.1E-05 rel	3.1E-03 unit/PF	
			0.50 ind	8.0E-06 rel	3.5E-03 unit/PF	
			0.70 ind	5.7E-06 rel	4.1E-03 unit/PF	
			0.90 ind	8.9E-06 rel	4.8E-03 unit/PF	
			1.00	0.0E+00 rel	5.2E-03 unit/PF	
			0.90 cap	8.9E-06 rel	2.7E-03 unit/PF	
			0.70 cap	5.7E-06 rel	2.7E-03 unit/PF	
			0.50 cap	8.0E-06 rel	2.7E-03 unit/PF	
			0.35 cap	2.3E-05 rel	2.7E-03 unit/PF	
			0.10 cap	8.0E-05 rel	2.7E-03 unit/PF	

Built-in Meters ⁵						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	PF	U1	U2	
AC Active Power @ 50/60 Hz – Measure ⁴	25 V to 60 V	0.15 mA to 1.5 mA 1.5 mA to 3.0 mA 3 mA to 5 mA 5 mA to 10 mA 10 mA to 20 mA 20 mA to 40 mA 40 mA to 80 mA 80 mA to 150 mA 0.15 A to 0.3 A 0.3 A to 0.6 A 0.6 A to 1.2 A 1.2 A to 2.5 A 2.5 A to 5 A 5 A to 10 A 10 A to 20 A 20 A to 32 A 32 A to 50 A	0.1 to 1	1.8E-03 rel	0.11 mW/P	PT-TAR-M-PAC1 Comparison Method (Note 1s, 8s, 16s)
			0.1 to 1	1.8E-03 rel	0.22 mW/P	
			0.1 to 1	1.8E-03 rel	0.36 mW/P	
			0.1 to 1	1.8E-03 rel	0.72 mW/P	
			0.1 to 1	1.8E-03 rel	1.5 mW/P	
			0.1 to 1	1.8E-03 rel	2.9 mW/P	
			0.1 to 1	1.8E-03 rel	5.8 mW/P	
			0.1 to 1	1.8E-03 rel	11 mW/P	
			0.1 to 1	1.8E-03 rel	22 mW/P	
			0.1 to 1	1.8E-03 rel	43 mW/P	
			0.1 to 1	1.8E-03 rel	87 mW/P	
			0.1 to 1	1.8E-03 rel	0.18 W/P	
			0.1 to 1	1.8E-03 rel	0.36 W/P	
			0.1 to 1	1.8E-03 rel	0.72 W/P	
			0.1 to 1	1.8E-03 rel	1.5 W/P	
			0.1 to 1	1.8E-03 rel	2.3 W/P	
0.1 to 1	2.4E-03 rel	4.3 W/P				



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

60 V to 130 V	0.15 mA to 1.5 mA	0.1 to 1	1.8E-03 rel	0.24 mW/P
	1.5 mA to 3.0 mA	0.1 to 1	1.8E-03 rel	0.47 mW/P
	3 mA to 5 mA	0.1 to 1	1.8E-03 rel	0.78 mW/P
	5 mA to 10 mA	0.1 to 1	1.8E-03 rel	1.6 mW/P
	10 mA to 20 mA	0.1 to 1	1.8E-03 rel	3.1 mW/P
	20 mA to 40 mA	0.1 to 1	1.8E-03 rel	6.3 mW/P
	40 mA to 80 mA	0.1 to 1	1.8E-03 rel	13 mW/P
	80 mA to 150 mA	0.1 to 1	1.8E-03 rel	24 mW/P
	0.15 A to 0.3 A	0.1 to 1	1.8E-03 rel	47 mW/P
	0.3 A to 0.6 A	0.1 to 1	1.8E-03 rel	94 mW/P
	0.6 A to 1.2 A	0.1 to 1	1.8E-03 rel	0.19 W/P
	1.2 A to 2.5 A	0.1 to 1	1.8E-03 rel	0.39 W/P
	2.5 A to 5 A	0.1 to 1	1.8E-03 rel	0.78 W/P
	5 A to 10 A	0.1 to 1	1.8E-03 rel	1.6 W/P
	10 A to 20 A	0.1 to 1	1.8E-03 rel	3.1 W/P
20 A to 32 A	0.1 to 1	1.8E-03 rel	5.0 W/P	
32 A to 50 A	0.1 to 1	2.4E-03 rel	9.4 W/P	

Built-in Meters ⁵						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	PF	U1	U2	
AC Active Power @ 50/60 Hz – Measure ⁴	130 V to 250 V	0.15 mA to 1.5 mA	0.1 to 1	1.8E-03 rel	0.45 mW/P	PT-TAR-M-PAC1 Comparison Method (Note 1s, 8s, 16s)
		1.5 mA to 3.0 mA	0.1 to 1	1.8E-03 rel	0.90 mW/P	
		3 mA to 5 mA	0.1 to 1	1.8E-03 rel	1.5 mW/P	
		5 mA to 10 mA	0.1 to 1	1.8E-03 rel	3.0 mW/P	
		10 mA to 20 mA	0.1 to 1	1.8E-03 rel	6.0 mW/P	
		20 mA to 40 mA	0.1 to 1	1.8E-03 rel	12 mW/P	
		40 mA to 80 mA	0.1 to 1	1.8E-03 rel	24 mW/P	
		80 mA to 150 mA	0.1 to 1	1.8E-03 rel	45 mW/P	
		0.15 A to 0.3 A	0.1 to 1	1.8E-03 rel	90 mW/P	
		0.3 A to 0.6 A	0.1 to 1	1.8E-03 rel	0.18 W/P	
		0.6 A to 1.2 A	0.1 to 1	1.8E-03 rel	0.36 W/P	
		1.2 A to 2.5 A	0.1 to 1	1.8E-03 rel	0.75 W/P	
		2.5 A to 5 A	0.1 to 1	1.8E-03 rel	1.5 W/P	
		5 A to 10 A	0.1 to 1	1.8E-03 rel	3.0 W/P	
		10 A to 20 A	0.1 to 1	1.8E-03 rel	6.0 W/P	
20 A to 32 A	0.1 to 1	1.8E-03 rel	9.7 W/P			
32 A to 50 A	0.1 to 1	2.4E-03 rel	18 W/P			



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Built-in Meters ⁵						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	PF	U1	U2	
	250 V to 400 V	0.15 mA to 1.5 mA	0.1 to 1	1.8E-03 rel	0.72 mW/P	
		1.5 mA to 3.0 mA	0.1 to 1	1.8E-03 rel	1.4 mW/P	
		3 mA to 5 mA	0.1 to 1	1.8E-03 rel	2.4 mW/P	
		5 mA to 10 mA	0.1 to 1	1.8E-03 rel	4.8 mW/P	
		10 mA to 20 mA	0.1 to 1	1.8E-03 rel	9.7 mW/P	
		20 mA to 40 mA	0.1 to 1	1.8E-03 rel	19 mW/P	
		40 mA to 80 mA	0.1 to 1	1.8E-03 rel	39 mW/P	
		80 mA to 150 mA	0.1 to 1	1.8E-03 rel	72 mW/P	
		0.15 A to 0.3 A	0.1 to 1	1.8E-03 rel	0.14 W/P	
		0.3 A to 0.6 A	0.1 to 1	1.8E-03 rel	0.29 W/P	
		0.6 A to 1.2 A	0.1 to 1	1.8E-03 rel	0.58 W/P	
		1.2 A to 2.5 A	0.1 to 1	1.8E-03 rel	1.2 W/P	
		2.5 A to 5 A	0.1 to 1	1.8E-03 rel	2.4 W/P	
		5 A to 10 A	0.1 to 1	1.8E-03 rel	4.8 W/P	
		10 A to 20 A	0.1 to 1	1.8E-03 rel	9.7 W/P	
		20 A to 32 A	0.1 to 1	1.8E-03 rel	15 W/P	
		32 A to 50 A	0.1 to 1	2.4E-03 rel	29 W/P	

Built-in Meters ⁵						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	PF	U1	U2	
AC Active Power @ 50/60 Hz – Measure ⁴ continued	400 V to 600 V	0.15 mA to 1.5 mA	0.1 to 1	1.8E-03 rel	1.1 mW/P	PT-TAR-M-PAC1 Comparison Method (Note 1s, 8s, 16s)
		1.5 mA to 3.0 mA	0.1 to 1	1.8E-03 rel	2.2 mW/P	
		3 mA to 5 mA	0.1 to 1	1.8E-03 rel	3.6 mW/P	
		5 mA to 10 mA	0.1 to 1	1.8E-03 rel	7.2 mW/P	
		10 mA to 20 mA	0.1 to 1	1.8E-03 rel	14 mW/P	
		20 mA to 40 mA	0.1 to 1	1.8E-03 rel	29 mW/P	
		40 mA to 80 mA	0.1 to 1	1.8E-03 rel	58 mW/P	
		80 mA to 150 mA	0.1 to 1	1.8E-03 rel	0.11 W/P	
		0.15 A to 0.3 A	0.1 to 1	1.8E-03 rel	0.22 W/P	
		0.3 A to 0.6 A	0.1 to 1	1.8E-03 rel	0.43 W/P	
		0.6 A to 1.2 A	0.1 to 1	1.8E-03 rel	0.87 W/P	
		1.2 A to 2.5 A	0.1 to 1	1.8E-03 rel	1.8 W/P	
		2.5 A to 5 A	0.1 to 1	1.8E-03 rel	3.6 W/P	
		5 A to 10 A	0.1 to 1	1.8E-03 rel	7.2 W/P	
		10 A to 20 A	0.1 to 1	1.8E-03 rel	14 W/P	
		20 A to 32 A	0.1 to 1	1.8E-03 rel	23 W/P	
		32 A to 50 A	0.1 to 1	2.4E-03 rel	43 W/P	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Built-in Meters ⁵						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	PF	U1	U2	
AC Active Power @ 50/60 Hz – Measure ⁴	600 V to 1000 V	0.5 mA to 5 mA	0.1 to 1	1.8E-03 rel	6.6 mW/P	PT-TAR-M-PAC1 Comparison Method (Note 1s, 8s, 16s)
		5 mA to 10 mA	0.1 to 1	1.8E-03 rel	13 mW/P	
		10 mA to 20 mA	0.1 to 1	1.8E-03 rel	27 mW/P	
		20 mA to 40 mA	0.1 to 1	1.8E-03 rel	53 mW/P	
		40 mA to 80 mA	0.1 to 1	1.8E-03 rel	0.11 W/P	
		80 mA to 150 mA	0.1 to 1	1.8E-03 rel	0.20 W/P	
		0.15 A to 0.3 A	0.1 to 1	1.8E-03 rel	0.40 W/P	
		0.3 A to 0.6 A	0.1 to 1	1.8E-03 rel	0.80 W/P	
		0.6 A to 1.2 A	0.1 to 1	1.8E-03 rel	1.6 W/P	
		1.2 A to 2.5 A	0.1 to 1	1.8E-03 rel	3.3 W/P	
		2.5 A to 5 A	0.1 to 1	1.8E-03 rel	6.6 W/P	
		5 A to 10 A	0.1 to 1	1.8E-03 rel	13 W/P	
		10 A to 20 A	0.1 to 1	1.8E-03 rel	27 W/P	
		20 A to 32 A	0.1 to 1	1.8E-03 rel	42 W/P	
32 A to 50 A	0.1 to 1	2.4E-03 rel	80 W/P			

Built-in Meters ⁵						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	PF	U1	U2	
AC Active Energy @ 50/60 Hz – Measure ⁴	60 V to 130 V	0.6 A to 1.2 A	0.1 to 1	1.8E-03 rel	0.19 Wh/E	PT-TAR-M-E1 Comparison Method (Note 1s, 11s)
		1.2 A to 2.5 A	0.1 to 1	1.8E-03 rel	0.39 Wh/E	
		2.5 A to 5 A	0.1 to 1	1.8E-03 rel	0.78 Wh/E	
		5 A to 10 A	0.1 to 1	1.8E-03 rel	1.6 Wh/E	
	130 V to 250 V	0.6 A to 1.2 A	0.1 to 1	1.8E-03 rel	0.36 Wh/E	
		1.2 A to 2.5 A	0.1 to 1	1.8E-03 rel	0.75 Wh/E	
		2.5 A to 5 A	0.1 to 1	1.8E-03 rel	1.5 Wh/E	
		5 A to 10 A	0.1 to 1	1.8E-03 rel	3.0 Wh/E	
	250 V to 400 V	0.6 A to 1.2 A	0.1 to 1	1.8E-03 rel	0.58 Wh/E	
		1.2 A to 2.5 A	0.1 to 1	1.8E-03 rel	1.2 Wh/E	
		2.5 A to 5 A	0.1 to 1	1.8E-03 rel	2.4 Wh/E	
		5 A to 10 A	0.1 to 1	1.8E-03 rel	4.8 Wh/E	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Built-in Meters ⁵						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Current	PF	U1	U2	
AC Active Energy @ 50/60 Hz – Measure ⁴ continued	400 V to 600 V	0.6 A to 1.2 A	0.1 to 1	1.8E-03 rel	0.87 Wh/E	PT-TAR-M-E1 Comparison Method (Note 1s, 11s)
		1.2 A to 2.5 A	0.1 to 1	1.8E-03 rel	1.8 Wh/E	
		2.5 A to 5 A	0.1 to 1	1.8E-03 rel	3.6 Wh/E	
		5 A to 10 A	0.1 to 1	1.8E-03 rel	7.2 Wh/E	

Generators ⁵				
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
		U1	U2	
DC Voltage – Measure ⁴	1 mV to 200 mV	1.3E-05 rel	0.24 µV/U	PT-TAR-G-VDC1 Direct Method (Note 1s, 2s)
	200 mV to 2 V	9.0E-06 rel	1.0 µV/U	
	2 V to 20 V	9.0E-06 rel	10 µV/U	
	20 V to 200 V	1.4E-05 rel	0.10 mV/U	
	200 V to 1000 V	1.4E-05 rel	1.2 mV/U	
	1000 V to 10 kV	1.3E-05 rel	0.24 µV/U	
	10 kV to 30 kV	1.5E-02 rel		

Generators ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
		Signal under evaluation U	
DC Ripple – Measure ⁴	1 mV to 200 mV	9.3E-4 %Vmean	PT-TAR-G-VDCrip1 Direct Method
	200 mV to 2 V	6.4E-4 %Vmean	
	2 V to 20 V	6.4E-4 %Vmean	
	20 V to 200 V	1.0E-3 %Vmean	
	200 V to 1000 V	1.0E-3 %Vmean	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Generators ⁵				
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
		U1	U2	
DC Current - Measure ⁴	20 µA to 200 µA	3.2E-05 rel	0.80 nA/I	PT-TAR-G-IDC1 Direct Method (Note 1s, 3s, 16s)
	200 µA to 2 mA	3.2E-05 rel	8.0 nA/I	
	2 mA to 20 mA	3.6E-05 rel	80 nA/I	
	20 mA to 200 mA	1.2E-04 rel	1.6 µA/I	
	200 mA to 2 A	4.5E-04 rel	32 µA/I	
	2 A to 20 A	1.0E-03 rel	0.80 mA/I	

RTD Simulator ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)	TECHNIQUE, REFERENCE STANDARD, EQUIPMENT
		U	
Meters - PT385, 100 Ohm – Generate ³ (Simulated)	-200 °C to -80 °C	13 m°C	PT-TAR-G-TEMP1 Direct Method (Note 17s)
	-80 °C to 100 °C	20 m°C	
	100 °C to 300 °C	24 m°C	
	300 °C to 400 °C	26 m°C	
	400 °C to 630 °C	33 m°C	
Meters - PT385, 1000 Ohm – Generate ³ (Simulated)	-200 °C to 0 °C	15 m°C	PT-TAR-G-TEMP1 Direct Method (Note 17s)
	0 °C to 100 °C	18 m°C	
	100 °C to 300 °C	24 m°C	
	300 °C to 400 °C	26 m°C	
	400 °C to 630 °C	33 m°C	

Thermocouple Simulators ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
		U	
Meters - Type B – Measure ⁴	600 °C to 800 °C	0.35 °C	PT-TAR-G-TEMP1 Direct Method (Note 17s)
	800 °C to 1550 °C	0.28 °C	
	1550 °C to 1820 °C	0.22 °C	
Meters - Type C - Measure ⁴	0 °C to 1000 °C	0.16 °C	
	1000 °C to 1800 °C	0.23 °C	
	1800 °C to 2000 °C	0.26 °C	
	2000 °C to 2316 °C	0.35 °C	
Meters - Type E - Measure ⁴	-250 °C to -200 °C	0.25 °C	
	-200 °C to -100 °C	0.12 °C	
	-100 °C to 0 °C	91 m°C	
	0 °C to 600 °C	81 m°C	
	600 °C to 1000 °C	0.10 °C	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Thermocouple Simulators continued ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
		U	
Meters - Type J - Measure ⁴	-210 °C to -100 °C -100 °C to 800 °C 800 °C to 1200 °C	0.14 °C 91 m°C 0.10 °C	PT-TAR-G-TEMP1 Direct Method (Note 17s)
Meters - Type K - Measure ⁴	-250 °C to -200 °C -200 °C to -100 °C -100 °C to 500 °C 500 °C to 800 °C 800 °C to 1372 °C	0.46 °C 0.16 °C 0.10 °C 0.10 °C 0.13 °C	
Meters - Type L - Measure ⁴	-200 °C to -100 °C -100 °C to 900 °C	0.10 °C 91 m°C	
Meters - Type N - Measure ⁴	-250 °C to -200 °C -200 °C to -100 °C -100 °C to 0 °C 0 °C to 100 °C 100 °C to 800 °C 800 °C to 1300 °C	0.73 °C 0.23 °C 0.12 °C 0.11 °C 0.10 °C 0.12 °C	
Meters - Type R - Measure ⁴	-50 °C to -25 °C -25 °C to 0 °C 0 °C to 100 °C 100 °C to 400 °C 400 °C to 600 °C 600 °C to 1000 °C 1000 °C to 1600 °C 1600 °C to 1767 °C	0.55 °C 0.45 °C 0.39 °C 0.28 °C 0.22 °C 0.21 °C 0.19 °C 0.23 °C	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Thermocouple Simulators continued ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Meters - Type S - Measure ⁴	-50 °C to -25 °C	0.51 °C	PT-TAR-G-TEMP1 Direct Method (Note 17s)
	-25 °C to 0 °C	0.43 °C	
	0 °C to 100 °C	0.38 °C	
	100 °C to 400 °C	0.29 °C	
	400 °C to 600 °C	0.23 °C	
	600 °C to 1000 °C	0.22 °C	
	1000 °C to 1600 °C	0.22 °C	
Meters - Type T - Measure ⁴	1600 °C to 1767 °C	0.26 °C	
	-250 °C to -200 °C	0.35 °C	
	-200 °C to -100 °C	0.16 °C	
	-100 °C to 0 °C	0.11 °C	
	0 °C to 200 °C	91 m°C	
Meters - Type U - Measure ⁴	200 °C to 400 °C	91 m°C	
	-200 °C to 0 °C	0.16 °C	
	0 °C to 200 °C	0.10 °C	
	200 °C to 600 °C	0.10 °C	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Generators ⁵					
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Frequency	U1	U2	
AC Voltage – Measure ⁴	1 mV to 200 mV	10 Hz to 40 Hz	3.2E-04 rel	50 µV /U	PT-TAR-G-VAC1 Direct Method (Note 1s, 2s)
		40 Hz to 100 Hz	2.7E-04 rel	50 µV/U	
		100 Hz to 2 kHz	2.7E-04 rel	24 µV/U	
		2 kHz to 10 kHz	3.3E-04 rel	50 µV/U	
		10 kHz to 30 kHz	7.9E-04 rel	0.10 mV/U	
		30 kHz to 100 kHz	1.7E-03 rel	0.24 mV/U	
	200 mV to 2 V	10 Hz to 40 Hz	2.7E-04 rel	24 µV/U	
	40 Hz to 100 Hz	2.2E-04 rel	24 µV/U		
	100 Hz to 2 kHz	1.8E-04 rel	24 µV/U		
	2 kHz to 10 kHz	2.7E-04 rel	24 µV/U		
	10 kHz to 30 kHz	5.2E-04 rel	50 µV/U		
	30 kHz to 100 kHz	1.3E-03 rel	0.24 mV/U		
	100 kHz to 300 kHz	6.0E-03 rel	2.4 mV/U		
	300 kHz to 1 MHz	2.0E-02 rel	24 mV/U		
	2 V to 20 V	10 Hz to 40 Hz	2.7E-04 rel	24 µV/U	
		40 Hz to 100 Hz	2.2E-04 rel	24 µV/U	
		100 Hz to 2 kHz	1.8E-04 rel	24 µV/U	
		2 kHz to 10 kHz	2.7E-04 rel	24 µV/U	
		10 kHz to 30 kHz	5.2E-04 rel	50 µV/U	
		30 kHz to 100 kHz	1.3E-03 rel	0.24 mV/U	
		100 kHz to 300 kHz	6.0E-03 rel	2.4 mV/U	
		300 kHz to 1 MHz	2.0E-02 rel	24 mV/U	
	20 V to 200 V	10 Hz to 40 Hz	2.7E-04 rel	24 µV/U	
		40 Hz to 100 Hz	2.2E-04 rel	24 µV/U	
		100 Hz to 2 kHz	1.8E-04 rel	24 µV/U	
		2 kHz to 10 kHz	2.7E-04 rel	24 µV/U	
		10 kHz to 30 kHz	5.2E-04 rel	50 µV/U	
		30 kHz to 100 kHz	1.3E-03 rel	0.24 mV/U	
		100 kHz to 300 kHz	6.0E-03 rel	2.4 mV/U	
		300 kHz to 1 MHz	2.0E-02 rel	24 mV/U	
	200 V to 1000 V	10 Hz to 40 Hz	2.9E-04 rel	50 µV/U	
		40 Hz to 10 kHz	2.8E-04 rel	50 µV/U	
		10 kHz to 30 kHz	5.3E-04 rel	0.10 mV/U	
		30 kHz to 100 kHz	1.4E-03 rel	0.10 mV/U	
	1000 V to 10 kV	50 Hz to 60 Hz	1.2E-0.2 rel	3.0 V/U	
	10 kV to 20 kV	50 Hz to 60 Hz	1.5E-0.2 rel		



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Generators ⁵					
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Current	Frequency	U1	U2	
AC Current – Measure ⁴	20 µA to 200 µA	10 Hz to 10 kHz	7.4E-04 rel	4.8 µA/I	PT-TAR-G-IAC1 Direct Method (Note 1s, 3s, 16s)
		10 kHz to 30 kHz	1.6E-03 rel	4.8 µA/I	
		30 kHz to 100 kHz	8.0E-03 rel	4.8 µA/I	
	200 µA to 2mA	10 Hz to 10 kHz	7.4E-04 rel	4.8 µA/I	
		10 kHz to 30 kHz	1.6E-03 rel	4.8 µA/I	
		30 kHz to 100 kHz	8.0E-03 rel	4.8 µA/I	
	2 mA to 20 mA	10 Hz to 10 kHz	7.4E-04 rel	4.8 µA/I	
		10 kHz to 30 kHz	1.6E-03 rel	4.8 µA/I	
		30 kHz to 100 kHz	8.0E-03 rel	4.8 µA/I	
	20 mA to 200 mA	10 Hz to 10 kHz	7.2E-04 rel	48 µA/I	
10 kHz to 30 kHz		1.5E-03 rel	48 µA/I		
200 mA to 2 A	10 Hz to 2 kHz	1.5E-03 rel	0.48 mA/I		
	2 kHz to 10 kHz	1.7E-03 rel	0.48 mA/I		
	10 kHz to 30 kHz	6.0E-03 rel	0.48 mA/I		
2 A to 20 A	10 Hz to 2 kHz	1.8E-03 rel	4.8 mA/I		
	2 kHz to 10 kHz	5.0E-03 rel	4.8 mA/I		
20 A to 30 A	50 Hz to 60 Hz	1.0E-02 rel	24 mA/I		
30 A to 60 A	50 Hz to 60 Hz	1.0E-02 rel	48 mA/I		
60 A to 100 A	50 Hz to 60 Hz	1.0E-02 rel	96 mA/I		
100 A to 1000 A	50 Hz to 60 Hz	5.0E-03 rel			

Generators ⁵					
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Frequency	Voltage	U1	U2	
Frequency (Voltage) – Measure ⁴	10 Hz to 100 Hz	1 mV to 1000 V	2.0E-05 rel	0.40 mHz/f	PT-TAR-G-FV1 Direct Method (Note 1s, 5s)
	100 Hz to 1000 Hz		2.0E-05 rel	4.0 mHz/f	
	1 kHz to 10 kHz		2.0E-05 rel	40 mHz/f	
	10 kHz to 100 kHz		2.0E-05 rel	0.40 Hz/f	
	100 kHz to 1 MHz		2.0E-05 rel	4.0 Hz/f	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Generators ⁵				
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Frequency	U	
AC Voltage Crest Factor – Measure ⁴	6 V to 60 V	50 Hz to 60 Hz	1.7E-02 rel	PT-TAR-G-CF1 Direct Method (Note 1s)
	60 V to 300 V	50 Hz to 60 Hz	1.9E-02 rel	

Generators ⁵				
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Time	Amplitude	U	
Rise Time (Voltage) – Measure ⁴	1 µs to 500 ms	5 mV to 7 kV	24E-05 rel	PT-TAR-G-TRFV1 Direct Method (Note 1s)
Rise Time (Current) – Measure ⁴	1 µs to 500 ms	50 mA to 5 kA	27E-06 rel	PT-TAR-G-TRF11 Direct Method (Note 1s, 16s)
Pulse duration (Voltage) – Measure ⁴	1 µs to 1 s	5 mV to 7 kV	10E-03 rel	PT-TAR-G-TDV1 Direct Method (Note 1s, 6s)

Generators ⁵				
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Time	Amplitude	U	
Pulse duration (Current) – Measure ⁴	1 µs to 1 s	50 mA to 5 kA	10E-03 rel	PT-TAR-G-TDI1 Direct Method (Note 1s)



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Current Generators ⁵				
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Value		U	
Inrush Current - Generate ³	20 mA to 1 A 1.5 A to 6600 A		1.5E-02 rel 3.8E-02 rel	PT-TAR-G-IAC1 Direct Method (Note 1s)

Generators ⁵				
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Voltage	Degrees	U	
Phase Shifting (Voltage – Sinusoidal @ 50/60 Hz) – Measure ⁴	6 V to 300 V	0° to 360°	27E-03°	PT-TAR-G-DSPV1 Direct Method (Note 1s, 6s)

Power Sources ⁵					
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Channel 1	Channel 2	φ	U	
Phase Displacement @ 50/60 Hz – Measure ⁴	10 V to 300 V	3 mA to 100 A	0° to 360°	0.20°	PT-TAR-Q-PD1 PT-TAR-Q-PD2 Direct Method

Voltage Generators ⁵					
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Par	IHRM Amplitude	IHRM Frequency	U	
AC Voltage Harmonics and Interharmonics – Measure ⁴	Fund	30 V to 1000 V	50 Hz/60 Hz	5.0E-03 rel	PT-TAR-G-HRMV1 Direct Method IEC 61000-4-7
	IHRM2 to IHRM2000	0.3 V to 60 V	55 Hz to 12 kHz	1.0E-02 rel	
	THD _v	Fund: 30 V to 1000 V	55 Hz to 12 kHz	7.5E-03 rel	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

					(Note 1s, 2s, 12s)
--	--	--	--	--	--------------------

Calibrators/Resistors ⁵					
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)	
		U1	U2		
DC Resistance – Measure ⁴	10 mΩ to 1 Ω	3.8E-04 rel	15 μΩ/R	PT-TAR-G-RDC2W1 PT-TAR-G-RDC4W1 PT-TAR-Q-RDC1 Direct Method (Note 1s, 4s)	
	1 Ω to 2 Ω	4.4E-05 rel	10 μΩ/R		
	2 Ω to 20 Ω	2.4E-05 rel	36 μΩ/R		
	20 Ω to 200 Ω	2.0E-05 rel	0.12 mΩ/R		
	200 Ω to 2 kΩ	2.0E-05 rel	1.2 mΩ/R		
	2 kΩ to 20 kΩ	2.0E-05 rel	12 mΩ/R		
	20 kΩ to 200 kΩ	2.0E-05 rel	0.12 Ω/R		
	200 kΩ to 2 MΩ	2.4E-05 rel	2.4 Ω/R		
	2 MΩ to 20 MΩ	5.0E-05 rel	0.24 kΩ/R		
	20 MΩ to 200 MΩ	3.0E-04 rel	24 kΩ/R		
200 MΩ to 2 GΩ	3.6E-03 rel	2.4 MΩ/R			



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Impedance module ⁵				
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Resistance	Frequency	U	
Impedance module – Measure ⁴	30 mΩ to 70 mΩ	20 Hz to 9 kHz	2.0E-02 rel	PT-TAR-Q-Z1 Direct Method (Note 1s)
	70 mΩ to 200 mΩ	20 Hz to 9 kHz	1.0E-02 rel	
	200 mΩ to 400 mΩ	20 Hz to 9 kHz	4.0E-03 rel	
		9 kHz to 100 kHz	2.0E-02 rel	
	400 mΩ to 1 Ω	20 Hz to 9 kHz	2.0E-03 rel	
		9 kHz to 100 kHz	2.0E-02 rel	
	1 Ω to 2 Ω	20 Hz to 100 Hz	2.0E-03 rel	
		100 Hz to 9 kHz	1.0E-03 rel	
		9 kHz to 100 kHz	1.0E-02 rel	
	2 Ω to 6 Ω	20 Hz to 9 kHz	1.0E-03 rel	
		9 kHz to 100 kHz	4.0E-03 rel	
		100 kHz to 1 MHz	2.0E-02 rel	
	6 Ω to 10 Ω	20 Hz to 9 kHz	1.0E-03 rel	
		9 kHz to 100 kHz	2.0E-03 rel	
		100 kHz to 1 MHz	1.0E-02 rel	
1 MHz to 3 MHz		2.0E-02 rel		
10 Ω to 30 Ω	20 Hz to 9 kHz	1.0E-03 rel		
	9 kHz to 100 kHz	2.0E-03 rel		
	100 kHz to 1 MHz	4.0E-03 rel		
	1 MHz to 3 MHz	2.0E-02 rel		
30 Ω to 5 kΩ	20 Hz to 100 kHz	1.0E-03 rel		
	100 kHz to 1 MHz	2.0E-03 rel		
	1 MHz to 3 MHz	2.0E-02 rel		
5 kΩ to 10 kΩ	20 Hz to 100 kHz	1.0E-03 rel		
	100 kHz to 1 MHz	2.0E-03 rel		
10 kΩ to 30 kΩ	20 Hz to 100 kHz	1.0E-03 rel		
	100 kHz to 1 MHz	4.0E-03 rel		
30 kΩ to 100 kΩ	20 Hz to 100 kHz	1.0E-03 rel		
	100 kHz to 1 MHz	1.0E-02 rel		
100 kΩ to 200 kΩ	20 Hz to 9 kHz	1.0E-03 rel		
	9 kHz to 100 kHz	2.0E-03 rel		
	100 kHz to 1 MHz	2.0E-02 rel		
200 kΩ to 400 kΩ	20 Hz to 9 kHz	1.0E-03 rel		
	9 kHz to 100 kHz	2.0E-03 rel		
400 kΩ to 1 MΩ	20 Hz to 100 Hz	2.0E-03 rel		
	100 Hz to 9 kHz	1.0E-03 rel		
	9 kHz to 100 kHz	4.0E-03 rel		
1 MΩ to 3 MΩ	20 Hz to 100 Hz	4.0E-03 rel		
	100 Hz to 9 kHz	2.0E-03 rel		



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Impedance module ⁵				
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Resistance	Frequency	U	
Impedance module – Measure ⁴ continued		9 kHz to 100 kHz	1.0E-02 rel	PT-TAR-Q-Z1 Direct Method (Note 1s)
	3 MΩ to 5 MΩ	20 Hz to 100 Hz 100 Hz to 9 kHz 9 kHz to 100 kHz	1.0E-02 rel 2.0E-03 rel 2.0E-02 rel	
	5 MΩ to 9 MΩ	20 Hz to 100 Hz 100 Hz to 9 kHz	1.0E-02 rel 2.0E-03 rel	
	9 MΩ to 20 MΩ	20 Hz to 100 Hz 100 Hz to 10 kHz	2.0E-02 rel 1.0E-02 rel	
	20 MΩ to 50 MΩ	100 Hz to 10 kHz	2.0E-02 rel	

Capacitance ⁵				
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Capacitance	Frequency	U	
Capacitance – Measure ⁴	3 pF to 5 pF	1 kHz to 100 kHz	2.0E-02 rel	PT-TAR-Q-C1 Direct Method (Note 1s)
	5 pF to 10 pF	1 kHz to 100 kHz	1.0E-02 rel	
	10 pF to 30 pF	1 kHz to 100 kHz	4.0E-03 rel	
	30 pF to 100 pF	1 kHz to 100 kHz	2.0E-03 rel	
	100 pF to 10 nF	1 kHz to 100 kHz	1.0E-03 rel	
	10 nF to 100 nF	40 Hz to 100 kHz	1.0E-03 rel	
	100 nF to 10 μF	40 Hz to 10 kHz	1.0E-03 rel	
	10 μF to 1 mF	40 Hz to 1 kHz	4.0E-03 rel	
	1 mF to 10 mF	20 Hz to 100 Hz	4.0E-03 rel	
	1 mF to 10 mF	100 Hz to 400 Hz 400 Hz to 1 kHz	1.0E-02 rel 2.0E-02 rel	
	10 mF to 100 mF	20 Hz to 100 Hz	2.0E-02 rel	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Inductance ⁵					
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Inductance	Frequency	U		
Inductance – Measure ⁴	2 µH to 10 µH	1 kHz to 10 kHz	2.0E-02 rel		PT-TAR-Q-L1 Direct Method (Note 1s)
	10 µH to 30 µH	100 Hz to 1 kHz	2.0E-02 rel		
	10 µH to 30 µH	1 kHz to 10 kHz	1.0E-02 rel		
	30 µH to 100 µH	50 Hz to 1 kHz	2.0E-02 rel		
	30 µH to 100 µH	1 kHz to 10 kHz	1.0E-02 rel		
	100 µH to 1 mH	50 Hz to 1 kHz	1.0E-02 rel		
	100 µH to 1 mH	1 kHz to 10 kHz	1.0E-03 rel		
	1 mH to 10 mH	40 Hz to 1 kHz	2.0E-03 rel		
		1 kHz to 10 kHz	1.0E-03 rel		
10 mH to 1 H	40 Hz to 10 kHz	1.0E-03 rel			

AC Dielectric Strength Meters ⁵					
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Amplitude	Frequency	U1	U2	
AC High Voltage – Measure ⁴	0.1 kV to 1 kV	50 Hz to 60 Hz	2.0E-03 rel	0.55 V/U	PT-TAR-DIEL1 Direct Method (Note 1s, 2s)
	1 kV to 10 kV	50 Hz to 60 Hz	5.0E-03 rel	5.0 V/U	
AC Voltage waveform – Measure ⁴	0.1 kV to 2 kV	50 Hz to 60 Hz	1.7E-02 rel	59 mV/U	PT-TAR-DIEL1 Direct Method (Note 1s, 2s)
AC Voltage THD	@ 1 kV	55 Hz to 2.4 kHz	4.0E-03 rel	58 µ% _{THD} /THD _v	PT-TAR-DIEL1 Direct Method (Note 1, 12)
Test Time – Measure ⁴	0.1 s to 120 s	50 Hz to 60 Hz	3.0E-03 rel	20 ms/T	PT-TAR-DIEL1 Direct Method (Note 1, 6)
	10 s to 999 s	50 Hz to 60 Hz	2.0E-04 rel		
AC Voltage Rise ramp – Measure ⁴	0.1 kV to 2 kV	50 Hz to 60 Hz	1.7E-02 rel	12 mV/U	PT-TAR-DIEL1 Direct Method (Note 1s, 2s)



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

AC Dielectric Strength Meters ⁵ continued					
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
AC Trip current and Leakage current Threshold – Measure ⁴	1 µA to 300 µA	50 Hz to 60 Hz	3.7E-03 rel	0.61 µA/I	PT-TAR-DIEL1 Direct Method (Note 1s, 3s, 16s)
	0.3 mA to 3 mA	50 Hz to 60 Hz	3.3E-03 rel	6.0 µA/I	
	3 mA to 30 mA	50 Hz to 60 Hz	4.5E-03 rel	60 µA/I	
	30 mA to 300 mA	50 Hz to 60 Hz	2.4E-03 rel	0.6 mA/I	
AC Short-Circuit current – Measure ⁴	20 mA to 1 A @ 1 kV to 5 kV	50 Hz to 60 Hz	1.5E-02 rel		PT-TAR-DIEL1 Direct Method (Note 1s)
500 VA Transformer capability – Measure ⁴	100 mA @ 5 kV	50 Hz to 60 Hz	5.7E-03 rel	2.1 V/U	PT-TAR-DIEL1 Direct Method (Note 1s, 2s)

DC Dielectric Strength Meters ⁵					
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
			U1	U2	
DC High Voltage – Measure ⁴	Amplitude				PT-TAR-DIEL1 Direct Method (Note 1s, 2s)
	0.1 kV to 1 kV	1 kV to 10 kV	2.4E-03 rel 3.0E-03 rel	0.55 V/U 5.0 V/U	
DC Voltage ripple – Measure ⁴	0.1 kV to 7 kV		1.7E-02 rel	12 mV/U	PT-TAR-DIEL1 Direct Method (Note 1, 2)



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

DC Dielectric Strength Meters ⁵ continued				
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Amplitude	U1	U2	
Test Time – Measure ⁴	0.1 s to 120 s 10 s to 999 s	3.0E-03 rel 2.0E-04 rel	2.0 ms/T	PT-TAR-DIEL1 Direct Method (Note 1s, 6s)
DC Voltage Rise ramp – Measure ⁴	0.1 kV to 2 kV	1.7E-02 rel	12 mV/U	PT-TAR-DIEL1 Direct Method (Note 1s, 2s)
DC Trip current and Leakage current Threshold – Measure ⁴	1 µA to 300 µA 0.3 mA to 3 mA 3 mA to 30 mA 30 mA to 300 mA	5.1E-03 rel 2.6E-03 rel 4.3E-03 rel 2.3E-03 rel	0.61 µA/I 6.0 µA/I 60 µA/I 0.16 mA/I	PT-TAR-DIEL1 Direct Method (Note 1s, 3s, 16s)
DC Short-Circuit current – Measure ⁴	20 mA to 1 A @ 1 kV to 5 kV	1.5E-02 rel		PT-TAR-DIEL1 Direct Method (Note 1s)



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Ground Bond Resistance Meters ⁵						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Current	Voltage	Resistance	U1	U2	
AC Current @ 50/60 Hz – Measure ⁴	5 A to 10 A 10 A to 20 A 20 A to 30 A 30 A to 60 A 60 A to 100 A			4.0E-04 rel 4.0E-04 rel 1.0E-02 rel 1.0E-02 rel 1.0E-02 rel	8.0 mA/I 16 mA/I 24 mA/I 48 mA/I 96 mA/I	PT-TAR-GBR1 Direct Method (Note 1s, 3s)
Resistance and Threshold @ 50/60 Hz – Measure ⁴	5 A to 10 A		1 mΩ to 20 mΩ 10 mΩ to 200 mΩ 100 mΩ to 2 Ω 1 Ω to 2.4 Ω	7.2E-04 rel 7.2E-04 rel 7.2E-04 rel 7.2E-04 rel	9.0 μΩ/R 86 μΩ/R 0.86 mΩ/R 3.2 mΩ/R	PT-TAR-GBR1 Comparison Method (Note 1s, 4s, 16s)
	10 A to 20 A		0.5 mΩ to 10 mΩ 5 mΩ to 100 mΩ 50 mΩ to 1 Ω 500 mΩ to 1.2 Ω	7.2E-04 rel 7.2E-04 rel 7.2E-04 rel 7.2E-04 rel	4.5 μΩ/R 43 μΩ/R 0.43 mΩ/R 1.6 mΩ/R	
	20 A to 30 A		0.3 mΩ to 5 mΩ 3.3 mΩ to 50 mΩ 33.3 mΩ to 0.5 mΩ 333.3 mΩ to 0.6 Ω	1.0E-02 rel 1.0E-02 rel 1.0E-02 rel 1.0E-02 rel	3.1 μΩ/R 29 μΩ/R 0.29 mΩ/R 1.1 mΩ/R	
	30 A to 60 A		0.2 mΩ to 3.3 mΩ 1.7 mΩ to 33.3 mΩ 16.7 mΩ to 0.33 Ω 166.7 mΩ to 0.4 Ω	1.0E-02 rel 1.0E-02 rel 1.0E-02 rel 1.0E-02 rel	1.6 μΩ/R 14 μΩ/R 0.14 mΩ/R 0.53 mΩ/R	
	60 A to 100A		0.1 mΩ to 1.7 mΩ 1 mΩ to 16.7 mΩ 10 mΩ to 0.17 Ω 100 mΩ to 0.2 Ω	1.0E-02 rel 1.0E-02 rel 1.0E-02 rel 1.0E-02 rel	1.2 μΩ/R 10 μΩ/R 0.10 mΩ/R 0.32 mΩ/R	
AC Voltage and Threshold @ 50/60 Hz – Measure ⁴		10 mV to 0.1 V 0.1 V to 1 V 1 V to 10 V 10 V to 12 V		6.0E-04 rel 6.0E-04 rel 6.0E-04 rel 6.0E-04 rel	40 μV/U 0.30 mV/U 3.0 mV/U 30 mV/U	PT-TAR-GBR1 Comparison Method (Note 1s, 2s, 16s)



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Ground Bond Resistance Meters ⁵ continued						
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Current	Voltage	Resistance	U1	U2	
Test Time @ 50/60 Hz – Measure ⁴	0.1 s to 120 s 10 s to 999 s			3.0E-03 rel 2.0E-04 rel	20 ms/T	PT-TAR-GBR1 Direct Method (Note 1s, 6s)

Insulation Resistance Meters ⁵						
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)	
	Voltage	Resistance	U1	U2		
DC Voltage – Measure ⁴	10 V to 100 V 100 V to 1000 V		1.9E-04 rel 1.6E-04 rel	40 mV/U 40 mV/U	PT-TAR-IR1 Direct Method (Note 1s, 2s)	
Resistance and Threshold – Generate ³	10 V to 500 V	200 kΩ to 999.9 kΩ 1 MΩ to 10 MΩ 10 MΩ to 999.9 MΩ 1 GΩ to 10 GΩ	2.0E-03 rel 3.0E-03 rel 5.0E-03 rel 1.0E-02 rel	58 Ω/R 0.58 kΩ/R 5.8 kΩ/R 0.58 MΩ/R	PT-TAR-IR1 Direct Method (Note 1s, 4s)	
	500 V to 1000 V	200 kΩ to 999.9 kΩ 1 MΩ to 10 MΩ 10 MΩ to 999.9 MΩ 1 GΩ to 10 GΩ	4.0E-03 rel 5.0E-03 rel 7.0E-03 rel 1.2E-02 rel	58 Ω/R 0.58 kΩ/R 5.8 kΩ/R 0.58 MΩ/R		
Test Time – Measure ⁴	0.1 s to 120 s 10 s to 999 s		3.0E-03 rel 2.0E-04 rel	20 ms/T	PT-TAR-IR1 Direct Method (Note 1s, 6s)	

Glow Wire test apparatus ⁵						
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)	
	Amplitude	Frequency	U1	U2		
AC Current – Measure ⁴	10 A to 100 A 100 A to 200 A	50 Hz to 60 Hz 50 Hz to 60 Hz	7.5E-03 rel 7.5E-03 rel	5.8 mA/I 58 mA/I	PT-TAR-GLOW_WIRE1 Comparison Method (Note 1s, 3s)	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

AC Current – Silver foil melting – Measure ⁴	120 A to 150 A	50 Hz to 60 Hz	7.5E-03 rel		PT-TAR-GLOW_WIRE1 Direct Method (Note 1s)
Temperature Simulation (K type) - Generate ³	120 °C to 1000 °C			0.13 °C	PT-TAR-GLOW_WIRE1 Direct Method
Temperature – Silver foil melting	950 °C to 970 °C			1.2 °C	PT-TAR-GLOW_WIRE1 Direct Method
Push force - carriage – Measure ⁴	0.9 N to 1.1 N		4.0E-03 rel		PT-TAR-GLOW_WIRE1 Direct Method (Note 1s)
Dimensional – millimeter scale – Measure ⁴	5 mm to 150 mm			0.29 mm	PT-TAR-GLOW_WIRE1 Direct Method
Ambient light – test chamber – Measure ⁴	2.5 Lux to 4000 Lux		3.1E-02 rel		PT-TAR-GLOW_WIRE1 Direct Method



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Tracking test apparatus ⁵					
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Amplitude	Frequency	U1	U2	
AC Voltage – Measure ⁴	60 V to 130 V 130 V to 250 V 250 V to 400 V 400 V to 600 V	45 Hz to 65 Hz 45 Hz to 65 Hz 45 Hz to 65 Hz 45 Hz to 65 Hz	4.0E-04 rel 4.0E-04 rel 4.0E-04 rel 4.0E-04 rel	0.10 V/U 0.20 V/U 0.32 V/U 0.48 V/U	PT-TAR-TRACKING1 Comparison Method (Note 1s, 2s)
AC Current and Current Threshold – Measure ⁴	15 mA to 0.15 A 0.15 A to 0.3 A 0.3 A to 0.6 A 0.6 A to 1.2 A	45 Hz to 65 Hz 45 Hz to 65 Hz 45 Hz to 65 Hz 45 Hz to 65 Hz	4.1E-04 rel 4.0E-04 rel 4.0E-04 rel 4.0E-04 rel	0.16 mA/I 0.24 mA/I 0.48 mA/I 0.96 mA/I	PT-TAR-TRACKING1 Comparison Method (Note 1s, 3s)
Current Threshold – Time – Measure ⁴	0.1 s to 120 s		3.0E-03 rel		PT-TAR-TRACKING1 Direct Method (Note 1s)
Dimensional - Electrodes – Measure ⁴	1 mm to 150 mm 0.1 mm to 25 mm 0 mm to 0.1 mm			30 µm 3.0 µm 5 µm	PT-TAR-TRACKING1 Direct Method
Angles – Electrodes – Measure ⁴	30° to 60°			2.1E-2°	PT-TAR-TRACKING1 Direct Method
Push force - Electrodes – Measure ⁴	0.9 N to 1.1 N		4.2E-03 rel		PT-TAR-TRACKING1 Direct Method AEP Transducers TCA AEP Transducers DFI (Note 1s)
Drops – Mass – Measure ⁴	0.5 g to 100 g			4.0 Mg	PT-TAR-TRACKING1 Direct Method
Drops – Erogation time – Measure ⁴	1350 s to 1590 s			80 µs	PT-TAR-TRACKING1 Direct Method
Draught proof – test chamber – Measure ⁴	0 m/s to 0.99 m/s			0.10 m/s	PT-TAR-TRACKING1 Direct Method



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Leakage Current Built-in Meters ⁵					
CALIBRATION AREA	RANGE		EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Current	Frequency	U1	U2	
Leakage current – Measure ⁴	10 µA to 100 µA 0.1 mA to 1 mA 1 mA to 10 mA 10 mA to 100 mA	50 Hz to 60 Hz	2.3E-03 rel 1.1E-03 rel 1.0E-03 rel 1.0E-03 rel	0.10 µA/I 0.70 µA/I 4.1 µA/I 40 µA/I	PT-TAR-LEAK1 Direct Method (Note 1s, 3s, 16s)

Leakage Current Networks/Other Networks ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Frequency	U	
Input impedance – Measure ⁴	DC	2.0E-05 rel	PT-TAR-Q-TRANSIMP2 Direct Method (Note 1s)
	20 Hz to 1 MHz	4.0E-03 rel	PT-TAR-Q-TRANSIMP2 Direct Method (Note 1s)
Transfer Impedance – Measure ⁴	DC	1.5E-04 rel	PT-TAR-Q-TRANSIMP2 Direct Method (Note 1s)
	20 Hz to 100 kHz	2.2E-02 rel	PT-TAR-Q-TRANSIMP2 Direct Method (Note 1s)
	100 kHz to 1 MHz	4.2E-02 rel	

Leakage Current Meters not built on power sources ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Frequency	U	
Volt per milliampere indication – Generate ³	DC	1.5E-04 rel	PT-TAR-LEAK1 Direct Method (Note 1s)



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Volt per milliampere indication – Generate ³	20 Hz to 100 kHz 100 kHz to 1 MHz	1.5E-02 rel 2.0E-02 rel	PT-TAR-LEAK1 Direct Method (Note 1s, 21s)
---	--------------------------------------	----------------------------	---

Digital Calipers ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2}	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
		(±) U	
External measurement	1 mm to 1500 mm (r = 0.01 mm)	13 µm	PT-TAR-CALIPER1 Direct Method (Note 18s)
Internal measurement	1 mm to 1500 mm (r = 0.01 mm)	13 µm	
Depth measurement	1 mm to 1500 mm (r = 0.01 mm)	12 µm	

Vernier and Dial Calipers ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2}	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
		(±) U	
External measurement	1 mm to 1500 mm (r = 0.02 mm)	24 µm	PT-TAR-CALIPER1 Direct Method (Note 18s)
Internal measurement	1 mm to 1500 mm (r = 0.02 mm)	24 µm	
Depth measurement	1 mm to 1500 mm (r = 0.02 mm)	24 µm	

Digital micrometers ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2}	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
		(±) U	
External measurement	1 mm to 25 mm (r = 0.001 mm)	1.5 µm	PT-TAR-MICROMETER1 Direct Method
	25 mm to 50 mm (r = 0.001 mm)	1.9 µm	

Analog micrometers ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2}	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
		(±) U	
	1 mm to 25 mm	3.1 µm	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

External measurement	(r = 0.0025 mm)		PT-TAR-MICROMETER1 Direct Method
	25 mm to 50 mm (r = 0.0025 mm)	3.3 µm	

Climatic Chambers ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Value	U	
Temperature - Measure ⁴	-45 °C to 0 °C (r=0.01 °C)	1.0 °C	PT-TAR-CLIMCHAMBER1 PT-TAR-CLIMCHAMBER2 PT-TAR-CLIMCHAMBER3 Comparison Method (Note 19s)
	0 °C to 50 °C (r=0.01 °C)	0.40 °C	
	50 °C to 180 °C (r=0.01 °C)	1.0 °C	
	180 °C to 260 °C (r=0.01 °C)	0.80 °C	
Relative Humidity - Measure ⁴	20 %RH to 95 %RH (r=0.01 %RH) @ 10 °C to 22 °C	2.1 %RH	
	20 %RH to 95 %RH (r=0.01 %RH) @ 22 °C to 24 °C	1.2 %RH	
	20 %RH to 95 %RH (r=0.01 %RH) @ 24 °C to 50 °C	2.1 %RH	

Temperature sensors – chained with indicator ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Value	U	
Temperature - Measure ⁴	-45 °C to 120 °C	0.06 °C	PT-TAR-M-TEMP2 Comparison Method
	120 °C to 140 °C	0.15 °C	
	140 °C to 650 °C	0.40 °C	

Platinum resistance PRT PT100 385 – without indicator - Customer			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Value	U	
Temperature - Measure ⁴	-45 °C to 120 °C	0.06 °C	PT-TAR-M-TEMP3 Comparison Method (Note 22s)
	120 °C to 140 °C	0.15 °C	
	140 °C to 650 °C	0.40 °C	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Thermocouple - without indicator ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
		U	
Temperature - Type B - Measure ⁴	-45 °C to 120 °C	0.36 °C	PT-TAR-M-TEMP3 Comparison Method (Note 22s)
	120 °C to 140 °C	0.38 °C	
	140 °C to 650 °C	0.53 °C	
Temperature - Type C - Measure ⁴	-45 °C to 120 °C	0.17 °C	
	120 °C to 140 °C	0.22 °C	
	140 °C to 650 °C	0.43 °C	
Temperature - Type E - Measure ⁴	-45 °C to 0 °C	0.11 °C	
	0 °C to 120 °C	0.10 °C	
	120 °C to 140 °C	0.17 °C	
	140 °C to 650 °C	0.41 °C	
Temperature - Type J - Measure ⁴	-45 °C to 120 °C	0.11 °C	
	120 °C to 140 °C	0.18 °C	
	140 °C to 650 °C	0.41 °C	
Temperature - Type K - Measure ⁴	-45 °C to 120 °C	0.12 °C	
	120 °C to 140 °C	0.18 °C	
	140 °C to 650 °C	0.41 °C	
Temperature - Type N - Measure ⁴	-45 °C to 120 °C	0.13 °C	
	120 °C to 140 °C	0.18 °C	
	140 °C to 650 °C	0.41 °C	
Temperature - Type R - Measure ⁴	-45 °C to 0 °C	0.46 °C	
	0 °C to 120 °C	0.40 °C	
	120 °C to 140 °C	0.32 °C	
	140 °C to 650 °C	0.45 °C	
Temperature - Type S - Measure ⁴	-45 °C to 0 °C	0.44 °C	
	0 °C to 120 °C	0.39 °C	
	120 °C to 140 °C	0.33 °C	
	140 °C to 650 °C	0.46 °C	
Temperature - Type T - Measure ⁴	-45 °C to 0 °C	0.13 °C	
	0 °C to 120 °C	0.11 °C	
	120 °C to 140 °C	0.18 °C	
	140 °C to 400 °C	0.41 °C	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Temperature source – temperature evaluation of one/multiple points inside a conditioned environment ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Value	U	
Temperature - Measure ⁴	-45 °C to 0 °C	1.0 °C	PT-TAR-M-TEMP4 Direct Method
	0 °C to 50 °C	0.40 °C	
	50 °C to 180 °C	1.0 °C	
	180 °C to 260 °C	0.80 °C	

Climatic central units – T, RH and Barometric Pressure ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Value	U	
Temperature - Measure ⁴	10 °C to 60 °C	0.20 °C	PT-TAR-M-CCU1 Comparison Method
Relative Humidity - Measure ⁴	10 %RH to 95 %RH	1.0 %RH	
Barometric Pressure - Measure ⁴	800 mbar to 1100 mbar	0.60 mbar	PT-TAR-M-BARP1 Comparison Method (Note 16s)

Non-automatic weighing instruments ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Value	U	
Weight	10 mg to 1 g	13 µg	PT-TAR-M-SCALE1 Direct Method
	1 g to 10 g	29 µg	
	10 g to 100 g	0.19 mg	
	100 g to 1 kg	0.97 mg	
	1 kg to 10 kg	9.0 mg	
	10 kg to 100 kg	90 mg	
	100 kg to 300 kg	0.14 g	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Manometer and pressure gauge ⁵				
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Value	U1	U2	
Pressure - Generate ³	1 mbar to 100 mbar		0.05 mbar	PT-TAR-M-PRE1 PT-TAR-M-PRE2 Comparison Method (Note 1s)
	-0.8 bar to 1 bar	5.0E-04	0.40 mbar	
	1 bar to 2 bar	1.5E-03 rel	0.70 mbar	
	2 bar to 5 bar		5.0 mbar	
	5 bar to 20 bar		7.0 mbar	
	20 bar to 50 bar		30 mbar	
	50 bar to 350 bar		0.18 bar	
	350 bar to 700 bar		0.50 bar	

Absolute Manometer ⁵				
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Value	U1	U2	
Absolute Pressure - Generate ³	0.2 bar to 20 bar		0.013 bar	PT-TAR-M-PRE3 Comparison Method (Note 1s, 26s)
	21 bar to 51 bar		30 mbar	
	51 bar to 351 bar		0.18 bar	
	351 bar to 701 bar		0.50 bar	

Torque meter/Slip Torque Meter				
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Value	U1	U2	
Torque meter/Slip Torque Meter - Measure ⁴	0.1 N·m to 2.5 N·m	1.0E-02 rel	5 mN·m /Tq	PT-TAR-M-TRQ1 Comparison Method (Note 1s, 14s)
	2.5 N·m to 25 N·m	5.0E-03 rel	13 mN·m /Tq	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Torque meter ⁵					
CALIBRATION AREA	RANGE			EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Weight	Lever	Torque	U	
Torque meter - Measure ⁴	0 g to 100 g	0 cm to 50 cm	0 N·m to 0.49 N·m	0.58 mN·m	PT-TAR-M-TRQ2 Direct Method
		50 cm to 100 cm	0 N·m to 0.98 N·m	0.59 mN·m	
		100 cm to 200 cm	0.1 N·m to 1.96 N·m	0.61 mN·m	
	100 g to 1 kg	0 cm to 20 cm	0 N·m to 1.96 N·m	0.97 mN·m	
		20 cm to 100 cm	0.2 N·m to 9.82 N·m	0.98 mN·m	
1 kg to 10 kg	100 cm to 200 cm	0.98 N·m to 19.63 N·m	0.99 mN·m		
	0 cm to 100 cm	0 N·m to 98.17 N·m	7.9 mN·m		
10 kg to 20 kg	100 cm to 200 cm	9.82 N·m to 196.34 N·m	8.1 mN·m		
	0 cm to 200 cm	0 N·m to 392.68 N·m	79 mN·m		
20 kg to 115 kg	0 cm to 200 cm	0 N·m to 2257.9 N·m	0.16 N·m		

Linear Dimension ⁵				
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Value	U1	U2	
Linear dimension - Measure ⁴ in contact	0.1 mm to 25 mm			PT-TAR-Q-LENGTH1 Direct Method
	25 mm to 150 mm			
Linear dimension - Measure ⁴ in contact/Optical	150 mm to 200 mm	50E-6 rel	6.0 µm	
	200 mm to 600 mm			
	600 mm to 2000 mm			
	2000 mm to 3000 mm			
	217 mm x 217 mm x 160 mm			
Degrees	0° to 90°		0.2°	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Digital Inclinometer				
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Value	U		
Degrees - Measure ⁴	0° to 90°	0.20°		PT-TAR-M-ANGLE1 Comparison Method

Tape Measure				
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Value	U		
Tape Measure - Tick distance	1 mm to 2000 mm	0.40 mm		PT-TAR-M-TAPE1 Comparison Method
	2000 mm to 3000 mm	0.50 mm		

Dynamometer ⁵				
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Value	U		
Dynamometer - Measure ⁴	0 N to 10 N	0.18 mN		PT-TAR-M-FRC1 Direct Method
	10 N to 100 N	2.6 mN		
	100 N to 1000 N	25 mN		

Impact Hammer				
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Value	U1	U2	
Energy	0.2 J to 1 J		0.013 J	PT-TAR-IMPHAM1 Direct Method (Note 1s, 15s)
Linear Dimension	0 mm to 230 mm		5.0 µm	
Releasing Force	≤10 N	2.0E-3 rel	5.0 mN/F	
Striking element Mass	250 g		1.0 g	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Weight			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Value	U	
Weight – Measure	0.5 g to 220 g	20 mg	PT-TAR-Q-MASS1 Direct Method
	220 g to 5700 g	2.0 g	
	5700 g to 30000 g	14 g	

Stopwatch - Manual Start/Stop ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Value	U	
Stopwatch - Measure ⁴	10 s to 600 s	5.0 ms	PT-TAR-M-TIME1 Comparison Method
	600 s to 10000 s	50 ms	
	10000 s to 30000 s	52 ms	

Stopwatch - Electrical Start/Stop ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Value	U	
Stopwatch - Measure ⁴	10 s to 600 s	5.0 ms	PT-TAR-M-TIME1 Comparison Method
	600 s to 10000 s	21 ms	
	10000 s to 30000 s	25 ms	

Angular Velocity ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Value	U	
Angular Velocity - Measure ⁴	6 rpm to 30000 rpm	0.21 rpm	PT-TAR-M-RPM2 Comparison Method

Angular Velocity – Manual count ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Value	U	
Angular Velocity - Measure ⁴	1 rpm to 10 rpm 10 rpm to 100 rpm	3.3E-03 rpm 33E-03 rpm	PT-TAR-Q-RPM1 Comparison Method



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Tachometer ⁵				
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Value	U		
Angular Velocity - Generate ³	6 rpm to 30000 rpm	95E-03 rpm		PT-TAR-M-RPM1 Direct Method

Chamber - Air velocity ⁵				
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Value	U		
Air Velocity measure ⁴	0.0 m/s to 0.55 m/s 0.55 m/s to 2.00 m/s 2.00 m/s to 5.00 m/s	30 mm/s 0.10 m/s 0.25 m/s		PT-TAR-Q-AIRVEL1 Direct Method

Chamber - Ambient light ⁵				
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Value	U1	U2	
Ambient light measure ⁴	2.5 lx to 40 lx 40 lx to 400 lx 400 lx to 4000 lx	3.0E-02 rel 3.0E-02 rel 3.0E-02 rel	0.20 lx 2.0 lx 20 lx	PT-TAR-Q-LUX1 Direct Method

Flow Meters ⁵				
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Value	U1	U2	
Water Flow - Measure ⁴	0 L/min to 10 L/min 10 L/min to 20 L/min 20 L/min to 38 L/min 38 L/min to 380 L/min	5.0E-03 rel 5.0E-03 rel 5.0E-03 rel 2.0E-02 rel	0.75 mL/min 1.4 mL/min 0.49 mL/min 10 mL/min	PT-TAR-M-FLW1 Comparison Method



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Water Counter ⁵				
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)		CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	Value	U1	U2	
Water Count - Measure ⁴	0.2 L/min to 2 L/min 2 L/min to 10 L/min	5.0E-03 rel 5.0E-03 rel	0.75 mL 1.4 mL	PT-TAR-M-WCNT1 Comparison Method

Fast Transient/Burst – Generator ⁵ -			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
		U	
Peak Voltage into 50 Ω - Measure ⁴	0.25 kV to 8 kV	43E-03 rel	PT-TAR-G-BURST IEC 61000-4-4 (Note 1s)
Peak Voltage into 1000 Ω - Measure ⁴	0.25 kV to 8 kV	43E-03 rel	
Rise Time into 50 Ω - Measure ⁴	5 ns	20 ps	
Rise Time into 1000 Ω - Measure ⁴	5 ns	35 ps	
Pulse Width into 50 Ω - Measure ⁴	50 ns	18 ps	
Pulse Width into 1000 Ω - Measure ⁴	50 ns to 150 ns	4.0 ps	
Frequency - Measure ⁴	5 kHz and 100 kHz	10E-09 rel	
Duration @ 5 kHz - Measure ⁴	15 ms	18 ns	
Duration @ 100 kHz - Measure ⁴	0.75 ms	18 ns	
Period - Measure ⁴	300 ms	18 ns	

Fast Transient/Burst - Coupling Decoupling Network ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
		U	
Peak Voltage into 50 Ω - Measure ⁴	0.25 kV to 8 kV	43E-03 rel	PT-TAR-G-BURSTIEC 61000-4-4 (Note 1s)
Rise Time into 50 Ω - Measure ⁴	5.5 ns	20 ps	
Pulse Width into 50 Ω - Measure ⁴	45 ns	18 ps	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Fast Transient/Burst - Capacitive Clamp ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2}	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
		(±)	
		U	
Peak Voltage into 50 Ω - Measure ⁴	2 kV	43E-03 rel	PT-TAR-G-BURST IEC 61000-4-4 (Note 1s)
Rise Time into 50 Ω - Measure ⁴	5 ns	20 ps	
Pulse Width into 50 Ω - Measure ⁴	50 ns	18 ps	

Surge Generator ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2}	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
		(±)	
		U	
Open Circuit Peak Voltage - Measure ⁴	0.25 kV to 7 kV	12E-03 rel	PT-TAR-G-SURGE IEC 61000-4-5 (Note 1s)
Front Time Voltage - Measure	1.2 μs	24E-05 rel	
Time to half value Voltage - Measure ⁴	50 μs	35E-06 rel	
Short Circuit Peak Current - Measure ⁴	20 A to 3.5 kA	38E-03 rel	
Front Time Current - Measure ⁴	8 μs	27E-06 rel	
Time to half value Current - Measure ⁴	20 μs	10E-06 rel	
Output Impedance	2 Ω to 12 Ω	5E-02 rel	

Surge Generator - Coupling Decoupling Network ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2}	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
		(±)	
		U	
Open Circuit Peak Voltage - Measure ⁴	0.25 kV to 7 kV	12E-03 rel	PT-TAR-G-SURGE IEC 61000-4-5 (Note 1s)
Front Time Voltage - Measure ⁴	1.2 μs	24E-05 rel	
Time to half value Voltage - Measure ⁴	50 μs	35E-06 rel	
Short Circuit Peak Current - Measure ⁴	8 A to 3.5 kA	38E-03 rel	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Front Time Current - Measure ⁴	2.5 μ s to 8 μ s	27E-06 rel	
Time to half value Current - Measure ⁴	20 μ s to 25 μ s	10E-06 rel	

Ring Wave Generator ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2}	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
		U	
Open Circuit Peak Voltage - Measure ⁴	0.25 kV to 7 kV	12E-03 rel	PT-TAR-G-RINGWAVE IEC 61000-4-12 (Note 1s)
Rise Time Voltage - Measure ⁴	0.500 μ s	24E-05 rel	
Frequency - Measure ⁴	100 kHz	16E-06 rel	
Decaying - Measure ⁴	Ratio from Peak voltage	0.02 Unit	
Short Circuit Peak Current - Measure ⁴	up to 500 A	38E-03 rel	
Rise Time Current - Measure ⁴	1 μ s	27E-06 rel	
Output Impedance	12 Ω to 30 Ω	5E-02 rel	

Flicker ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2}	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
		U	
P_{st} - Measure ⁴	$P_{st}=1$	30E-03 Unit	PT-TAR-M-FLICKER IEC 61000-3-3, IEC 61000-3-11, IEC 61000-4-15
$\Delta U/U$ - Measure ⁴	0.4 % to 3.2 %	150E-03 % $\Delta U/U$	
Flicker Impedance Resistive Part - Measure ⁴	0.032 Ω to 0.400 Ω	0.9E-03 Ω	
Flicker Impedance Imaginary Part - Measure ⁴	0.020 j to 0.250 j	0.5E-03 j	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

ElectroStatic Discharge			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
		U	
Air discharge DC output voltage- Measure ⁴	2 kV to 30 kV	15E-03 rel	PT-TAR-G-ESD IEC 61000-4-2, ISO 10605 (Note 1s)
Peak Current - Measure ⁴ Discharge Circuit	7.5 A to 112.50 A	58E-03 rel	
Current @ 30 ns - Measure ⁴ Discharge Circuit	4 A to 60 A	58E-03 rel	
Current @ 60 ns - Measure ⁴ Discharge Circuit	2 A to 30 A	58E-03 rel	
Rise Time - Measure ⁴ Discharge Circuit	600 ps to 1000 ps	80 ps	

Power frequency magnetic field immunity – Generator/Antenna ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)	TECHNIQUE, REFERENCE STANDARD, EQUIPMENT
		U	
Magnetic Field	1 A/m to 100 A/m @ 50-60 Hz	1.0E-01 rel	PT-TAR-Q- MAGNETIC_FIELD IEC 61000-4-8
Magnetic Field Uniformity	1 A/m to 100 A/m @ 50-60 Hz	1.2E-01 rel	
THD	100A total rms Fund @ 50-60 Hz	2.5E-02 rel	
Coil Factor	1 A/m to 100 A/m @ 50-60 Hz	1.0E-01 rel	

S-Parameters ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)	TECHNIQUE, REFERENCE STANDARD, EQUIPMENT
		U	
Transmission Measurements - Measure ⁴	9 kHz to 50 MHz -50 dB to 0 dB -360° to 360°	0.20 dB 2.8°	PT-TAR-Q- S_PARAMETERS Direct Method
	50 MHz to 3 GHz -70 dB to -50 dB -360° to 360°	0.30 dB 3.6°	
	50 MHz to 3 GHz -50 dB to 0 dB -360° to 360°	0.20 dB 2.8°	
	3 GHz to 6 GHz -70 dB to -50 dB	0.31 dB	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

	-360° to 360°	3.6°	
	3 GHz to 6 GHz -50 dB to 0 dB -360° to 360°	0.21 dB 2.8°	

S-Parameters ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)	TECHNIQUE, REFERENCE STANDARD, EQUIPMENT
		U	
Transmission Measurements - Measure ⁴	100 kHz to 200 MHz -60 dB to -50 dB -360° to 360°	0.20 dB 2.1°	PT-TAR-Q-S_PARAMETERS Direct Method
	100 kHz to 200 MHz -50 dB to -40 dB -360° to 360°	0.11 dB 1.8°	
	100 kHz to 200 MHz -40 dB to -20 dB -360° to 360°	0.089 dB 1.7°	
	100 kHz to 200 MHz -20 dB to 0 dB -360° to 360°	0.083 dB 1.6°	
	200 MHz to 4 GHz -60 dB to -50 dB -360° to 360°	0.24 dB 2.3°	
	200 MHz to 4 GHz -50 dB to -40 dB -360° to 360°	0.12 dB 1.8°	
	200 MHz to 4 GHz -40 dB to -20 dB -360° to 360°	0.094 dB 1.7°	
	200 MHz to 4 GHz -20 dB to 0 dB -360° to 360°	0.088 dB 1.6°	
	4 GHz to 10 GHz -60 dB to -50 dB -360° to 360°	0.25 dB 3.9°	
	4 GHz to 10 GHz -50 dB to -40 dB -360° to 360°	0.13 dB 3.7°	
	4 GHz to 10 GHz -40 dB to -20 dB -360° to 360°	0.10 dB 3.6°	
	4 GHz to 10 GHz -20 dB to 0 dB -360° to 360°	0.096 dB 3.6°	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

S-Parameter⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY^{1,2} (±)	TECHNIQUE, REFERENCE STANDARD, EQUIPMENT
Transmission Measurements - Measure ⁴	10 GHz to 20 GHz -60 dB to -50 dB -360° to 360°	0.26 dB 4.9°	PT-TAR-Q-S_PARAMETERS Direct Method
	10 GHz to 20 GHz -50 dB to -40 dB -360° to 360°	0.14 dB 4.7°	
	10 GHz to 20 GHz -40 dB to -20 dB -360° to 360°	0.12 dB 4.6°	
	10 GHz to 20 GHz -20 dB to 0 dB -360° to 360°	0.11 dB 4.6°	
	20 GHz to 26.5 GHz -60 dB to -50 dB -360° to 360°	0.34 dB 5.2°	
	20 GHz to 26.5 GHz -50 dB to -40 dB -360° to 360°	0.17 dB 4.8°	
	20 GHz to 26.5 GHz -40 dB to -20 dB -360° to 360°	0.13 dB 4.7°	
	20 GHz to 26.5 GHz -20 dB to 0 dB -360° to 360°	0.12 dB 4.7°	
	26.5 GHz to 35 GHz -60 dB to -50 dB -360° to 360°	0.34 dB 5.6°	
	26.5 GHz to 35 GHz -50 dB to -40 dB -360° to 360°	0.17 dB 5.2°	
	26.5 GHz to 35 GHz -40 dB to -20 dB -360° to 360°	0.14 dB 5.2°	
	26.5 GHz to 35 GHz -20 dB to 0 dB -360° to 360°	0.13 dB 5.1°	
	35 GHz to 40 GHz -60 dB to -50 dB -360° to 360°	0.99 dB 8.3°	
	35 GHz to 40 GHz -50 dB to -40 dB -360° to 360°	0.34 dB 5.6°	
	35 GHz to 40 GHz -40 dB to -20 dB -360° to 360°	0.17 dB 5.2°	
	35 GHz to 40 GHz -20 dB to 0 dB -360° to 360°	0.14 dB 5.2°	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

S-Parameters⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY^{1,2} (±)	TECHNIQUE, REFERENCE STANDARD, EQUIPMENT
Reflection Measurements - Measure ⁴	9 kHz to 6 GHz 0 dB to -15 dB -360° to 360°	0.41 dB 3.6°	PT-TAR-Q-S_PARAMETERS Direct Method
	9 kHz to 6 GHz -15 dB to -25 dB -360° to 360°	1.0 dB 6.4°	
	9 kHz to 6 GHz -25 dB to -35 dB -360° to 360°	3.0 dB 20°	
	100 kHz to 10 GHz 0 dB to -15 dB -360° to 360°	0.13 dB 3.6°	
	100 kHz to 10 GHz -15 dB to -25 dB -360° to 360°	0.30 dB 3.9°	
	100 kHz to 10 GHz -25 dB to -35 dB -360° to 360°	0.89 dB 7.5°	
	10 GHz to 20 GHz 0 dB to -15 dB -360° to 360°	0.20 dB 4.7°	
	10 GHz to 20 GHz -15 dB to -25 dB -360° to 360°	0.49 dB 5.3°	
	10 GHz to 20 GHz -25 dB to -35 dB -360° to 360°	1.3 dB 12°	
	20 GHz to 40 GHz 0 dB to -15 dB -360° to 360°	0.27 dB 5.3°	
	20 GHz to 40 GHz -15 dB to -25 dB -360° to 360°	0.62 dB 6.3°	
	20 GHz to 40 GHz -25 dB to -35 dB -360° to 360°	1.8 dB 15°	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

Normalized Site Attenuation and Site Voltage Standing Wave Ratio ⁵			
CALIBRATION AREA	RANGE	EXPANDED UNCERTAINTY ^{1,2} (±)	TECHNIQUE, REFERENCE STANDARD, EQUIPMENT
		U	
Normalized Site Attenuation Horizontal	30 MHz to 200 MHz -10 dB to 10 dB	0.74 dB	PT-TAR-Q-NSA PT-TAR-Q-SVSWR Direct Method CISPR 16-1-4
Normalized Site Attenuation Vertical	30 MHz to 200 MHz -10 dB to 10 dB	0.90 dB	
Normalized Site Attenuation Horizontal	200 MHz to 1 GHz -10 dB to 10 dB	0.70 dB	
Normalized Site Attenuation Vertical	200 MHz to 1 GHz -10 dB to 10 dB	0.70 dB	
Site Voltage Standing Wave Ratio Horizontal and Vertical	1 GHz to 6 GHz 0 dB to 10 dB	1.9 dB	
Site Voltage Standing Wave Ratio Horizontal and Vertical	6 GHz to 18 GHz 0 dB to 10 dB	1.9 dB	

General Notes:

¹The uncertainty covered by the Calibration and Measurement Capability (CMC) is expressed as the expanded uncertainty having a coverage probability of approximately 95 %. It is the smallest measurement uncertainty that a laboratory can achieve within its scope of accreditation when performing calibrations of a best existing device. The measurement uncertainty reported on a calibration certificate may be greater than that provided in the CMC due to the behavior of the calibration item and other factors that may contribute to the uncertainty of a specific calibration.

²When uncertainty is stated in relative terms (such as percent, a multiplier expressed as a decimal fraction or in scientific notation), it is in relation to instrument reading or instrument output, as appropriate, unless otherwise indicated. When Calibration and Measurement Capability (CMC) is expressed in two parts as **U1** and **U2**, the total CMC is the root sum squared of U1 and U2. Given below is an example calculation of total CMC expressed in % of reading for DC Voltage at a calibration point of 50 mV:

$$CMC_{[\%rdg]} = 100 \cdot \sqrt{U1^2_{[rel]} + \left(\frac{U2_{[V]}}{Calpoint_{[V]}}\right)^2}$$

$$CMC_{[\%rdg]} = 100 \cdot \sqrt{(1.5 \cdot 10^{-3})^2_{[rel]} + \left(\frac{1.4 \cdot 10^{-6}[V]}{50 \cdot 10^{-3}[V]}\right)^2} = 0.15_{[\%rdg]}$$

³Capability is suitable for the calibration of measuring devices in the stated ranges.

⁴Capability is suitable for the calibration of devices intended to generate the indicated quantity in the stated ranges.

⁵Also available as site calibration. Note that actual measurement uncertainties achievable at a customer's site can normally be expected to be larger than the uncertainties listed on this Scope of Accreditation.

Specific Notes:

- 1s rel is value declared in relative form.
- 2s U indicates voltage expressed in Volt.
- 3s I indicates current expressed in Ampere.
- 4s R indicates resistance expressed in Ohm.



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

- 5s *f* indicates frequency expressed in Hertz.
- 6s *T* indicates time expressed in second.
- 7s *PF* indicates power factor expressed in unit.
- 8s *P* indicates active power expressed in Watt.
- 9s *Q* indicates reactive power expressed in var.
- 10s *S* indicates apparent power expressed in VA.
- 11s *E* indicates active energy expressed in Wh.
- 12s *THD* indicates THD expressed in %_{THD}.
- 13s *L* indicates Length expressed in meter.
- 14s *Tq* indicates Torque expressed in Nm
- 15s *F* Indicates Force expressed in N
- 16s Ranges are defined considering the measurement capability of reference standards. If the quantity is obtained using auxiliary equipment (e.g. resistive loads) there could be limitations on measuring points definition.
- 17s Uncertainties do not take into account temperature sensor uncertainty contributors.
- 18s The coverage factor k is 1.83 (trapezoidal distribution).
- 19s Uncertainty includes contributors of fluctuation, uniformity, repeatability and resolution.
- 20s The waveform applied is perfectly sinusoidal $P = S \cdot \cos(\varphi)$ and $PF = \cos(\varphi)$ so $PF = P/S$ where P is Real Power and S is Apparent Power.
- 21s Uncertainty does not include resolution
- 22s Measurements are performed considering nominal constants to transduce temperature
- 23s Uncertainties are evaluated for each measurement
- 24s Power factor values are intended for both inductive and capacitive measurement
- 25s *C* indicates Capacitance in F
- 26s *P_amb* indicates Ambient pressure in bar
- 27s Stated uncertainties are valid for the ranges of frequencies given, but the actual frequency applied by the calibrator may be dependent on the measurement device under calibration.



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

CL-171

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Dimensional			
Calibration of steel balls	0 mm to 60 mm	62 µm	Cal procedure – ULID-003973 Using Digital Caliper by direct method
Mechanical			
Measurement of weights	500 g to 4000 g	2.9 g	Cal procedure – ULID-003974 Using Weights and Measuring Scale by comparison method
Tension meter calibration (for Surface Tension)	48 mN/m to 250 mN/m	0.18 mN/m	Cal procedure – ULID-008346 Using Reference weights & Thermometer by comparison method
Water flow meters	0.5 L/min to 5 L/min 5 L/min to 22 L/min	0.05 L/min 0.36 L/min	Cal procedure – ULID-003972 Using master water flow meter by comparison methods
Sound Level Meters	94 dB 114 dB	0.3 dB 0.6 dB	Cal Procedure – ULID-003968 Using Sound Level Calibrator by direct method (Fixed point)
Thermal			
Thermocouple	-40 °C to 140 °C 140 °C to 300 °C 300 °C to 700 °C	0.4 °C 1.8 °C 2.6 °C	Cal Procedure – ULID-005825 Using Dry Block Calibrator & Pt-100 / Type N reference sensor
Electrical – DC/LF			
Thermocouple Simulation Generate and Measure			Cal Procedure – ULID-003960 Using Multifunction Calibrator by direct method
Type J	-210 °C to -100 °C -100 °C to -30 °C -30 °C to 150 °C 150 °C to 760 °C 760 °C to 1200 °C	0.6 °C 0.5 °C 0.71 °C 0.71 °C 0.71 °C	
Type T	-250 °C to -150 °C -150 °C to 0 °C 0 °C to 120 °C 120 °C to 400 °C	0.73 °C 0.29 °C 0.33 °C 0.33 °C	
Type K	-200 °C to -100 °C	0.39 °C	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	-100 °C to -25 °C -25 °C to 120 °C 120 °C to 1000 °C 1000 °C to 1372 °C	0.22 °C 0.20 °C 0.31 °C 0.69 °C	
AC Current generate	15 A to 50 A (45 Hz to 1 kHz) 55 A to 120 A (45 Hz to 1 kHz)	700 x 10 ⁻⁶ + 2.4 mA x 5 1400 x 10 ⁻⁶ + 5.8 mA x 5	Cal Procedure -- ULID-003961 and ULID-003962 Using Multifunction Calibrator with Current coil of 5/10/20 turns by direct method
AC Current generate	30 A to 100 A (45 Hz to 1 kHz) 110 A to 200 A (45 Hz to 100 Hz)	700 x 10 ⁻⁶ + 2.4 mA x 10 1400 x 10 ⁻⁶ + 5.8 mA x 10	
AC Current generate	60 A to 200 A (45 Hz to 100 Hz) 220 A to 400 A (45 Hz to 100 Hz)	700 x 10 ⁻⁶ + 2.4 mA x 20 1400 x 10 ⁻⁶ + 5.8 mA x 20	
Shunt Calibration	1.00 A to 99.00 A	0.13 A	Cal Procedure – ULID-003963 Using Shunt Calibration Unit by direct method
Fixed Reference DC Resistances (4-wire reference resistance)	1 mΩ 5 mΩ 10 mΩ 20 mΩ 100 mΩ	2.4 % 1.2 % 1.2 % 1.2 % 1.2 %	Cal Procedure – ULID-003960 Using Reference Resistance Unit by direct method
Fixed Reference DC Resistances (2-wire reference resistance)	1.8 Ω 3 Ω 9 Ω 18 Ω 100 Ω 200 Ω 1.5 kΩ 10 kΩ 20 kΩ 90 kΩ 180 kΩ	1.2 % 1.2 % 1.2 % 1.2 % 1.2 % 1.2 % 1.2 % 1.2 % 1.2 % 1.2 % 1.2 %	Cal Procedure – ULID-003960 Using Reference Resistance Unit by direct method Cal Procedure – ULID-003960 Using Reference Resistance Unit by direct method
DC Resistance - Measure ⁴ (3-wire method)	0 Ω to 500 Ω 500 Ω to 5 kΩ 5 kΩ to 50 kΩ	0.064 % + 0.087 Ω 0.064 % + 0.87 Ω 0.064 % + 12 Ω	Cal Procedure – ULID-003960 Using Reference Resistance Unit by direct method



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
DC Resistance - Measure ⁴ (4-wire method)	(50 mA & 500 mV) 1 Ω to 10 Ω (DC)	0.032 % + 10 mΩ	Cal Procedure – ULID-003960 Using Fluke 5520A as current source, Fluke 289 to measure voltage drop V/I method
	(50 Hz)	0.35 % + 12 mΩ	
	(1 kHz)	0.46 % + 16 mΩ	
	(10 kHz)	0.52 % + 26 mΩ	
	(50 mA & 5 V) 10 Ω to 100 Ω (DC)	0.032 % + 7.6 mΩ	
	(50 Hz)	0.35 % + 74 mΩ	
	(1 kHz)	0.69 % + 61 mΩ	
	(10 kHz)	0.73 % + 110 mΩ	
	(5 mA & 5 V) 100 Ω to 1000 Ω (DC)	0.032 % + 130 mΩ	
	(50 Hz)	0.35 % + 750 mΩ	
	(1 kHz)	0.69 % + 750 mΩ	
	(10 kHz)	0.73 % + 910 mΩ	
	(0.5 mA & 5 V) 1 kΩ to 10 kΩ (DC)	0.032 % + 10 Ω	
	(50 Hz)	0.37 % + 12 Ω	
	(1 kHz)	0.70 % + 12 Ω	
	(10 kHz)	0.90 % + 13 Ω	
DC Voltage - Measure ⁴	0 mV to 50 mV	0.059 % + 0.024 mV	Cal Procedure – ULID-019807 Using Multimeter by direct method.
	0 mV to 500 mV	0.03 % + 0.024 mV	
	0 V to 5 V	0.03 % + 0.24 mV	
	0 V to 50 V	0.031 % + 2.4 mV	
	0 V to 500 V	0.035 % + 24 mV	
	0 V to 1000 V	0.036 % + 240 mV	
DC Current - Measure ⁴	0 μA to 500 μA	0.095 % + 0.24 μA	Cal Procedure – ULID-003966 Using Multimeter by direct method
	0 μA to 5000 μA	0.089 % + 0.24 μA	
	0 μA to 50 mA	0.062 % + 0.012 mA	
	0 mA to 400 mA	0.18 % + 0.024 mA	
	0 A to 5 A	0.36 % + 1.2 mA	
	0 A to 10 A	0.36 % + 2.4 mA	
AC Voltage - Measure ⁴	0 mV to 50 mV (45 Hz to 65 Hz)	3500 x 10 ⁻⁶ + 29 μV	Cal Procedure – ULID-003966 Using Multimeter by direct method
	(65 Hz to 10 kHz)	4700 x 10 ⁻⁶ + 29 μV	
AC Voltage - Measure ⁴ (continued)	(10 kHz to 20 kHz)	8100 x 10 ⁻⁶ + 47 μV	Cal Procedure – ULID-003966 Using Multimeter by direct method
	(20 kHz to 100 kHz)	41 x 10 ⁻³ + 47 μV	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	0 mV to 500 mV (45 Hz to 65 Hz)	$3500 \times 10^{-6} + 290 \mu\text{V}$	
	(65 Hz to 10 kHz)	$4700 \times 10^{-6} + 290 \mu\text{V}$	
	(10 kHz to 20 kHz)	$8100 \times 10^{-6} + 470 \mu\text{V}$	
	(20 kHz to 100 kHz)	$41 \times 10^{-3} + 470 \mu\text{V}$	
	0 V to 5 V (45 Hz to 65 Hz)	$3500 \times 10^{-6} + 2.9 \text{ mV}$	
	(65 Hz to 10 kHz)	$7000 \times 10^{-6} + 2.9 \text{ mV}$	
	0 V to 50 V (45 Hz to 65 Hz) (65 Hz to 10 kHz)	$3500 \times 10^{-6} + 29 \text{ mV}$ $4700 \times 10^{-6} + 29 \text{ mV}$	
	0 V to 500 V (45 Hz to 65 Hz)	$3500 \times 10^{-6} + 290 \text{ mV}$	
	0 V to 500 V (65 Hz to 10 kHz)	$4700 \times 10^{-6} + 290 \text{ mV}$	
	0 V to 1000 V (45 Hz to 65 Hz) (65 Hz to 10 kHz)	$3500 \times 10^{-6} + 2.9 \text{ V}$ $4700 \times 10^{-6} + 2.9 \text{ V}$	
AC Current - Measure ⁴	0.01 mA to 5 mA (45 Hz to 1 kHz)	$7100 \times 10^{-6} + 0.59 \mu\text{A}$	Cal Procedure – ULID-003966 Using Multimeter by direct method
	(1 kHz to 20 kHz)	$8300 \times 10^{-6} + 1.2 \mu\text{A}$	
	5 mA to 50 mA (45 Hz to 1 kHz) (1 kHz to 20 kHz)	$7300 \times 10^{-6} + 24 \mu\text{A}$ $7700 \times 10^{-6} + 24 \mu\text{A}$	
	50 mA to 400 mA (45 Hz to 1 kHz) (1 kHz to 20 kHz)	$7 \times 10^{-3} + 59 \mu\text{A}$ $36 \times 10^{-3} + 120 \mu\text{A}$	
	400 A to 5 A (45 Hz to 1 kHz)	$9500 \times 10^{-6} + 2.4 \text{ mA}$	



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 101, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	5 A to 10 A (45 Hz to 1 kHz)	9500 x 10 ⁻⁶ + 5.9 mA	
Sine Frequency - Measure ⁴	0 Hz to 99.999 Hz 0 Hz to 999.99 Hz 0 kHz to 9.9999 kHz 0 kHz to 99.999 kHz 0 kHz to 999.99 kHz	240 x 10 ⁻⁶ + 5.9 mHz 59 x 10 ⁻⁶ + 0.059 Hz 59 x 10 ⁻⁶ + 0.59 Hz 59 x 10 ⁻⁶ + 5.9 Hz 59 x 10 ⁻⁶ + 59 Hz	Cal Procedure – ULID-003966 Using Multimeter by direct method
Chemical & Gas			
Conductivity meter	1413 uS/cm 12.88 mS/cm	1.3% 1.3%	Cal Procedure – ULID-008347 Using Omega CA 150 Ref. electrolytic solution by direct method

¹The uncertainty covered by the Calibration and Measurement Capability (CMC) is expressed as the expanded uncertainty having a coverage probability of approximately 95 %. It is the smallest measurement uncertainty that a laboratory can achieve within its scope of accreditation when performing calibrations of a best existing device. The measurement uncertainty reported on a calibration certificate may be greater than that provided in the CMC due to the behavior of the calibration item and other factors that may contribute to the uncertainty of a specific calibration.

²When uncertainty is stated in relative terms (such as percent, a multiplier expressed as a decimal fraction or in scientific notation), it is in relation to instrument reading or instrument output, as appropriate, unless otherwise indicated.

³Capability is suitable for the calibration of measuring devices in the stated ranges.

⁴Capability is suitable for the calibration of devices intended to generate the indicated quantity in the stated ranges.

